# $PETCT\_NEMA\_2007\_IQ.Adult\_PET4$

27. februar 2013

# Indhold

1	Top	ogram	1	
	1.1	Routine	1	
	1.2	Scan	]	
<b>2</b>	AC CT NEMA			
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2		
3	Pau	se	4	
4	PE	Γ ΝΕΜΑ ΙQ	4	
	4.1	Routine	4	
	4.2	Scan	4	
	4.3	Recons	4	
		4.3.1 Recon 1	4	

# 1 Topogram

# 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

# 2 AC CT NEMA

## 2.1 Routine

- Eff. mAs: 50
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 3.65446mGy

• Scan time: 8.105 s

• Delay: 2.000 s

• Slice: 5 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 120

• Eff. mAs: 50

• kV: 120

 $\bullet$  Scan time: 8.105 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 5 mm

• Pitch: 0.80

• Direction: Craniocaudal

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT NEMA RTD

• Slice: 4

• Kernel: B19f

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 4.000

• No. of images: 56

#### 2.3.2 Recon 2

• Series description: AC CT NEMA

• Slice: 5

• Kernel: B19f

• Window: Abdomen

• Extended FoV: On

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 75

# 3 Pause

# 4 PET NEMA IQ

## 4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 370000000 Bequerels
- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 7.600000 Minutes

## 4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET NEMA IQ
- Recon range (bed): 1 to 1
- Output image type: Corrected

• Recon method: Iterative3D

• Iterations: 2

• Subsets: 24

• Image size: 200

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 5

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: Off ()

• Scatter correction: On

• Match CT slice location: Off

• Save intermediate data: Off