WB_Flourid_PET_5.MlAdult

27. september 2012

Indhold

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 34 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 34 mm
- Direction: Caudocranial
- Tube position: Top
- API: None
- Kernel: T80s sharp
- Window: Topogram Body

2 CT WB

2.1 Routine

- Eff. mAs: 40
- kV: 38
- CARE Dose4D: Off
- CTDlvol: 2.70069mGy
- Scan time: 0.8000000 s

- Delay: 4.000 s

- Slice: 3.00 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 40 grader

2.2 Scan

- Quality ref. mAs: 50

- Eff. mAs: 40

- kV: 120

- Scan time: 38.240 s

- Rotation time: 0.500 s

- Delay: 4.000 s

- Slice: 0.6 mm

- Pitch: 0.8000000

- Direction: Caudocranial

2.3 Recon

2.3.1 Recon 1

- Series description: AC CT

- Slice: 3.00

- Kernel: "B19f PET very smooth"

- Window: "Abdomen"

- Extended FoV: On

- FoV: 700

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 3.000

- No. of images: 574

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18

- Pharm.: Fluoride

- Inj. Dose: 200 MegaBequerels

- Inj. date (date/month - year): 21/08 - 2012

- Inj. time: 14:00:00

- Scan mode: Sinogram

- Scan range: Match CT FOV

- No. of beds: 11

- Scan duration/bed: 120 PtSecond

4.2 Scan

- Autoload: On

- Rebinner LUT: Off

- Scan output: Sinogram

- Sinogram mode: Trues

- Input trigger signal: None

- LLD (keV): 435

- ULD (keV): 650

4.3 Recon

4.3.1 Recon 1

- Series description: PET knogleskanning Image check

- Recon range (bed): ups..

- Output image type: Corrected

- Recon method: Iterative

- Iterations: 4

- Subsets: 8

- Image size: 168

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 4

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.2   Recon 2

- Series description: PET knogleskanning True X 2 mm

- Recon range (bed): ups..

- Output image type: Corrected

- Recon method: TrueX

- Iterations: 3

- Subsets: 21

- Image size: 336

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 2

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.3   Recon 3

- Series description: no AC

- Recon range (bed): ups..

- Output image type: Uncorrected

- Recon method: Iterative

- Iterations: 3

- Subsets: 21

- Image size: 336

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 4

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.4 Recon 4

- Series description: PET WB UDEN Trie X

- Recon range (bed): ups..

- Output image type: Corrected

- Recon method: Iterative

- Iterations: 3

- Subsets: 21

- Image size: 336

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 4

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off