$4_PET_CT_WB_FDG_Terapi_med_IV_Bolus.Adult$

23. oktober 2012

Indhold

Τ	Topogram			
	1.1	Routine	1	
	1.2	Scan	1	
2	PreMonitoring			
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
3	Monitoring			
	3.1	Routine	3	
	3.2	Scan	4	
	3.3	Recons	4	
		3.3.1 Recon 1	4	
4	WB CT			
	4.1	Routine	5	
	4.2	Scan	5	
	4.3	Recons	6	
		4.3.1 Recon 1	6	
		4.3.2 Recon 2	6	
		4.3.3 Recon 3	7	
		4.3.4 Recon 4	8	
		4.3.5 Recon 5	8	
5	Pau	se	9	
6	PE	T WB	9	
	6.1	Routine	9	
	6.2	Scan	9	
	6.3	Recons	10	
		6.3.1 Recon 1	10	
		6.3.2 Recon 2	11	
			11	

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

2 PreMonitoring

2.1 Routine

- Eff. mAs: 20
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDlvol: 1.186mGy

• Scan time: 0.500 s

• Delay: 2.000 s

• Slice: 10 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 20

• kV: 120

 \bullet Scan time: 0.500 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 10 mm

• Pitch: 0.00

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: PreMonitoring 10.0 eFoV

• Slice: 10

• Kernel: B19s

• Window: Abdomen

• Extended FoV: true

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 10.000

• No. of images: 1

3 Monitoring

3.1 Routine

• Eff. mAs: 20

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 35.58mGy

 \bullet Scan time: 0.500 s

• Delay: 10.000 s

• Slice: 10 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

3.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 20

• kV: 120

 \bullet Scan time: 0.500 s

• Rotation time: 0.500 s

• Delay: 10.000 s

• Slice: 10 mm

• Pitch: 0.00

• Direction: Craniocaudal

3.3 Recons

3.3.1 Recon 1

• Series description: Monitoring 10.0 eFoV

• Slice: 10

• Kernel: B19s

• Window: Abdomen

• Extended FoV: true

• FoV: 780

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 10.000

• No. of images: 30

4 WB CT

4.1 Routine

• Eff. mAs: 200

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 14.56mGy

 \bullet Scan time: 32.530 s

• Delay: 4.000 s

• Slice: 2 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

4.2 Scan

• Quality ref. mAs: 200

• Eff. mAs: 200

• kV: 120

• Scan time: 32.530 s

• Rotation time: 0.500 s

 \bullet Delay: 4.000 s

• Slice: 2 mm

• Pitch: 0.80

• Direction: Craniocaudal

4.3 Recons

4.3.1 Recon 1

- Series description: tmp WB CT RTD
- Slice: 4
- Kernel: B19f
- Window: Abdomen
- Extended FoV: false
- FoV: 500
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 245

4.3.2 Recon 2

- Series description: AC CT 2 mm
- Slice: 2
- Kernel: B19f
- Window: Abdomen
- Extended FoV: true
- FoV: 780
- Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 489

4.3.3 Recon 3

• Series description: CT WB 2.0 B30f

• Slice: 2

• Kernel: B30f

• Window: Abdomen

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 489

4.3.4 Recon 4

• Series description: CT WB 2.0 cor B30f

• Slice: 2

• Kernel: B30f

• Window: Abdomen

• Extended FoV: false

• FoV: 977

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: 3D

• Recon Axis: Coronal

• Image order: AP

• Recon increment: 2.000

• No. of images: 251

4.3.5 Recon 5

• Series description: CT WB 2.0 sag B30f

• Slice: 2

• Kernel: B30f

• Window: Abdomen

• Extended FoV: false

• FoV: 977

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: 3D

• Recon Axis: Sagittal

• Image order: LeftToRight

• Recon increment: 2.000

• No. of images: 489

5 Pause

6 PET WB

6.1 Routine

• Isotope: F-18

• Pharm.: Fluorodeoxyglucose

 $\bullet\,$ Inj. Dose: 244000000.000000 Bequerels

 $\bullet\,$ Inj. date (date/month - year): 29/05 - 2012

• Inj. time: 09:14:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

 $\bullet\,$ No. of beds: Not given. Check recon range

• Scan duration/bed: 2.000000 Minutes

6.2 Scan

• Autoload: true

• Rebinner LUT: N/A

• Scan output: Sinogram

- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- \bullet ULD (keV): N/A

6.3 Recons

6.3.1 Recon 1

- Series description: PET WB Image check HD
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 12
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

6.3.2 Recon 2

- Series description: PET WB Corr TrueX+TOF 2 mm
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

6.3.3 Recon 3

- Series description: PET WB Corr TrueX
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 24
- Image size: 400

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 2.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: Off (1)

• Scatter correction: true

• Match CT slice location: true

• Save intermediate data: false