

PROJ\_Dotatoc\_PET\_5.MlAdult\_PET5

3. december 2012

# Indhold

<b>1</b>	<b>Topogram</b>	<b>1</b>
1.1	Routine . . . . .	1
1.2	Scan . . . . .	1
<b>2</b>	<b>Lavdos CT WB</b>	<b>1</b>
2.1	Routine . . . . .	1
2.2	Scan . . . . .	2
2.3	Recons . . . . .	2
2.3.1	Recon 1 . . . . .	2
<b>3</b>	<b>Pause</b>	<b>3</b>
<b>4</b>	<b>PET WB</b>	<b>3</b>
4.1	Routine . . . . .	3
4.2	Scan . . . . .	4
4.3	Recons . . . . .	4
4.3.1	Recon 1 . . . . .	4
4.3.2	Recon 2 . . . . .	5

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API: None
- Kernel: 80s shar
- Window: Topogram Body

# **2 Lavdos CT WB**

## **2.1 Routine**

- Eff. mAs: 40
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 2.70069mGy

- Scan time: 18.670 s
- Delay: 4.000 s
- Slice: 3.00 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 50
- Eff. mAs: 40
- kV: 120
- Scan time: 18.670 s
- Rotation time: 0.500 s
- Delay: 4.000 s
- Slice: 3.00 mm
- Pitch: 0.8000000
- Direction: Craniocaudal

## **2.3 Recons**

### **2.3.1 Recon 1**

- Series description: AC CT
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 274

### **3    Pause**

## **4    PET WB**

### **4.1    Routine**

- Isotope: Ga-68
- Pharm.: Dotatoc
- Inj. Dose: 1 MegaBequerels
- Inj. date (date/month - year): 23/03 - 2011
- Inj. time: 13:06:00
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 3 Minutes

## **4.2 Scan**

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

## **4.3 Recons**

### **4.3.1 Recon 1**

- Series description: PET WB Corrected
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

### 4.3.2 Recon 2

- Series description: PET Dotatoc WB Corr. True X 2 mm
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off