# $4\_PETCT\_WholeBody\_Dynamic.Adult$

8. oktober 2012

## Indhold

1	Topogram 1				
	1.1	Routine	1		
	1.2	Scan	1		
2	$\mathbf{CT}$	$_{ m VB}$	1		
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recon	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	3		
		2.3.3 Recon 3	4		
		2.3.4 Recon 4	4		
3	Pau	e	5		
4	PET	WB	5		
	4.1	Routine	5		
	4.2	Scan	5		
	4.3	Recon	6		
		4.3.1 Recon 1	6		
		4.3.2 Recon 2	7		

## 1 Topogram

## 1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

#### 1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API:
- Kernel: T80s
- Window: Topogra

## 2 CT WB

## 2.1 Routine

- Eff. mAs: 210
- kV: 120
- CARE Dose4D: On
- CTDlvol: 15.288mGy
- Scan time: 28.450 s

• Delay: 2.000 s

• Slice: 5 mm

 $\bullet$  No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 210

• Eff. mAs: 210

• kV: 120

• Scan time: 28.450 s

 $\bullet$  Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 5 mm

• Pitch: 0.80

• Direction: Craniocaudal

## 2.3 Recon

#### 2.3.1 Recon 1

• Series description: tmp CT WB RTD

• Slice: 4

• Kernel: "B19f"

• Window: "Abdomen"

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 213

#### 2.3.2 Recon 2

- Series description: AC CT WB
- Slice: 5
- Kernel: "B19f"
- Window: "Abdomen"
- Extended FoV: true
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 284

#### 2.3.3 Recon 3

• Series description: CT WB 3.0 B30f

• Slice: 3

• Kernel: "B30f"

• Window: "Abdomen"

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 426

#### 2.3.4 Recon 4

• Series description: CT LUNG 3.0 B70f

• Slice: 3

• Kernel: "B70f"

• Window: "Lung"

• Extended FoV: false

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 2.000

• No. of images: 426

## 3 Pause

## 4 PET WB

## 4.1 Routine

• Isotope: F-18

• Pharm.: Fluorodeoxyglucose

• Inj. Dose: 3

 $\bullet\,$  Inj. date (date/month - year): 06/09 - 2010

 $\bullet\,$  Inj. time: 05:42:01

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: N/A

• Scan duration/bed: 3.000000 Minutes

#### 4.2 Scan

• Autoload: true

• Rebinner LUT: N/A

• Scan output: Sinogram

- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

## 4.3 Recon

#### 4.3.1 Recon 1

- Series description: PET WB Dynamic
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

#### 4.3.2 Recon 2

- Series description: PET WB Dynamic Uncorrected
- Recon range (bed): ups...
- Output image type: Uncorrected
- Recon method: Iterative3D
- Iterations: 2
- Subsets: 24
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false