4_PET_CT_LM_FET_Brain_Terapi_stat.Adult_PET4

11. januar 2013

Indhold

1 Topogram

1.1 Routine

- mA: 35

- kV: 120

- Topogram length: 256 mm

- Tube position: Lateral

1.2 Scan

- mA: 35

- kV: 120

- Delay: 4s

- Topogram length: 256 mm

- Direction: Craniocaudal

- Tube position: Lateral

- API: API

- Kernel: 20

- Window: Topogram

2 CT Brain

2.1 Routine

- Eff. mAs: 340

- kV: 120

- CARE Dose4D: Off

- CareDoseType: CareDoseAEC

- CTDlvol: 58.072mGy

- Scan time: 48.120 s

- Delay: 2.000 s

- Slice: 1 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 0.0 grader

2.2   Scan

- Quality ref. mAs: 350

- Eff. mAs: 340

- kV: 120

- Scan time: 48.120 s

- Rotation time: 1.000 s

- Delay: 2.000 s

- Slice: 1 mm

- Pitch: 0.80

- Direction: Caudocranial

2.3   Recons

2.3.1   Recon 1

- Series description: tmp CT Brain RTD

- Slice: 5

- Kernel: H19s

- Window: Cerebrum

- Extended FoV: Off

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Caudocranial

- Recon increment: 5.000

- No. of images: 45

2.3.2 Recon 2

- Series description: AC CT 1mm Brain

- Slice: 1

- Kernel: H19s

- Window: Cerebrum

- Extended FoV: Off

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Caudocranial

- Recon increment: 1.000

- No. of images: 224

2.3.3   Recon 3

- Series description: CT Brain 1.0 H31s

- Slice: 1

- Kernel: H31s

- Window: Cerebrum

- Extended FoV: Off

- FoV: 250

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Caudocranial

- Recon increment: 1.000

- No. of images: 224

2.3.4   Recon 4

- Series description: CT Brain 3.0 H31s

- Slice: 3

- Kernel: H31s

- Window: Cerebrum

- Extended FoV: Off

- FoV: 250

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Caudocranial

- Recon increment: 3.000

- No. of images: 75

3   Pause

4   PET Brain LM

4.1   Routine

- Isotope: F-18

- Pharm.: FET

- Inj. Dose: 205000000 Bequerels

- Inj. date (date/month - year): 19/09 - 2012

- Inj. time: 12:54:04

- Scan mode: List mode

- Scan range:

- No. of beds: Not given. Check recon range

- Scan duration/bed: 20.000000 Minutes

4.2   Scan

- Autoload: Off

- Rebinner LUT: Off

- Scan output: List mode

- Sinogram mode: false

- Input trigger signal: None

- LLD (keV): n/a

- ULD (keV): n/a

4.3   Recons

4.3.1   Recon 1

- Series description: FET stat 20-40min 5mm CT3mm

- Recon range (bed): Automatic/All/dunno

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 4

- Subsets: 12

- Image size: 400

- Zoom: 2.500000

- Filter: Gaussian

- FWHM (mm): 5.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: Off ()

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.2 Recon 2

- Series description: FET stat 25-35min 5mm CT3mm

- Recon range (bed): Automatic/All/dunno

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 4

- Subsets: 12

- Image size: 400

- Zoom: 2.500000

- Filter: Gaussian

- FWHM (mm): 5.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: Off ()

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off