$4_PETCT_AC_WholeBody_noswa.Adult$

8. oktober 2012

Indhold

1	Top	1 .0															1									
	1.1	Routin	e																							1
	1.2	Scan .											•		•					•	•	•		•		1
2	CT WB															1										
	2.1	Routin	e																							1
	2.2	Scan .																								2
	2.3	Recon																								2
		2.3.1	Recon 1																							2
		2.3.2	Recon 2																							3
3	Pau	ıse																								4
4	PET WB															4										
	4.1	Routin	e																							4
	4.2	Scan .																								4
	4.3	Recon																								5
		4.3.1	Recon 1																							5
		4.3.2	Recon 2																	 						5

1 Topogram

1.1 Routine

- mA: 35
- kV: 120
- Topogram length: 1024 mm
- Tube position: Top

1.2 Scan

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 1024 mm
- Direction: Craniocaudal
- Tube position: Top
- API:
- Kernel: T80s
- Window: Topogra

2 CT WB

2.1 Routine

- Eff. mAs: 50
- kV: 120
- CARE Dose4D: Off
- CTDlvol: 3.64mGy
- Scan time: 26.780 s

• Delay: 3.000 s

• Slice: 5 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 120

• Eff. mAs: 50

• kV: 120

• Scan time: 26.780 s

 \bullet Rotation time: 0.500 s

• Delay: 3.000 s

• Slice: 5 mm

• Pitch: 0.80

• Direction: Craniocaudal

2.3 Recon

2.3.1 Recon 1

• Series description: tmp AC CT Body RTD

• Slice: 4

• Kernel: "B19f"

• Window: "Abdomen"

• Extended FoV: false

• FoV: 500

• Center X: 0

• Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 201

2.3.2 Recon 2

- Series description: AC CT Body
- Slice: 5
- Kernel: "B19f"
- Window: "Abdomen"
- Extended FoV: true
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 3.000
- No. of images: 267

3 Pause

4 PET WB

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 3
- Inj. date (date/month year): 07/12 2008
- Inj. time: 01:57:55
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: N/A
- Scan duration/bed: 2.000000 Minutes

4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recon

4.3.1 Recon 1

- Series description: PET WB
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 2
- Subsets: 24
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

4.3.2 Recon 2

- Series description: PET WB NAC
- Recon range (bed): ups..
- Output image type: Uncorrected
- Recon method: Iterative3D
- Iterations: 2
- Subsets: 24

• Image size: 200

• Zoom: 1

• Filter: Gaussian

• FWHM (mm): 5

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: true

• Match CT slice location: true

• Save intermediate data: false