# $Hjerne\_FET\_PET\_CT\_Stat\_LM\_5.MlAdult$

1. oktober 2012

## Indhold

1	Top	ogram	1		
			1		
	1.2	Scan	1		
<b>2</b>	CT WB				
	2.1	Routine	1		
	2.2	Scan	2		
	2.3	Recon	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	3		
3	Pau	${f se}$	4		
4	PET	$\Gamma  ext{ WB}$	4		
	4.1	Routine	4		
	4.2	Scan	4		
	4.3	Recon	1		
		4.3.1 Recon 1	5		

## 1 Topogram

#### 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: None

• Kernel: T20s standard

• Window: Topogram Head

## 2 CT WB

## 2.1 Routine

• Eff. mAs: 340

• kV: 17

• CARE Dose4D: Off

• CTDlvol: 57.596mGy

 $\bullet$  Scan time: 1.2000000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 340 grader

## 2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 340

• kV: 120

• Scan time: 17.000 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 0.6 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

#### 2.3 Recon

#### 2.3.1 Recon 1

 $\bullet\,$  Series description: AC CT

• Slice: 3.00

• Kernel: "H19s PET very smooth"

• Window: "Cerebrum"

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 3.000
- No. of images: 74

#### 2.3.2 Recon 2

- Series description: CT Brain 3.0 H31s
- Slice: 3.00
- Kernel: "H31s medium smooth +"
- Window: "Cerebrum"
- Extended FoV: Off
- FoV: 300
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Caudocranial
- Recon increment: 3.000
- No. of images: 74

## 3 Pause

## 4 PET WB

## 4.1 Routine

- Isotope: F-18
- Pharm.: FET
- Inj. Dose: 1 MegaBequerels
- $\bullet\,$  Inj. date (date/month year): 04/06 2012
- Inj. time: 15:08:00
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: 1
- Scan duration/bed: 20 PtMinute

#### 4.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: List mode
- Sinogram mode: Trues
- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

## 4.3 Recon

#### 4.3.1 Recon 1

- Series description: PET Brain
- Recon range (bed): ups..
- Output image type: NoRecon
- Recon method: Iterative3D
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 3
- Filter: Gaussian
- FWHM (mm): 3
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off