# TUMOR\_HUND\_ATSM.MlAdult\_PET5

20. februar 2013

# Indhold

1	Topogram 1				
	1.1	Routine	]		
	1.2	Scan	]		
2	CT TUMOR				
	2.1	Routine	]		
	2.2	Scan	2		
	2.3	Recons	2		
		2.3.1 Recon 1	2		
		2.3.2 Recon 2	٠		
3	Pau	${f se}$	4		
4	PE	Γ STATIC	4		
	4.1	Routine	4		
	4.2	Scan	4		
	4.3	Recons	4		
		4.3.1 Recon 1	4		
		4.3.2 Recon 2	Ę		

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 0 mm

• Tube position: Top

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 0 mm

• Direction: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

## 2 CT TUMOR

## 2.1 Routine

• Eff. mAs: 170

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 11.4448mGy

• Scan time: 4.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

 $\bullet$  No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 170

• Eff. mAs: 170

• kV: 120

 $\bullet$  Scan time: 4.000 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

#### 2.3.2 Recon 2

• Series description: CT TUMOR 3.0 B30f

• Slice: 3.00

• Kernel: B30f medium smooth

• Window: Abdomen

• Extended FoV: Off

• FoV: 500

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

## 3 Pause

## 4 PET STATIC

## 4.1 Routine

• Isotope: Cu-64

• Pharm.: ATSM

• Inj. Dose: 226 MegaBequerels

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 20 Minutes

## 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

#### 4.3 Recons

#### 4.3.1 Recon 1

• Series description: PET TUMOR Image check

• Recon range (bed): -1 to -1

• Output image type: Corrected

- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 3
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- Series description: PET TUMOR True X
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: Off

• Match CT slice location: On

• Save intermediate data: Off