Fantom.MlAdult_PET5

20. februar 2013

Indhold

1	Top	ogram	1
	1.1	Routine	1
	1.2	Scan	1
2	\mathbf{CT}	Uden spejl	1
	2.1	Routine	1
	2.2	Scan	2
	2.3	Recons	2
		2.3.1 Recon 1	2
		2.3.2 Recon 2	3
3	Pau	ase	4
4	\mathbf{CT}	med spejl	4
	4.1	Routine	4
	4.2	Scan	4
	4.3	Recons	5
		4.3.1 Recon 1	5
		4.3.2 Recon 2	5
5	Pau	ise	6
6	PE	Γ med spejl	6
	6.1	Routine	6
	6.2	Scan	6
	6.3	Recons	7
		6.3.1 Recon 1	7
		6.3.2 Recon 2	8

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Top

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

2 CT Uden spejl

2.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 15.1475mGy

• Scan time: 4.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

• Scan time: 4.000 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: tmp CT Uden spejl RTD

• Slice: 5.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 17.300

• No. of images: 14

2.3.2 Recon 2

• Series description: AC CT Uden Spejl

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

3 Pause

4 CT med spejl

4.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 15.1475mGy

• Scan time: 4.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

4.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

• Scan time: 4.000 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 1.2000000

• Direction: Craniocaudal

4.3 Recons

4.3.1 Recon 1

- Series description: tmp CT med spejl RTD
- Slice: 5.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 17.300
- No. of images: 14

4.3.2 Recon 2

- Series description: AC CT Med Spejl
- Slice: 3.00
- Kernel: B19f PET very smooth
- Window: Abdomen
- Extended FoV: On
- FoV: 700
- Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 74

5 Pause

6 PET med spejl

6.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 25 MegaBequerels

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 5 Minutes

6.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

6.3 Recons

6.3.1 Recon 1

- Series description: AC PET m spejl CT u spejl
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

6.3.2 Recon 2

- Series description: AC PET m spejl CT m spejl
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 6
- Subsets: 16
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off