# WB\_Flourid\_PET\_5.MlAdult\_PET5

23. januar 2013

# Indhold

| 1        | 1 Topogram                          | 1     |
|----------|-------------------------------------|-------|
|          | 1.1 Routine                         | <br>1 |
|          | 1.2 Scan                            | <br>1 |
| <b>2</b> | 2 Lav CT knogle                     | 1     |
|          | 2.1 Routine                         | <br>1 |
|          | 2.2 Scan                            | <br>2 |
|          | 2.3 Recons                          | <br>2 |
|          | 2.3.1 Recon $1 \dots \dots$         | <br>2 |
| 3        | 3 Pause                             | 3     |
| 4        | 4 PET knogle                        | 3     |
|          | 4.1 Routine                         | <br>3 |
|          | 4.2 Scan                            | <br>3 |
|          | 4.3 Recons                          | <br>4 |
|          | $4.3.1$ Recon $1 \dots \dots \dots$ | <br>4 |
|          | $4.3.2$ Recon $2 \dots \dots \dots$ | <br>5 |
|          | 4.3.3 Recon 3                       | <br>5 |
|          | 101 D                               | <br>6 |

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 34 mm

• Tube position: Top

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 34 mm

• Direction: Caudocranial

• Tube position: Top

• API: None

• Kernel: 80s shar

• Window: Topogram Body

## 2 Lav CT knogle

### 2.1 Routine

• Eff. mAs: 40

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 2.70069mGy

• Scan time: 38.240 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 50

• Eff. mAs: 40

• kV: 120

 $\bullet$  Scan time: 38.240 s

• Rotation time: 0.500 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 0.8000000

• Direction: Craniocaudal

### 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: B19f PET very smooth

• Window: Abdomen

• Extended FoV: On

• FoV: 700

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 574

## 3 Pause

## 4 PET knogle

## 4.1 Routine

• Isotope: F-18

• Pharm.: Fluoride

• Inj. Dose: 200 MegaBequerels

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 11

• Scan duration/bed: 120 Seconds

### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

- Input trigger signal: None
- LLD (keV): 435
- ULD (keV): 650

#### 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET knogleskanning Image check
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 4
- Subsets: 8
- Image size: 168
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.2 Recon 2

- Series description: PET knogleskanning True X 2 mm
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.3 Recon 3

- Series description: no AC
- Recon range (bed): -1 to -1
- Output image type: Uncorrected
- Recon method: Iterative
- Iterations: 3
- Subsets: 21
- Image size: 336

- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 4.3.4 Recon 4

- Series description: PET WB UDEN Trie X
- Recon range (bed): -1 to -1
- Output image type: Corrected
- Recon method: Iterative
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off