3_PET_AC_FDG_Brain_Neurostat.Adult_PET4

3. december 2012

Indhold

1	Topogram 1					
	1.1	Routine	1			
	1.2	Scan	1			
2	CT Brain					
	2.1	Routine	1			
	2.2	Scan	2			
	2.3	Recons	2			
		2.3.1 Recon 1	2			
		2.3.2 Recon 2	3			
3	Pau	se	4			
4	PE	' Brain	4			
	4.1	Routine	4			
	4.2	Scan	4			
	4.3	Recons	5			
		4.3.1 Recon 1	5			
		4.3.2 Recon 2	5			

1 Topogram

1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: API

• Kernel: 20

• Window: Topogram

2 CT Brain

2.1 Routine

• Eff. mAs: 30

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

• CTDlvol: 5.124mGy

• Scan time: 2.640 s

• Delay: 2.000 s

• Slice: 3 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

2.2 Scan

• Quality ref. mAs: 30

• Eff. mAs: 30

• kV: 120

 \bullet Scan time: 2.640 s

• Rotation time: 1.000 s

• Delay: 2.000 s

• Slice: 3 mm

• Pitch: 1.50

• Direction: Craniocaudal

2.3 Recons

2.3.1 Recon 1

• Series description: tmp CT Brain RTD

• Slice: 5

• Kernel: H19s

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 5.000

• No. of images: 1

2.3.2 Recon 2

• Series description: AC CT

• Slice: 3

• Kernel: H19s

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Craniocaudal

• Recon increment: 3.000

• No. of images: 1

3 Pause

4 PET Brain

4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 200000000.000000 Bequerels
- Inj. date (date/month year): 15/05 2012
- Inj. time: 09:15:23
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 10.000000 Minutes

4.2 Scan

- Autoload: On
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- \bullet Series description: PET Brain Corrected 3 mm zoom 3
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 24
- Image size: 400
- Zoom: 3.000000
- Filter: Gaussian
- FWHM (mm): 3.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

4.3.2 Recon 2

- Series description: Neurostat
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 8

• Image size: 128

• Zoom: 3.000000

• Filter: Gaussian

• FWHM (mm): 8.000000

• Offset X: 0 mm

• Offset Y: 0 mm

• Attenuation correction: On (1)

• Scatter correction: On

• Match CT slice location: On

• Save intermediate data: Off