

# PETCT\_Cardiac.Adult\_PET4

23. januar 2013

# Indhold

|          |                    |          |
|----------|--------------------|----------|
| <b>1</b> | <b>Topogram</b>    | <b>1</b> |
| 1.1      | Routine . . . . .  | 1        |
| 1.2      | Scan . . . . .     | 1        |
| <b>2</b> | <b>CT Cardiac</b>  | <b>1</b> |
| 2.1      | Routine . . . . .  | 1        |
| 2.2      | Scan . . . . .     | 2        |
| 2.3      | Recons . . . . .   | 2        |
| 2.3.1    | Recon 1 . . . . .  | 2        |
| 2.3.2    | Recon 2 . . . . .  | 3        |
| <b>3</b> | <b>Pause</b>       | <b>4</b> |
| <b>4</b> | <b>PET Cardiac</b> | <b>4</b> |
| 4.1      | Routine . . . . .  | 4        |
| 4.2      | Scan . . . . .     | 4        |
| 4.3      | Recons . . . . .   | 4        |
| 4.3.1    | Recon 1 . . . . .  | 4        |

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 512 mm
- Tube position: Top

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 512 mm
- Direction: Craniocaudal
- Tube position: Top
- API: API
- Kernel: 80
- Window: Topogram

# **2 CT Cardiac**

## **2.1 Routine**

- Eff. mAs: 11
- kV: 120
- CARE Dose4D: Off
- CareDoseType: CareDoseAEC
- CTDIvol: 0.8008mGy

- Scan time: 4.630 s
- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 170
- Eff. mAs: 11
- kV: 120
- Scan time: 4.630 s
- Rotation time: 0.500 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 1.50
- Direction: Craniocaudal

## **2.3 Recons**

### **2.3.1 Recon 1**

- Series description: tmp CT Cardiac RTD
- Slice: 4
- Kernel: B18f
- Window: Abdomen
- Extended FoV: Off
- FoV: 500
- Center X: 0

- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 56

### **2.3.2 Recon 2**

- Series description: AC CT Cardiac
- Slice: 3
- Kernel: B18f
- Window: Abdomen
- Extended FoV: On
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 111

## **3 Pause**

## **4 PET Cardiac**

### **4.1 Routine**

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 370000000 Bequerels
- Scan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 10.000000 Minutes

### **4.2 Scan**

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

### **4.3 Recons**

#### **4.3.1 Recon 1**

- Series description: PET Cardiac
- Recon range (bed): 1 to 1
- Output image type: Corrected

- Recon method: Backprojection
- Iterations: 2
- Subsets: 24
- Image size: 200
- Zoom: 2.000000
- Filter: Gaussian
- FWHM (mm): 8.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off