

# 4\_PETCT\_Cardiac\_Gated.Adult

8. oktober 2012

# Indhold

<b>1</b>	<b>Topogram</b>	<b>1</b>
1.1	Routine . . . . .	1
1.2	Scan . . . . .	1
<b>2</b>	<b>CT WB</b>	<b>1</b>
2.1	Routine . . . . .	1
2.2	Scan . . . . .	2
2.3	Recon . . . . .	2
2.3.1	Recon 1 . . . . .	2
2.3.2	Recon 2 . . . . .	3
<b>3</b>	<b>Pause</b>	<b>4</b>
<b>4</b>	<b>PET WB</b>	<b>4</b>
4.1	Routine . . . . .	4
4.2	Scan . . . . .	4
4.3	Recon . . . . .	5
4.3.1	Recon 1 . . . . .	5
4.3.2	Recon 2 . . . . .	5

# **1 Topogram**

## **1.1 Routine**

- mA: 35
- kV: 120
- Topogram length: 256 mm
- Tube position: Top

## **1.2 Scan**

- mA: 35
- kV: 120
- Delay: 4s
- Topogram length: 256 mm
- Direction: Craniocaudal
- Tube position: Top
- API:
- Kernel: T80s
- Window: Topogra

# **2 CT WB**

## **2.1 Routine**

- Eff. mAs: 11
- kV: 120
- CARE Dose4D: Off
- CTDIvol: 0.8008mGy
- Scan time: 4.700 s

- Delay: 2.000 s
- Slice: 3 mm
- No. of images: Samme som i foerste recon, slet?(y/n)
- Tilt: 0.0 grader

## **2.2 Scan**

- Quality ref. mAs: 170
- Eff. mAs: 11
- kV: 120
- Scan time: 4.700 s
- Rotation time: 0.500 s
- Delay: 2.000 s
- Slice: 3 mm
- Pitch: 1.50
- Direction: Craniocaudal

## **2.3 Recon**

### **2.3.1 Recon 1**

- Series description: tmp CT Cardiac RTD
- Slice: 4
- Kernel: "B19f"
- Window: "Abdomen"
- Extended FoV: false
- FoV: 500
- Center X: 0
- Center Y: 0

- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 4.000
- No. of images: 57

### **2.3.2 Recon 2**

- Series description: AC CT Cardiac
- Slice: 3
- Kernel: "B19f"
- Window: "Abdomen"
- Extended FoV: true
- FoV: 780
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: Axial
- Recon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 113

### **3 Pause**

## **4 PET WB**

### **4.1 Routine**

- Isotope: Rb-82
- Pharm.: Rubidium chloride
- Inj. Dose: 3
- Inj. date (date/month - year): 06/09 - 2010
- Inj. time: 08:01:56
- Scan mode: List mode
- Scan range: Match CT FOV
- No. of beds: N/A
- Scan duration/bed: 10.000000 Minutes

### **4.2 Scan**

- Autoload: true
- Rebinner LUT: N/A
- Scan output: List mode
- Sinogram mode: true
- Input trigger signal: ECG
- LLD (keV): N/A
- ULD (keV): N/A

## **4.3 Recon**

### **4.3.1 Recon 1**

- Series description: PET Cardiac
- Recon range (bed): ups..
- Output image type: Uncorrected
- Recon method: Backprojection
- Iterations: 4
- Subsets: 12
- Image size: 200
- Zoom: 2.000000
- Filter: Gaussian
- FWHM (mm): 8.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false

### **4.3.2 Recon 2**

- Series description: PET Cardiac Gated
- Recon range (bed): ups..
- Output image type: Corrected
- Recon method: Backprojection
- Iterations: 4
- Subsets: 12

- Image size: 200
- Zoom: 2.000000
- Filter: Gaussian
- FWHM (mm): 8.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false