# $4\_DOHAN\_Terapi\_PET\_CT\_FDG\_nr2.Adult\_PET4$

30. maj 2013

## Indhold

1	Top	Topogram													1											
	1.1	Routine																						 		1
	1.2	Scan		•		•														•	•	•		 	•	1
2	PreMonitoring														1											
	2.1	Routine																 						 		1
	2.2	Scan																 						 	•	1
	2.3	Recons																 						 		2
		2.3.1 Recon 1																						 		2
3	Moi	Monitoring														2										
	3.1	Routine																						 		2
	3.2	Scan																 						 		2
	3.3	Recons																								3
		3.3.1 Recon 1																								3
4	WB CT													3												
	4.1	Routine																 						 		3
	4.2	Scan																 						 	•	3
	4.3	Recons																								4
																										4
		4.3.2 Recon 2																 						 		4
		4.3.3 Recon 3																 						 		4
																										5
																										5
5	Pau	se																								6
6	PE	ГWB																								6
	6.1	Routine																 						 		6
	6.2	Scan																 						 	•	6
	6.3	Recons																								6
																										6
																										7
																										7
																										7
		6.3.5 Recon 5		•		•	•		•	•	•	•	•	•	•	 •	•	 •	•	•	•	•	•	 •	•	8

## 1 Topogram

#### 1.1 Routine

mA: 35kV: 120

• Topogram length: 512 mm

• Tube position: Top

#### 1.2 Scan

mA: 35kV: 120Delay: 4s

Topogram length: 512 mmDirection: Craniocaudal

• Tube position: Top

API: APIKernel: 80

• Window: Topogram

## 2 PreMonitoring

#### 2.1 Routine

• Eff. mAs: 20

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

CTDlvol: 1.18627mGyScan time: 0.500 s

• Delay: 2.000 s

• Slice: 10 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 2.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 20

• kV: 120

Scan time: 0.500 sRotation time: 0.500 s

Delay: 2.000 sSlice: 10 mmPitch: 0.00

• Direction: Craniocaudal

#### 2.3 Recons

#### 2.3.1 Recon 1

• Series description: PreMonitoring 10.0 eFoV

Slice: 10Kernel: B19s

Window: AbdomenExtended FoV: On

FoV: 780Center X: 0Center Y: 0Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CraniocaudalRecon increment: 10.000

• No. of images: 1

## 3 Monitoring

#### 3.1 Routine

• Eff. mAs: 20

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

CTDlvol: 35.5881mGy
 Scan time: 0.500 s

Delay: 10.000 sSlice: 10 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 3.2 Scan

• Quality ref. mAs: 225

Eff. mAs: 20kV: 120

Scan time: 0.500 sRotation time: 0.500 s

Delay: 10.000 sSlice: 10 mm

• Pitch: 0.00

• Direction: Craniocaudal

#### 3.3 Recons

#### 3.3.1 Recon 1

• Series description: Monitoring 10.0 eFoV

Slice: 10Kernel: B19s

Window: AbdomenExtended FoV: On

FoV: 780Center X: 0Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CraniocaudalRecon increment: 10.000

• No. of images: 30

## 4 WB CT

#### 4.1 Routine

• Eff. mAs: 225

• kV: 120

• CARE Dose4D: On

• CareDoseType: CareDoseAEC

CTDlvol: 16.4451mGyScan time: 32.745 s

 $\bullet$  Delay: 4.000 s

• Slice: 2 mm

 $\bullet\,$  No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

#### 4.2 Scan

• Quality ref. mAs: 225

• Eff. mAs: 225

• kV: 120

Scan time: 32.745 sRotation time: 0.500 s

• Delay: 4.000 s

Slice: 2 mmPitch: 0.80

• Direction: Craniocaudal

#### 4.3 Recons

#### 4.3.1 Recon 1

• Series description: WB CT RTD

• Slice: 4

• Kernel: B19f

Window: AbdomenExtended FoV: Off

FoV: 500Center X: 0Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

Image order: CraniocaudalRecon increment: 4.000

• No. of images: 245

#### 4.3.2 Recon 2

• Series description: AC CT 2 mm

• Slice: 2

• Kernel: B19f

Window: AbdomenExtended FoV: On

FoV: 780Center X: 0Center Y: 0Mirroring: None

• Extended CT scale: Standard

Recon job: AxialRecon Axis: Axial

• Image order: Craniocaudal

 $\bullet$  Recon increment: 2.000

• No. of images: 489

#### 4.3.3 Recon 3

• Series description: CT WB 2.0 B30f

• Slice: 2

• Kernel: B30f

- Window: AbdomenExtended FoV: Off
- FoV: 500Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: AxialRecon Axis: Axial
- Image order: Craniocaudal
- Recon increment: 2.000
- No. of images: 489

#### 4.3.4 Recon 4

- Series description: CT WB 2.0 cor B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
   Extended FoV: Off
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Coronal
- Image order: AP
- Recon increment: 2.000
- No. of images: 251

#### 4.3.5 Recon 5

- Series description: CT WB 2.0 sag B30f
- Slice: 2
- Kernel: B30f
- Window: Abdomen
- Extended FoV: Off
- FoV: 977
- Center X: 0
- Center Y: 0
- Mirroring: None
- Extended CT scale: Standard
- Recon job: 3D
- Recon Axis: Sagittal

- Image order: LeftToRight
  Recon increment: 2.000
- No. of images: 489

## 5 Pause

## 6 PET WB

#### 6.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 0 BequerelsScan mode: Sinogram
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 2.000000 Minutes

## 6.2 Scan

- Autoload: On
- Rebinner LUT: Off
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): n/a
- ULD (keV): n/a

#### 6.3 Recons

#### 6.3.1 Recon 1

- Series description: PET WB Image check HD
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 12
- Image size: 200
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On

- Match CT slice location: On
- Save intermediate data: Off

#### 6.3.2 Recon 2

- Series description: PET WB Corr TrueX+TOF 2 mm
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueXTOF
- Iterations: 2
- Subsets: 21
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 6.3.3 Recon 3

- Series description: PET WB Corr TrueX
- Recon range (bed): 1 to 7
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 2
- Subsets: 24
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 2.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 6.3.4 Recon 4

- Series description: PET WB UDEN True X + TOF
- Recon range (bed): 1 to 7
- Output image type: Corrected

- Recon method: Iterative3D
- Iterations: 3Subsets: 24
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mmOffset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off

#### 6.3.5 Recon 5

- Series description: PET WB no AC
- Recon range (bed): 1 to 7
- Output image type: Uncorrected
- Recon method: Iterative3D
- Iterations: 3
- Subsets: 24
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: Off ()
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off