PETCT_WholeBody_TEST.MlAdult_PET6

27. februar 2013

Indhold

1 Topogram

1.1 Routine

- mA: 36

- kV: 140

- Topogram length: 1536 mm

- Tube position: Bottom

1.2 Scan

- mA: 36

- kV: 140

- Delay: 4s

- Topogram length: 1536 mm

- Direction: Caudocranial

- Tube position: Bottom

- API: None

- Kernel: 80s shar

- Window: Spine

2 CT WB

2.1 Routine

- Eff. mAs: 170

- kV: 100

- CARE Dose4D: On

- CareDoseType: CareDoseAEC

- CTDlvol: 6.65856mGy

- Scan time: 17.580 s

- Delay: 5.000 s

- Slice: 8.00 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 0.0 grader

2.2   Scan

- Quality ref. mAs: 170

- Eff. mAs: 170

- kV: 100

- Scan time: 17.580 s

- Rotation time: 0.500 s

- Delay: 5.000 s

- Slice: 8.00 mm

- Pitch: 0.8500000

- Direction: Craniocaudal

2.3   Recons

2.3.1   Recon 1

- Series description: CT WB 1

- Slice: 8.00

- Kernel: B19f PET very smooth

- Window: Larynx

- Extended FoV: On

- FoV: 700

- Center X: 0

- Center Y: 0

- Mirroring: LeftRight

- Extended CT scale: Extended

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 8.000

- No. of images: 103

2.3.2   Recon 2

- Series description: CT WB 2

- Slice: 8.00

- Kernel: B30f medium smooth

- Window: Abdomen

- Extended FoV: Off

- FoV: 500

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 8.000

- No. of images: 103

3 Pause

4 PET WB

4.1 Routine

- Isotope: Cu-64

- Pharm.: Liquid Solution

- Inj. Dose: 10 Microcuries

- Scan mode: List mode

- Scan range: Match CT FOV

- No. of beds: 1

- Scan duration/bed: 3 Seconds

4.2 Scan

- Autoload: On

- Rebinner LUT: Off

- Scan output: List mode

- Sinogram mode: Trues

- Input trigger signal: ECG

- LLD (keV): 435

- ULD (keV): 650

4.3 Recons

4.3.1 Recon 1

- Series description: PET WB

- Recon range (bed): -1 to -1

- Output image type: NoRecon

- Recon method: Iterative

- Iterations: 4

- Subsets: 8

- Image size: 168

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 5

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.2   Recon 2

- Series description: PET WB Uncorrected

- Recon range (bed): -1 to -1

- Output image type: NoRecon

- Recon method: Iterative

- Iterations: 4

- Subsets: 8

- Image size: 168

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 5

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off