


T-DOC



GETINGE

vsázka

226758

Aut 4

Cykl 28937

program P01


GETINGE


2026.08.19 08:53:38

2026.08.19 09:58:57

systémová chyba

0






46-Series

DATUM : 2026.08.19
ZACATEK PROCESU : 08:53:38
SIGNALY
0 None
NAZEV PRISTROJE : Aut 4
CITAC CYKLU : 28937

PARAMETRY
CAS EXPOZICE : 00:07:00
TEPLOTA EXPOZICE : 134.0 C
PROGRAM : P01 Balené výrobky 134

CAS PROG P01		
00:00:00	75.2	0.951
START		
00:00:01	75.2	0.951
NEG PULZACE		
00:00:14	75.1	0.963
00:00:21	75.1	0.965
00:01:31	58.8	0.320
00:02:00	60.3	0.213
00:03:22	63.3	0.079
00:04:00	70.3	0.386
00:04:40	93.0	0.823
00:06:00	80.0	0.263
00:08:00	72.0	0.102
00:08:35	71.3	0.088
00:09:54	94.5	0.818
00:09:56	94.7	0.815
00:10:00	94.3	0.781
00:12:00	85.4	0.199
00:14:00	80.5	0.101
00:14:21	80.2	0.094
00:14:35	80.3	0.090
POZ PULZACE		
00:15:43	94.6	0.821
00:16:00	99.1	0.983
00:17:27	118.3	1.865
00:17:29	118.4	1.859
00:17:50	104.9	1.131
00:17:54	104.2	1.147
00:18:00	105.8	1.253
00:18:51	118.3	1.867
00:18:53	118.4	1.864
00:19:14	104.8	1.121
00:19:18	103.9	1.133
00:20:00	115.7	1.727
00:20:14	118.3	1.868
00:20:16	118.4	1.865
00:20:37	104.7	1.117
00:20:41	103.8	1.129
00:21:40	118.3	1.868
00:21:45	118.5	1.861
00:22:00	108.2	1.205



00:22:04	104.6	1.115
00:22:08	103.8	1.127
00:23:04	118.3	1.868
00:23:08	118.5	1.861
00:23:27	104.8	1.116
00:23:28	104.1	1.133

ZAHRIVANI

00:23:29	104.1	1.133
00:24:00	113.8	1.626
00:26:00	127.0	2.450
00:28:00	131.4	2.792
00:30:00	133.7	2.985

STERILIZACE

00:30:18	134.1	3.036
00:30:30	134.5	3.058
00:31:00	134.7	3.074
00:31:30	134.9	3.093
00:32:00	135.2	3.122
00:32:19	135.4	3.133
00:32:27	135.4	3.138
00:32:30	135.4	3.136
00:33:00	135.2	3.121
00:33:30	135.2	3.116
00:34:00	135.1	3.111
00:34:30	135.1	3.106
00:35:00	135.0	3.107
00:35:30	135.0	3.105
00:36:00	135.0	3.101
00:36:30	135.0	3.105
00:37:00	135.0	3.100

KONECNE UPRAVY


00:37:18	135.0	3.097
----------	-------	-------

SUSENI

00:37:25	134.9	3.091
00:38:00	125.2	2.224
00:38:54	90.1	0.618
00:40:00	100.0	0.270
00:40:38	100.9	0.199
00:42:00	97.0	0.120
00:44:00	93.6	0.068
00:44:27	93.4	0.060
00:46:00	94.3	0.053
00:47:27	94.1	0.049

SUSICI PAR PULZ

00:47:40	94.0	0.049
00:47:59	95.5	0.231
00:48:00	95.6	0.225
00:48:36	98.1	0.097
00:50:00	92.7	0.053
00:50:56	90.3	0.050
00:51:06	89.9	0.050
00:51:21	92.4	0.220
00:51:52	96.4	0.101
00:52:00	96.3	0.085
00:54:00	91.0	0.051
00:54:05	90.6	0.050
00:54:23	89.6	0.053



00:54:37	91.9	0.211
00:55:04	95.6	0.102
00:56:00	94.9	0.053
00:56:40	91.9	0.050
00:57:34	88.0	0.050
00:57:51	90.9	0.222
00:58:00	93.0	0.159
00:58:53	96.0	0.056
01:00:00	91.3	0.051
01:00:40	88.4	0.050
VYROVNAVANI		
01:00:50	87.6	0.051
01:02:00	84.8	0.523
PROCES UKONCEN		
01:03:43	81.7	0.931
01:04:00	81.3	0.930
01:05:19	80.2	0.948