

Certificate of Calibration

Instrument: Rapid Angle Measurement Kit
Model number: AFLS1
Serial Number: 012
Manufacturer: Pine Instrument Company

Calibration Date: 10/26/10
Recalibration Date: 10/26/11
Temperature(° C): 21
Rel. Hum.: 53%

	Run 1	Run 2	Run 3	Average
Angle (°)	1.24 °	1.24 °	1.24 °	1.24 ° +/-0.011°

The equipment used to perform this calibration/confirmation/verification is traceable to the National Institute of Standards and Technology (NIST).

The uncertainty of the Rapid Angle Measurement System is estimated at +/-0.011°

Calibration Standard: Angle Calibration Tube

Model Number: AFLSC01

Serial Number: 070101

Calibration Date: 10/26/2010

Recalibration Date: 10/26/2011

Standard Angle Value and Uncertainty: 1.234 ° +/- 0.011°

Tested By: Brent Booher

Approved By: 

Pine Instrument Company
101 Industrial Drive
Grove City, PA 16127