

## Certificate of Calibration

Certificate Number: 0907005G  
Calibration Interval: 12 months

Date of Calibration: October 29, 2010  
Next Calibration Due: October 29, 2011

---

**Instrument Under Test:**

Manufacturer: Busch

Model: B-40, Vacuum Gauge

Serial Number: 0907005G

**Standard Used:**

NIST Traceable, Busch, B-40

Serial Number: 0907044G

Next Calibration: 11/06/10

---

**Certificate Information:**

As Found: In Tolerance

Technician: MC

As Left: In Tolerance

Test Conditions:


Temperature  
23.9°C

Relative Humidity  
30%RH

Test Point	Standard	Reading	Units	Condition
1	35	35.2	TORR	In Tolerance
2	20	20.1	TORR	In Tolerance
3	5	5	TORR	In Tolerance

---

This Absolute Vacuum Gauge was calibrated against National Institute of Standard and Technology Traceable Instrumentation maintained by InstroTek®. The results contained herein relate only to the calibrated instrument. This certificate shall only be reproduced in full. Copies can be obtained by contacting InstroTek, Inc®.



Testing Inspector Signature

10/29/10

Date