Offset at 110°C

AASHTO MATERIALS REFERENCE LABORATORY 4441 Buckeystown Pike, Frederick, MD 21704

REPORT ON ASSESSMENT OF MATERIALS LABORATORY

Assessors Present	Assessor Roles (circle one)	Assessment #
1.	Normal Training Combo Cross-Training	Start Date
2.	Normal Training Combo Cross-Training	End Date
3.	Normal Training Combo Cross-Training	Old Report #
4.	Normal Training Combo Cross-Training	Old Re. Date
		Lab AMRL #

1	. Location of La	aboratory:					
	Phone No. (includ	le area code):					
	Name of Laborate	ory:					
	Address						
3.	<u>Distribution o</u>	f Report: <u>Original</u>				<u>Copy</u>	
	Title:		NIV				
	Email:						
		<u>Circ</u>	cle one:				
LAB	ACCREDITED?		Yes	Seek	ing No		
Сору	to FHWA (Main	State DOTs only)?	Yes	No			
Out-	of-Sequence Labo	ratory (OOS)?	Yes	No	Supplemental	_	
Ki	t Letter	Temp. Probe A	Temp. Probe B		Uncertainty		
Offse	et at 20°C				± 1°C		
Offse	et at 60°C				± 1°C		

 $\pm 1^{\circ}C$

AMRL PRE-ASSESSMENT CHECKLIST

	<u>nent</u> - Review		•			
New methods /	scopes reque	sted:				
Accredited met	hods not requ	uested:				
Currently susp	ended / revok	xed:				
Reason for susp	oension / revo	ocation:				
Aggregate asses	ssed by:	AMRL	CCRL	mixture	none	
Note: CCRL doe	rsn't offer C12	252 (uncompa	ucted voids) or any	aggregate "D" m	ethods except fo	or D2419 (Sand
required for thei	r scope of acc	reditation.	ment. Check that			
BAC MAR	PGB HVM	EML HMG	AGC SOL	AGF RVL	HMS CBR	HMI PNT
	PSP results. (Check for cur	rent low ratings (0,			
Notes: Pata Review current that are consiste	PSP results. Ontly below or	Check for cur consistently a	rent low ratings (0, above the average re	esult. Compare to	o laboratory's in	ternal audits.
Notes: Pata Review current	PSP results. Ontly below or	Check for cur consistently a	rent low ratings (0,	esult. Compare to		ternal audits.
Notes: Pata Review current that are consiste	PSP results. Ontly below or	Check for cur consistently a	rent low ratings (0, above the average re	esult. Compare to	o laboratory's in	ternal audits.
Notes: Pata Review current that are consiste	PSP results. Ontly below or	Check for cur consistently a	rent low ratings (0, above the average re	esult. Compare to	o laboratory's in	ternal audits.
Notes: Pata Review current that are consiste	PSP results. Ontly below or	Check for cur consistently a	rent low ratings (0, above the average re	esult. Compare to	o laboratory's in	ternal audits.
Notes: Pata Review current that are consiste Test Method	PSP results. Ontly below or	Check for cur consistently a	rent low ratings (0, above the average re	esult. Compare to	o laboratory's in	ternal audits.
Notes: Pata Review current that are consiste Test Method vious Assessmen	PSP results. On Sample Sample Report	Check for cur consistently a	rent low ratings (0, above the average real strings	Com	o laboratory's in	ternal audits.
Notes:	PSP results. On the below or sample state	Check for cur consistently a e # F	rent low ratings (0, above the average real real real real real real real rea	Com D3740	E329	ds
Notes:P Data Review current that are consiste Test Method	PSP results. On the below or service of the service	Check for cur consistently a le # F	D3666 applicable). Have a	D3740 copy available to	E329 o reference while	ds e writing the
Notes:	PSP results. On the below or service of the service	Check for cur consistently a le # F	D3666 applicable). Have a	Com D3740	E329 o reference while	ds
Notes:	PSP results. On the property of the property o	Check for cur consistently a le # F	D3666 applicable). Have a	D3740 copy available to	E329 o reference while	ds e writing the

AMRL LABORATORY ASSESSMENT OPENING MEETING

Date:	
sessor: Upon arrival at the laboratory, arrange and conduct an opening meeting with the contact person person in charge of the facility as well as any other key staff that may be involve in the assessme item from the checklist below with the laboratory:	
Purpose of Assessment and Accreditation Criteria The laboratory is being assessed to evaluate the competency to perform test conformance to AASHTO, ASTM, and other governing standards as requellaboratory. Laboratories seeking AASHTO accreditation will additionally be conformance to the criteria specified in AASHTO R18, the AASHTO Accreditation Requirements Manual, and additional quality system standards (C1077, D3666, It as requested by the lab.	ested by the evaluated for tion Program
Scope of the Assessment Review the assessment request form and (1) verify that the laboratory is in agreer planned scope of testing, and (2) if accredited, that the scope requested is in agreer current accreditation scope. Note any tests to be added or dropped. Discuss any cope of the assessment and how it will affect the laboratory's accreditation status.	reement with
ethods Added:	
ethods Dropped:	
Assessment Schedule - Discuss/review the following as it relates to the schedule: Length of Assessment Quality Manual and Records Review schedule Special Circumstances – preference of scope order, field testing require requirements, etc. Approximate close-out meeting time 	ements, staff
Confidentiality Explain that all information obtained during the assessment will be held in confidence to be shared with other parties.	ence and will
Ask that a member of the laboratory staff who was present at the opening meeti worksheets, indicating that this material was discussed.	ng initial the
Review with (Laboratory Representative):	
Name (Print): Initials:	

Assessor Notes- for additional ASTM quality standards, the following criteria apply:

- C1077 demonstrate ASTM methods C117, C127, C128, and C136 during the AMRL or CCRL on-site assessment.
 - D3666 demonstrate at least one ASTM Hot Mix test method during the on-site assessment.
- **D3740** demonstrate at least 5 ASTM Soil tests (**Note:** any soil tests we assess for are acceptable, except for ASTM D2419).
- **E329** lab must meet the requirements of ASTM C1077, D3666, or D3740. (**Note:** A laboratory may cover ASTM C1077 through CCRL.) Alternatively, they can become accredited for E329 for SFRM testing only, if SFRM testing is performed during the on-site assessment.

AMRL LABORATORY ASSESSMENT CLOSING MEETING

				Date:			
			a preliminary report and that things may change on the final report. 90 calendar days of the date of issuance of the final report.	Explain that findings must			
	Explain	Nonconfo	rmities, Observations, and Informational as described on page 1 of the	report.			
	Distribute and review the preliminary report. Record names of all in attendance on assessment worksheets.						
	prelimi	nary report.	AP system for responding to findings – make sure you provide a copy. Give an explanation of how to resolve deficiencies. (Must present in ents, completed records, etc. Purchase orders will not suffice.)				
			pecifier role (DOT's, FHWA, multiple branch labs, other specifiers) the information, access to assessment reports, PSP results, etc, with specifiers.				
			personnel that they can provide feedback by visiting the "About Us" ality and Information Manager.	page on the website or by			
			bility of AASHTO accreditation for ISO/IEC 17025, General Requirements Testing Laboratories. An additional visit by an AMRL 17025 audito				
	Mention	n that anyo	ne can sign up to receive our newsletter by sending an email to subscr	ibe@amrl.net.			
	Explain other pa		formation obtained during the assessment will be held in confidence	and will not be shared with			
	Thank t	he laborato	bry for participating in the AMRL program(s).				
	Confirm	n email add	ress for final report.				
			WORKSHEET & TEST METHOD SUGGESTIONS				
			ave for changes to the worksheets in the table below. Suggestions for change the sele: Agg-11 T85 suggestion) Please discuss anything you write here w				
Set	/ p#	Test	Suggestions				
Fill out i	if handin	g suggestic	ons to APS:				
Assesso	r Name:		Report Number:				

Date: _____

PERSONNEL

1.	Laboratory workers involved in the assessment:	
Name:		Position:
		Position:
Name: .	(C/AF	Position: Position: Position: Position:
		Position:
		Position:
For Co	mbo labs, please record which assessor observed w	
Assesso	r 2 Name:	Observed:

C	\mathbf{R}^{\prime}	N.	\mathbf{E}^{*}	R	Δ	T	

HOURS

Please record the actual number of hours for each activity indicated below. Do not include lunch breaks or similar breaks if not auditing. For combo labs, please account accurately for each assessor's time. Do not record hours for trainees or cross-trainees.

- Auditing Hours time spent on testing, measuring equipment, and meetings (do not include QS Review or Reporting hours in Auditing Hours)
- Quality System Review Hours time spent reviewing quality system paperwork
- Reporting Hours time spent writing the report in the field

	Auditing Hours	Reporting Hours	QS Review Hours
Day 1			
Day 2			
Day 3			
Day 4			
Day 5			
Total Hours			

MECHANICAL SIEVING DEVICES *

Manufacturer	ID#	Standarization Time	Set Time	Elapsed Run Time	Condition OK?
					11

^{*} Only record devices that you have seen in operation

OVENS

			Temperature Range (°C)			110 ± 5°C
<u>Manufacturer</u>	Model No.	Serial No.	Min.	Max.	Ave.	$(230 \pm 9 \text{F})$?

Note: If the average temperature is not within the allowable range specified in the method (110 \pm 5°C), or if the min/max temperature recorded is beyond the allowable limits specified in the method by more than (min/max outside of 110 \pm 10°C), report the deficiency below. The temperature of the oven should be recorded for a minimum of 10 minutes.

LITERATURE

Soil					
AASHTO books		ASTM books			
AASHTO	Date	ASTM	Date		
R58-11		D421-85			
T88-13		D422-63			
T89-13		D4318-10			
T90-00		D4318-10			
T99-10		D698-12			
T100-06		D854-10			
T134-05		D558-11			
T135-13		D559-03	*		
T136-13		D560-03	*		
T146-96					
T176-08		D2419-09			
T180-10		D1557-12			
T190-13		D2844-13			
T191-13		D1556-07			
T193-13		D1883-07			
T208-10		D2166-13			
T215-70	*	D2434-68			
T216-07		D2435-11			
T217-13	2	D4944-11			
T224-10		D4718-87			
T236-08		D3080-11			
T265-12		D2216-10			
T267-86		D2974-13	THE RESERVE		
T288-12					
T289-91					
T296-10		D2850-03a			
T297-94	*	D4767-11			
T310-13		D6938-10			
T311-00					
		D1140-00			
		D2487-11			
		D2488-09a			
		D2937-10			
		D4546-08			
		D4643-08			
		D4644-08			
		D4829-11			
		D4943-08			
		D4972-13			
		D5084-10			
		D5731-08			
		D6913-04			
	ethod present	D7012-13	<u> </u>		

Date:	_

Aggregate							
AASHTO books		ASTM books					
AASHTO	Date	ASTM	Date				
T2-91		D75-13					
T11-05		C117-13					
T19-09		C29-09					
T21-05		C40-11					
T27-11		C136-06					
T37-07		D546-10					
T84-13		C128-12					
T85-13		C127-12					
T96-02	*	C131-06					
T104-99		C88-13					
T112-00		C142-10					
T113-06		C123-12					
T176-08		D2419-09					
T210-10		D3744-11a					
T248-11		C702-11					
T255-00		C566-13					
T304-11		C1252-06					
T327-12		D6928-10					
T335-09		D5821-13					
67)	C535-12					
A	~~~~~	D4791-10					
	Q	D7172-06					
Trans. See Se		D7370-09					
		D7428-08					

Quality Systems					
AASHTO Date ASTM Date					
R18-10		C1077-14			
		D3666-13			
		D3740-12a			
		E329-13c			

Non AASHTO/ASTM Standards			
FM 5-515		(Available free	
Printed 2000		online)	

* red standards are no longer printed in current books

Note: For each Test Method presented, look for the most current edition. If the test method was not demonstrated, mark it with a line. **COMMENTS:**

Date: _____

LITERATURE

Hot Mix Asphalt					
AASHTO books		ASTM books			
AASHTO	Date	ASTM	Date		
R47-08					
R59-11		D1856-09			
T30-13		D5444-08			
T110-03		D1461-11			
T164-13		D2172-11			
T166-13		D2726-13			
T167-10		D1074-09			
T209-12		D2041-11			
T245-13		D6926-10			
T245-13		D6927-06			
T246-10		D1560-09a			
T247-10		D1561-13			
T269-11		D3203-11			
T275-07		D1188-07			
T283-07		D4867-09			
T287-06		D4125-10			
T305-09		D6390-11			
T308-10		D6307-10			
T312-12	1	D6925-09			
T324-11			\y		
T329-13			<u> </u>		
T331-13		D6752-11			
		D1075-11			
\	-47	D2950-11			
		D4013-09			
		D5404-12			
		D6931-12			

Emulsions				
AASHTO books	AASHTO books ASTM books			
AASHTO	Date	ASTM	Date	
T59-13		D6929-10		
T59-13		D6930-10		
T59-13		D6933-13		
T59-13		D6934-08		
T59-13		D6935-11		
T59-13		D6936-09		
T59-13		D6937-08		
T59-13		D6997-12		
T59-13		D6998-11		
T59-13		D7402-09		
T59-13		D7496-11		
		D7000-11		

COM	MEN	ΓS:
-----	-----	-----

Asphalt Cement			
AASHTO books		ASTM books	
AASHTO	Date	ASTM	Date
R28-12		D6521-13	
R29-08			
T44-13		D2042-09	
T48-06	*	D92-12a	
T49-07	*	D5-13	
T50-09	*	D139-12	
T51-09	*	D113-07	
T53-09	*	D36-09	
T55-02	*	D95-13	
T78-10		D402-08	
T79-12		D3143-13	
T179-05		D1754-09	
T201-10	*	D2170-10	
T202-10	*	D2171-10	
T228-09	*	D70-09	
T240-13		D2872-12	
T295-13		D3142-11	
T300-11			
T301-13		D6084-06	
T313-12		D6648-08	
T314-12	7	D6723-12	
T315-12		D7175-08	
T316-13		D4402-12	
		D243-08	
		D3289-08	
		D5801-95	
		D7405-10a	

Quality Systems			
AASHTO books			
AASHTO	Date	ASTM	Date
R18-10		C1077-14	
		D3666-13	
		D3740-12a	
		E329-13c	

D7553-10

Sprayed	Sprayed Fire-Resistive Materials				
		E605-93			
		E736-00			

METALS LITERATURE

Ensure the laboratory also has the **specification** in the shaded box above each section.

Test Name	AASHTO	Date	ASTM	Date
Deformed and Plain Billet-Steel Bars			A615-13	Butte
Low-Alloy Steel, Deformed and Plain Bars			A706-09b	
CA Testing Mechanical and Welded Splices			CT670-11	
Zinc (Hot-Dip Galvanized) Coatings on Iron and Steel	M111-11		A123-13	
Thickness of Zinc Coating	T65-13		A90-11	
Steel Welded Wire, Plain	M55-09		A1064-13	
Steel Welded Wire, Deformed	M221-09		A1064-13	
Bend Test				
Weld Shear				
Tensile Strength	T244-13		A370-12a	
High-Strength Bolts for Structural Steel Joints	none		A325-10	
Rotational Capacity				
Brinell Hardness	none		E10-12	
Rockwell Hardness	none		E18-14	
Tensile Strength	T244-13		F606-13	
Proof Load Determination	T244-13		F606-13	
Steel Strand, Uncoated 7-Wire	M203-12		A416-12a	
Tensile Strength	T244-13		A1061-09	
Gray Iron Castings	M105-09		A48-03	
Tensile Strength	none		E8-13	
Steel Wire, Deformed	M225-09		A1064-13	
Steel Wire, Plain	M32-09		A1064-13	
Bend Test				
Tensile Strength	T244-13		A370-12a	
Deformed and Plain Billet-Steel Bars	M31-10		A615-13	
Low-Alloy Steel, Deformed and Plain Bars	none		A706-09b	
Rail-Steel and Axle-Steel Deformed Bars	M322-10		A996-14	
Testing Mechanical Splices			A1034-10a	11
Tensile Strength	T244-13		A370-12a	
Bend Test	T285-89		E290-13	
Structural Steel	M270-12		A709-13a	
Tension Test	none		A6-13a	
Bend Test	none		A6-13a	
Charpy V-Notch	T266-08*	*	E23-12c	
Zinc Coated Steel Wire Rope & Fittings for Guardrail	M30-02		A741-11	
Ductility (Wrap Test) / Adherence of Coating				
Tensile Strength	T244-13		A370-12a	
Mass of Zinc Coating	T65-13		A90-13	
Epoxy Coated Reinforcing Bars	M284-09 *	*	A775-07b	
 Coating Flexibility (Bend Test) 				
Film Thickness			G12-07	
Continuity of Coating (Holidays)			G62-07	
Headed Steel Bars			A970-13a	
Bend Test				
Tensile Strength			A370-12a	

^{^ =} reprint of ASTM version except for small changes

A6, A48, A90, A123, A325, A370, A416, A615, A709, A741, A775, A970, A996, A1061, A1064, E8, E10, E18, E23, E290, F606, G12, G62.

M30, M31, M32, M55, M105, M111, M203, M221, M225, M270, M284*, M322, T65, T244, T266*, T285.

Pipe: D2412, D2444, D2680, D3034, M252, M264, M278, M294, F405, F2306

Other standards we reference: D88, D2168, D4753, E1, E4, E11, E77, E100, M92, M231, M320, PP57, R30, T72, T224, TP71

SAMPLE SIZE REDUCERS (T248/C702)

Date:		
Date:		

1. Jones or Flat Riffle Samplers:

Manufacturer (Record)			
Chutes:			
(a) Coarse: 8 or more? (record number)			
Fine: 12 or more? (record number)			
(b) Equal size openings?			
(c) Minimum distance between dividers?			
(d) Adjacent chutes discharge oppositely?			
(d) Dividers in good condition?			
Miscellaneous:			
(a) Splitter level?			
(b) Number of splits per run is two?			
Feeder: (Only required for AASHTO)			
(a) Feeder width equal to or slightly less (1/4 in.) than total chute width?			
(b) Edge of feeder straight?			
Discharge Pans:			
(a) Length equals or exceeds total chute width?		_	
(b) Pans in good condition?			

2. Miscellaneous Splitters

Manufacturer	Type	Nom. Opening	Nom. Particle Size	Condition OK?

3.	Quartering Equipment:
	(a) (Optional) Quartering cloth, 6 x 8 ft. (2 x 2.5 m)?
	(b) Straightedge scoop?
	(c) Shovel or trowel?
	(d) Brush or broom?
4.	Miniature Stockpile Sampling:
	(a) Sample thief, small scoop, or small spoon?

SIEVES (as specified in M92/E11)

Date:		
Date.		

	Set	Metric (mm)	Size (in.) or Number	Min. opening size (mm)	Max. opening size (mm)	No. Inspected	No. OK	Remarks
		75	3	72.8	78.1			
	S	63	2 1/2	61.1	65.7			
1	S	50	2	48.5	52.3			
1	2 S	37.5	1 1/2	36.4	39.3			
	S	31.5	1 1/4	30.6	33.1			
1	S	25.0	1	24.2	26.4			
1	2 S	19.0	3/4	18.4	20.1			
	S	16.0	5/8	15.5	17.0			
	S	12.5	1/2	12.12	13.33			
1	2 S	9.5	3/8	9.21	10.18			
	S	8.0	5/16	7.75	8.60			
		6.3	1/4 (No. 3)	6.10	6.81			
1	2 S	4.75	No. 4	4.60	5.16			
	S	4.00	5	3.87	4.37			
	2 S	2.36	8	2.284	2.61			
1		2.00	10	1.935	2.23			
		1.7	12	1.644	1.90			
	2 S	1.18	16	1.140	1.34			
1		0.850	20	0.821	0.977			
	2 S	0.600	30	0.579	0.701			
1		0.425	40	0.410	0.506	1		
	2 S	0.300	50	0.289	0.365			
1		0.250	60	0.240	0.308			
	2 S	0.150	100	0.143	0.193			11
1		0.106	140	0.101	0.141			
1	2	0.075	200	0.071	0.104			
					Pans OK?		Covers OK?	

Note: Each sieve shall have a label marked with the following information: (1) Test Sieve; (2) ASTM E-11; (3) Standard sieve designation; (4) Name of manufacturer or distributor; (5) Alternative sieve designation (optional); and (6) a unique serial number permanently engraved or etched onto the sieve frame, skirt, or nameplate.

Note: 1 – T88/D422 Hydrometer, sieve set 1 [ASTM only: 1 1/2 in., 3/4 in., No. 20, No. 60, and No. 140 sieves are required.]

2 – T88/D422 Hydrometer, sieve set 2

S - T104/C88 Sulfate Soundness sieve set

THERMOMETERS

Date: _____

General Condition of Thermometers (Reference ASTM E77 and E1)

A CITIM				Imm. Line		#		#
ASTM Liquid in glass	IP	Group	Range	T=total	Divisions	insp.	Serial Number(s)	OK
7F		2	30 to 580°F	Т	2°F	•	. ,	
7C	5C	2	-2 to 300°C	Т	1°C			
8F		2	30 to 760°F	Т	2°F			
8C	6C	2	-2 to 400°C	Т	1°C			
9F	15F	2	20 to 230°F	57 mm	1°F			
9C	15C	2	-5 to 110°C	57 mm	0.5°C			
11F	28F	2	20 to 760°F	25 mm	5°F			
11C	28C	2	-6 to 400°C	25 mm	2°C			
12F	64F	2	-5 to 215°F	T	0.5°F			
12C	64C	2	-20 to 102°C	T	0.2°C			
13C	47C	2	155 to 170°C	T	0.5°C			
15F		2	30 to 180°F	T	0.5°F			
15C	60C	2	-2 to 80°C	T	0.2°C			
16F		2	85 to 392°F	T	1°F			
16C	61C	2	30 to 200°C	T	0.5°C			
17F		2	66 to 80°F	T	0.2°F			
17C		2	19 to 27°C	T	0.1°C			
19F		2	120 to 134°F	T	0.2°F			
19C		2	49 to 57°C	T	0.1°C			
20F		2	134 to 148°F	T	0.2°F			
20C		2	57 to 65°C	T	0.1°C			
47F	35F	1	137.5 to 142.5°F	T	0.1°F			
47C	35C	1	58.6 to 61.4°C	T	0.05°C			
62C	19	/ 1 //	-38 to 2°C	T	0.1°C			
63F		1//	18 to 89°F	T	0.2°F	1		
63C	Pri	1	-8 to 32°C	T	0.1°C	- 3		
64F		1	77 to 131°F	T	0.2°F			
64C		1/1	25 to 55°C	T	0.1°C			
110F		1	272.5 to 277.5°F	T	0.1°F			
110C	93C	1	133.6 to 136.4°C	T	0.05°C			
			0		0			
digital	liquid	in glass						
			0		0			_
digital	liquid	in glass						
			0		0			+
digital	liquid	in glass						
	<u> </u>		0		0			
digital	liquid	in glass						
			0		0			+
digital	liquid	in glass						

Note to Assessors: Total immersion thermometers (T) must be submerged up to the top of the liquid (mercury, etc) column to get an accurate reading. Partial immersion thermometers must be submerged to the immersion line (ex: 25 mm). For write-in thermometers please mark the type of thermometer.

Group 1 – May be read to fractions of a division, often by using a magnifying aid. Without magnification, consider as Group 2. **Group 2** – May be read to nearest half division.

GENERAL PURPOSE BALANCES (M231/D4753)

_		
Date:		
Date.		

Before testing the balance, check the levelness of the balance (most have a built-in bubble level) and clean off the balance pan.

Class	Read	lability / Se	ability / Sensitivity			Accuracy*			
G1 (ASTM: 0	GP1)	0.01 g				0.02 g or 0.1 %			
G2 (ASTM: 0	GP2)	0.1 g				0.2	g or 0.	1 %	
G5 (ASTM:	GP5)	1 g				2 g	or 0.1	%	
G20 (ASTM:	GP10)	5 g				5 g	or 0.1	%	
G100		20 g				20	g or 0.	1%	
ASTM GP1	00	50 g				50	g or 0.	1%	
		_					1		1
Serial #:	1	5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g	
Mfgr:	Туре:	2 kg		100 g		5 g		0.2 g	
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g	
	•						•		
Serial #:	/	5 kg	Į	500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg	9	200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g	Z	5 g		0.2 g	
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g	
								11	
Serial #:		5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g	
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g	
Serial #:		5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g	
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g	

^{*}Accuracy equal to the mass stated or 0.1 percent of the test load, whichever is greater, throughout the range of use.

^{*}The accuracy requirements shall be met for application of a test load on any point on the pan. **COMMENTS:**

Sensitivity OK?

GENERAL PURPOSE BALANCES (M231/D4753)

_		
Date:		
Date.		

Before testing the balance, check the levelness of the balance (most have a built-in bubble level) and clean off the balance pan.

Class	Rea	Readability / Sensitivity			Accuracy*				
G1 (ASTM: C	GP1)	0.01 g	0.01 g			0.02	0.02 g or 0.1 %		
G2 (ASTM: C	GP2)	0.1 g			0.2 g or 0.1 %				
G5 (ASTM: C	GP5)	1 g				2 g	or 0.1	%	
G20 (ASTM: C	GP10)	5 g				5 g	or 0.1	%	
G100		20 g				20 ;	g or 0.	1%	
ASTM GP1	00	50 g				50	g or 0.	1%	
							•		
Serial #:		5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g	
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g	
Serial #:	/	5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg	9	200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g	V	5 g		0.2 g	
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g	
								11	
Serial #:		5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g	
Sensitivity OK? Accuracy OK?		1 kg		50 g		2 g		0.1 g	
Serial #:		5 kg		500 g		20 g		1 g	
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g	
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g	
G 44 44 OFF								0.4	

^{*}Accuracy equal to the mass stated or 0.1 percent of the test load, whichever is greater, throughout the range of use.

1 kg

50 g

2 g

Accuracy OK?

0.1 g

^{*}The accuracy requirements shall be met for application of a test load on any point on the pan. COMMENTS:

GENERAL PURPOSE BALANCES (M231/D4753)

_		
Date:		
Date.		

Before testing the balance, check the levelness of the balance (most have a built-in bubble level) and clean off the balance pan.

Class	Read	Readability / Sensitivity				Accuracy*				
G1 (ASTM: 0	GP1)	0.01 g				0.02 g or 0.1 %				
G2 (ASTM: 0	GP2)	0.1 g			0.2 g or 0.1 %					
G5 (ASTM:	GP5)	1 g			2 g or 0.1 %					
G20 (ASTM:	GP10)	5 g			5 g or 0.1 %					
G100		20 g			20 g or 0.1%					
ASTM GP1	00	50 g			50 g or 0.1%					
G			<u> </u>	- 0.0		•••	<u> </u>		<u> </u>	
Serial #:		5 kg		500 g		20 g		1 g		
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g		
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g		
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g		
Serial #:		5 kg		500 g		20 g		1 g		
Class G/GP:	Capacity:	3 kg	9	200 g		10 g		0.5 g		
Mfgr:	Type:	2 kg		100 g	Z	5 g		0.2 g		
Sensitivity OK?	Accuracy OK?	1 kg	92000	50 g		2 g		0.1 g		
								11		
Serial #:		5 kg		500 g		20 g		1 g		
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g		
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g		
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g		
Serial #:		5 kg		500 g		20 g		1 g		
Class G/GP:	Capacity:	3 kg		200 g		10 g		0.5 g		
Mfgr:	Type:	2 kg		100 g		5 g		0.2 g		
Sensitivity OK?	Accuracy OK?	1 kg		50 g		2 g		0.1 g		

^{*}Accuracy equal to the mass stated or 0.1 percent of the test load, whichever is greater, throughout the range of use.

^{*}The accuracy requirements shall be met for application of a test load on any point on the pan. **COMMENTS:**