

# Certificate of Calibration



**MASTER SETTING RING**

**089**

QA/05/03.01

Identification :

Z-8682

Date :

20.12.02

Specified Dimensions	Observed Dimensions
<u>5.75000 in</u>	<u>5.75015 in</u>

The above gauges are calibrated at 20°C under standards room conditions. Uncertainty of measurement is 0.001 mm. Reference standards traceable to International Standards vide Certificate No. E01110355/C27/16 traceable to NPL UK valid upto February 2005.

CALIBRATED BY



APPROVED BY