Table 1: Reconstruction fidelity across interpolation factors and the three study cohorts. The layout follows a dataset-grouped comparison rather than a method leaderboard. No method is designated as "ours," and best-value typography is intentionally omitted because reconstruction fidelity is a context variable for the reliability experiments, not the paper's claimed contribution.

		fastMRI brain T2				IXI brain T2				NingXia knee T2			
Factor	Method	MAE ↓	PSNR ↑	SSIM ↑	NCC ↑	MAE ↓	PSNR ↑	SSIM ↑	NCC ↑	MAE ↓	PSNR ↑	SSIM ↑	NCC ↑
×2	Linear	98.66	16.77	.7707	.7816	—	—	—	—	136.78	18.12	.7611	.7697
	Cubic	114.51	15.63	.7432	.7414	—	—	—	—	153.41	17.29	.7423	.7412
	ACSI-Net	—	—	—	—	—	—	—	—	—	—	—	—
	ACVTT	97.73	16.94	.7794	.7911	—	—	—	—	131.78	18.45	.7711	.7850
	ArSSR	80.03	18.19	.8342	.8506	—	—	—	—	133.56	18.28	.7662	.7773
	CycleINR	80.25	18.17	.8346	.8496	—	—	—	—	133.56	18.28	.7662	.7773
	HAT	68.03	19.59	.8909	.8940	—	—	—	—	122.16	18.91	.7963	.8082
	I3Net	45.61	22.57	.9459	.9473	—	—	—	—	124.09	18.82	.7860	.8046
	Pix2PixGAN	56.64	20.91	.9204	.9220	—	—	—	—	129.67	18.35	.7744	.7855
	Res-SRDiff	89.32	17.49	.8203	.8272	—	—	—	—	479.32	10.40	.5459	.6322
	ResUNet	72.73	19.12	.8785	.8814	—	—	—	—	132.33	18.40	.7697	.7828
	SAINT	72.08	19.31	.8857	.8874	—	—	—	—	130.89	18.48	.7742	.7876
	SMamba-UNet	67.94	19.52	.8887	.8916	—	—	—	—	129.36	18.57	.7764	.7906
	SwinIR	65.69	19.78	.8948	.8978	—	—	—	—	127.41	18.67	.7829	.7961
×4	Linear	134.54	14.24	.5884	.5953	65.99	20.34	.9003	.9030	175.93	16.14	.6350	.6408
	Cubic	153.24	13.21	.5443	.5411	69.87	19.99	.8984	.8977	197.41	15.20	.5998	.5968
	ACSI-Net	—	—	—	—	47.13	23.12	.9480	.9482	—	—	—	—
	ACVTT	122.23	15.38	.6676	.6886	—	—	—	—	162.77	16.85	.6640	.6832
	ArSSR	114.21	15.38	.6659	.6888	—	—	—	—	165.77	16.51	.6509	.6655
	CycleINR	114.39	15.38	.6659	.6888	—	—	—	—	172.33	16.30	.6417	.6501
	HAT	86.09	17.80	.8277	.8327	50.18	22.51	.9414	.9413	149.41	17.25	.6961	.7153
	I3Net	70.87	19.03	.8714	.8757	39.26	25.03	.9645	.9644	162.62	16.80	.6332	.6836
	Pix2PixGAN	92.41	16.95	.8076	.8067	—	—	—	—	183.70	15.73	.6094	.6177
	Res-SRDiff	95.98	16.75	.7781	.7865	—	—	—	—	225.44	14.93	.5519	.5661
	ResUNet	92.68	17.27	.8014	.8087	—	—	—	—	164.52	16.77	.6620	.6785
	SAINT	80.14	18.30	.8477	.8511	—	—	—	—	154.55	17.15	.6904	.7069
	SMamba-UNet	85.88	17.68	.8194	.8267	—	—	—	—	150.46	17.22	.6868	.7091
	SwinIR	90.37	17.37	.8028	.8121	50.55	22.46	.9406	.9404	148.66	17.27	.6975	.7152
×6	Linear	148.66	13.42	.5042	.5098	—	—	—	—	201.72	15.02	.5343	.5378
	Cubic	148.66	13.42	.5042	.5098	—	—	—	—	204.51	14.92	.5298	.5329
	ACSI-Net	—	—	—	—	—	—	—	—	—	—	—	—
	ACVTT	124.63	15.00	.6264	.6485	—	—	—	—	185.36	15.89	.5626	.5881
	ArSSR	128.03	14.51	.5814	.6057	—	—	—	—	186.41	15.57	.5523	.5704
	CycleINR	128.25	14.47	.5774	.6033	—	—	—	—	198.16	15.15	.5400	.5466
	HAT	96.75	16.82	.7727	.7826	—	—	—	—	172.24	16.23	.5903	.6229
	I3Net	81.61	17.82	.8256	.8333	—	—	—	—	188.46	15.65	.5102	.5723
	Pix2PixGAN	95.61	16.61	.7913	.7911	—	—	—	—	218.87	14.43	.4596	.4733
	Res-SRDiff	102.11	16.22	.7655	.7655	—	—	—	—	254.89	13.91	.4379	.4685
	ResUNet	92.93	17.14	.7940	.8013	—	—	—	—	180.15	16.01	.5812	.6030
	SAINT	90.87	17.35	.8071	.8116	—	—	—	—	186.94	15.89	.5670	.5960
	SMamba-UNet	97.94	16.86	.7741	.7849	—	—	—	—	176.48	16.17	.5814	.6161
	SwinIR	97.73	16.83	.7725	.7835	—	—	—	—	172.19	16.20	.5853	.6202

Values are means on the common generated-plane foreground protocol: 24 fastMRI examinations, 115 IXI examinations, and 22 out-of-fold NingXia series from eight patients. IXI has a complete recovered audit only at ×4 for the listed available methods; "—" denotes an unavailable or non-auditable dataset–factor–method combination, not a zero result. Intensities differ across cohorts, so MAE is interpreted within cohort; no values are pooled across datasets.