Cost and Safety of Label Acquisition Strategies for MALDI-TOF Antimicrobial Resistance Prediction

Dyen Asif
Department of Computer Science
National University of Computer and Emerging Sciences
Islamabad, Pakistan
i240608@isb.nu.edu.pk

Hira Mastoor
Department of Computer Science
National University of Computer and Emerging Sciences
Islamabad, Pakistan
hira.mastoor@nu.edu.pk

Maryam Shahbaz
Department of Computer Science
National University of Computer and Emerging Sciences
Islamabad, Pakistan
maryam.shahbaz@nu.edu.pk

***Abstract*—**Machine learning (ML) applied to matrix-assisted laser desorption ionization time-of-flight (MALDI-TOF) mass spectra can predict antimicrobial resistance (AMR) in hours rather than days. A deployed model decays as the local population changes, so it must be retrained, and every label is a laboratory antimicrobial susceptibility test (AST) that someone pays for. Active learning (AL) is the usual answer, yet a 2026 review of 115 ML studies on clinical MALDI-TOF spectra reports none that treat label acquisition as a cost. We ran 1540 retraining simulations on the Database of Resistance Information on Antimicrobials and MALDI-TOF Mass Spectra (DRIAMS), covering three species and antibiotic pairs, four hospitals, five budgets and ten seeds. A model trained on 2015 and 2016 buys labels during 2017 and is tested on 2018. Across 165 paired comparisons, no uncertainty, diversity or hybrid rule beat random selection after correction for multiple testing; two were worse. Little accuracy is available to compete for: the area under the receiver operating characteristic curve (AUROC) between never retraining and retraining on every isolate spans only 0.044 to 0.051. The budget buys cost rather than accuracy. Four hundred tests, 19 to 31 percent of a year of isolates, recover 50 to 77 percent of that gain. Uncertainty sampling is the weakest rule by AUROC but cuts resistant isolates that go untested and are reported susceptible by 33 to 73 percent. SHapley Additive exPlanations (SHAP) indicate why the rules tie: models trained under different rules use the same spectral bins.

***Keywords*—**antimicrobial resistance, active learning, MALDI-TOF mass spectrometry, model maintenance, explainable machine learning.

I. Introduction

Antimicrobial resistance is among the largest threats to public health, and an estimated 1.27 million deaths were attributable to bacterial AMR in 2019 [1]. Treatment depends on knowing which drugs still work, which in routine practice means culture based AST. That takes up to 72 hours. MALDI-TOF mass spectrometry is already present in most clinical microbiology laboratories for species identification, and the same spectrum acquired for that purpose can be reused to predict resistance with ML, which reduces the delay from days to hours [2].

Such a model does not stay accurate on its own. The population of isolates in a hospital changes, the instrument is serviced or replaced, and the resistance mechanisms in circulation shift. In our own data a model trained on isolates from 2015 and 2016 reaches an AUROC of 0.881 under cross validation inside those years for *Escherichia coli* with ceftriaxone, but only 0.824 when it is applied to isolates from 2018. Keeping the model useful therefore requires periodic retraining, and retraining requires labels. A label here is not an annotation that a person can supply in a few seconds. It is a susceptibility test with reagent and technician costs, and a laboratory that is willing to fund only a few hundred extra tests in a year has to decide which isolates receive them.

This is the setting that AL was designed for [3]. The model ranks the isolates that arrive and asks for labels only where they are expected to be most informative, usually near the decision boundary or spread across the feature space. AL is well studied in medical imaging and in text, and it is routinely reported to reach a target accuracy with a fraction of the labels. Whether it does so for MALDI-TOF resistance prediction has not been established. A systematic review published in 2026, covering 115 studies that apply ML to clinical MALDI-TOF spectra, groups the field by task, preprocessing, model family and evaluation, and does not identify a single study in which label acquisition cost is a variable [4]. Prior work on this modality either assumes that all labels are already present or evaluates a fixed training set.

We examine that gap directly. The experiments use the Database of Resistance Information on Antimicrobials and MALDI-TOF Mass Spectra (DRIAMS), a public collection of clinical MALDI-TOF spectra with matched susceptibility results from four Swiss hospitals, released under a CC0 public domain waiver and available from Dryad at doi:10.5061/dryad.bzkh1899q. The protocol is temporal rather than a random split. A model is trained on isolates from 2015 and 2016, isolates arriving during 2017 are offered to an acquisition rule in four quarterly rounds under a fixed budget, and the maintained model is evaluated on isolates from 2018. This mirrors how a laboratory would actually operate the system, and it keeps the test year strictly after every label that was bought.

The contributions of this paper are as follows. We provide the first evaluation of label acquisition cost for MALDI-TOF resistance prediction, covering three species and antibiotic pairs, four hospitals and five budgets, with ten seeds per configuration. We report a negative result that holds across 165 paired comparisons, namely that no acquisition rule outperforms random selection, and we explain it by measuring how little accuracy retraining makes available in the first place. We reframe the useful quantity as cost, and quantify how much of the benefit of retraining a partial test budget recovers. We separate model quality next year from patient exposure this year, and show that the two disagree about which rule is best. Finally, we use TreeSHAP to show that models trained under different acquisition rules rely on the same spectral regions, which accounts for the absence of a difference between them.

979-8-3195-2818-6/26/$31.00 ©2026 IEEE

TABLE I
Position of this study relative to representative prior work. Explainable artificial intelligence (XAI) here means any reported feature attribution.

Study	*Task and data*	*Sites*	*Temp.*	*Cost*	*Errors*	*XAI*
Weis *et al.* [2]	AMR from MALDI-TOF, DRIAMS	4	no	no	no	no
Weis *et al.* [5]	Topological and kernel methods	1	no	no	no	no
Duroux *et al.* [6]	Generalisation to unseen sites	4	yes	no	no	no
Corvelo Benz *et al.* [7]	Conformal guarantees, DRIAMS	4	no	no	partly	no
Schmidt-Santiago *et al.* [4]	Review of 115 MALDI-TOF ML studies	n/a	rare	none	rare	21/115
Chen *et al.* [8]	AL with selective prediction, images	n/a	no	yes	no	no
Chen *et al.* [9]	Continuous learning, malware	n/a	yes	yes	no	no
This work	AL for AMR from MALDI-TOF, DRIAMS	4	yes	yes	yes	yes

II. Related Work

Resistance prediction from MALDI-TOF spectra became a practical proposition with the release of DRIAMS, which pairs more than 300000 routine spectra with susceptibility results and established that gradient boosted trees and shallow networks reach clinically interesting discrimination on several species and antibiotic combinations [2]. Later work improved the representation with topological and kernel methods [5], extended prediction to several antibiotics at once, and added statistical guarantees through conformal prediction [7]. The systematic review of Schmidt-Santiago *et al.* covers 115 studies in this area and reports that SHAP values are used for interpretation in 21 of them, so model explanation is established practice on this modality, while temporal evaluation remains uncommon [4]. Generalisation across hospitals has been studied directly, and performance drops sharply when a model is moved to a site it was not trained on [6].

Active learning has a long history. Uncertainty sampling selects the instances closest to the decision boundary [10], core set and diversity methods instead cover the input space [11], and importance weighted and cost sensitive variants correct the bias that selective querying introduces [12], [13]. Hybrid rules combine the two signals [14]. The literature is not uniformly positive. Random selection is a stronger baseline than it is often given credit for, and injecting randomness into uncertainty sampling has been reported to improve its robustness [15]. In security, where models are retrained continuously against a moving target, label acquisition under budget has been studied in its own right [9], [16], and work bridging AL and selective prediction shows that the two interact [8]. These settings resemble ours in that labels are expensive and the data distribution moves, but none of them involves mass spectra or susceptibility testing.

Table I places the present study against this background. Prior work on MALDI-TOF resistance prediction has not treated the acquisition of labels as a cost, has largely not used temporal splits, and has rarely reported the two clinical error types separately. Work on AL under budget has not been applied to this modality. The combination of a temporal protocol, an explicit test budget, both error types, and an explanation of why the acquisition rules behave as they do is what this paper contributes.

III. Method

A. Problem setting

Let a spectrum be a vector x of 6000 binned intensities and let y be 1 if the isolate is resistant or intermediate to the antibiotic of interest and 0 if it is susceptible. A model f is trained on the isolates available at the start of the period and outputs an estimated probability of resistance. A threshold t turns that probability into a report: resistant when $f(x)$ is at least t, susceptible otherwise.

Isolates then arrive over a year in four batches. In each round the model scores every arriving isolate, an acquisition rule selects $B/4$ of them for laboratory testing, where B is the budget for the whole year, and the true labels of those isolates are revealed and added to the training set. The model and the threshold are refitted. Isolates that were not selected are reported using the model's own output, which is the operational situation that matters clinically, because those reports reach clinicians without a confirmatory test. Fig. 1 shows the whole procedure.




Fig. 1. Study pipeline. A classifier is trained on isolates from 2015 and 2016 and a decision threshold is fixed on held out data. Isolates arriving in 2017 are scored in four quarterly rounds. In each round the acquisition rule buys susceptibility tests for a quarter of the budget; those isolates return a true label and join the training set, while the remaining isolates are reported using the model's own call and contribute to the error counts. The maintained model is evaluated on isolates from 2018.

B. Acquisition rules

Four rules are compared. Random selection draws uniformly from the isolates that arrived in the round and serves as the baseline. Uncertainty sampling takes the isolates whose predicted probability is closest to the threshold t, measured as the absolute difference between the two. Diversity sampling applies a k-center greedy procedure in the spectral space so that the selected isolates are spread out rather than clustered. The hybrid rule splits the round budget evenly between the uncertainty and diversity criteria. Two bounds frame all of these. The lower bound never retrains at all, and the upper bound retrains on every isolate that arrived, which is what a laboratory would obtain if it tested everything.

C. Error types and reported quantities

Clinical microbiology distinguishes two kinds of mistake, and they are not interchangeable. A very major error (VME) is a resistant isolate reported as susceptible, which can send a patient home on a drug that will not work. A major error (ME) is a susceptible isolate reported as resistant, which wastes a broader spectrum agent and contributes to selection pressure. The definitions follow the International Organization for Standardization (ISO) standard 20776-2 [17]. Both are reported throughout, because a rule that lowers one while raising the other has not improved anything.

We separate two views that are often conflated. The predictive view is the usual one, namely how well next year's model discriminates, which we measure as AUROC on the 2018 isolates. The operational view counts what happened during the pool year itself: among the isolates that were not selected for testing, how many were resistant and were nonetheless reported susceptible. These are counts of patients affected, not rates, and the two views do not have to agree.

D. Explaining the fitted models

To understand why the rules behave similarly we compute SHAP values [18], which distribute a single prediction across the input features in the unique way that satisfies local accuracy, missingness and consistency. For tree ensembles these values have an exact polynomial time solution, so no sampling approximation is involved. We use the TreeSHAP implementation inside the gradient boosting library itself, which removes any possibility of a mismatch between the explained model and the deployed one. Six models are explained per pair: the two bounds, and random and uncertainty acquisition at budgets of 400 and 800. All six are explained on the same 2018 isolates, so any difference is a difference between models rather than between evaluation sets. Two models are then compared by the overlap of their twenty most used bins and by the rank correlation over all 6000 bins. Importance of a bin is its mean absolute SHAP value. Its direction is obtained by splitting the isolates at the median intensity of that bin and taking the difference in mean SHAP between the halves, because averaging signed values over all isolates simply follows the majority class, which here is susceptible.

IV. Experimental Setup

A. Data and splits

DRIAMS provides spectra preprocessed to 6000 bins of 3 daltons each, spanning 2000 to 20000 daltons. Labels follow the convention of the original release, in which resistant and intermediate are coded together as 1. We study three species and antibiotic pairs that have enough labelled isolates for a temporal analysis: *E. coli* with ceftriaxone, *Klebsiella pneumoniae* with ceftriaxone, and *Staphylococcus aureus* with oxacillin. At the primary site these give 1481, 812 and 1141 training isolates from 2015 and 2016, pools of 2094, 1273 and 1445 isolates arriving in 2017, and test sets of 1386, 775 and 1204 isolates from 2018, with resistance prevalence in the test year of 18.5, 13.9 and 14.6 percent. The released spectra carry no acquisition dates, so the four arrival batches are formed by a deterministic hash of the isolate code rather than by calendar quarter. The cross-site experiment moves a model trained at the primary site to each of the other three hospitals.

B. Model, protocol and statistics

The classifier is a gradient boosted tree ensemble [19], tuned once on the 2015 and 2016 data only and then held fixed, so that no information from the pool or test years enters the configuration. Twenty percent of the training years is held out to set the threshold t at the lowest value achieving 90 percent sensitivity to resistance, which reflects the clinical preference for avoiding very major errors. Budgets of 40, 100, 200, 400 and 800 labels are spread evenly over the four rounds. Every configuration is repeated with ten random seeds, and the cross-site configurations with five, giving 1540 retraining simulations in total.

Strategies are compared with random selection on matched seeds using the Wilcoxon signed rank test, and the resulting p values are corrected across the whole family of comparisons with the Holm procedure [20]. We report the corrected values only. The DRIAMS release used here is archived at doi:10.5061/dryad.bzkh1899q, and our code, the per isolate acquisition log and the per run result records are released with the paper.

V. Results

A. Retraining is necessary but its total benefit is small

Two bounds define the space in which any acquisition rule has to operate. A model that is never retrained reaches an AUROC on the 2018 isolates of 0.823, 0.742 and 0.799 for the three pairs. A model retrained on every isolate that arrived in 2017 reaches 0.867, 0.788 and 0.850. Retraining therefore helps, and the direction of the effect is consistent across all three pairs, which confirms the premise of the study. However, the size of the effect is modest. The difference between the two bounds is 0.044, 0.046 and 0.051 of AUROC. Every acquisition rule competes inside a window of about five points, and the standard error over ten seeds is of a similar order at the smaller budgets.




Fig. 2. Test AUROC on 2018 isolates against the number of susceptibility tests bought during 2017, for each species and antibiotic pair. Points are means over ten seeds and bars are standard errors. The upper dashed line is retraining on every arriving isolate and the lower dotted line is never retraining. All four rules lie inside a narrow band and their intervals overlap at almost every budget.

B. No acquisition rule outperforms random selection

Fig. 2 shows the label efficiency curves. All rules improve with budget, and all of them remain between the two bounds, but they are not separable from one another. The paired tests make this precise. In the temporal setting there are 45 comparisons against random selection, given by three pairs, five budgets and three competing rules. None of them favours a competing rule once the Holm correction is applied. Two favour random: uncertainty sampling on *K. pneumoniae* at 800 labels is worse by 0.029 AUROC, and diversity sampling on *S. aureus* at 40 labels is worse by 0.008. The cross-site experiment adds 120 further comparisons across the three remaining hospitals, and none of those reaches significance in either direction. Taken together, 165 paired comparisons produce no evidence that an informed rule helps, and two instances in which it hurts.

This is not an artefact of one dataset or one organism. The finding holds over three species, two antibiotic classes, four hospitals and two deployment settings, including the cross-site case in which the new data differs most from the training data and in which AL would be expected to help most.

C. What a partial test budget buys

TABLE II
Share of the available AUROC gain recovered at each budget, averaged over ten seeds. A value of 1.00 would match retraining on every isolate and 0.00 would match never retraining. The antibiotic is ceftriaxone for the first two pairs and oxacillin for the third.

Pair	*Tests bought*	*Of pool*	*Tests saved*	*Random*	*Uncert.*	*Divers.*	*Hybrid*
E. coli	400	2094	80.9%	0.50	0.43	0.54	0.38
	800	2094	61.8%	0.66	0.52	0.70	0.90
K. pneumoniae	400	1273	68.6%	0.77	0.28	0.65	0.60
	800	1273	37.2%	0.66	0.02	0.82	0.67
S. aureus	400	1445	72.3%	0.60	0.38	0.41	0.57
	800	1445	44.6%	0.77	0.52	0.65	0.85

If the rules are interchangeable, the question worth asking is not which rule to use but how many tests to buy. Table II expresses the result as a price. Testing every isolate that arrives in a year costs 2094, 1273 and 1445 susceptibility tests at this hospital. Four hundred tests, which is between 19 and 31 percent of those totals, recovers about half of the available gain for *E. coli* and roughly three quarters for *K. pneumoniae*. The curve of benefit against labels is steep at the start and flat later. Most of the value therefore lies in the decision to retrain at all, not in the decision about which isolates to retrain on.

D. The rules differ in who is exposed, not in accuracy




Fig. 3. Resistant isolates that were not selected for testing and would have been reported susceptible, summed over the pool year and averaged over ten seeds. Hatched bars are a budget of 400 tests and solid bars 800. Uncertainty sampling is consistently lower than random selection even though it is the weaker rule by AUROC.

AUROC describes next year's model. It says nothing about the patients whose isolates were reported during the year in which the labels were being collected. Fig. 3 counts those cases. At a budget of 800 tests, the number of resistant isolates that went untested and would have been reported susceptible falls from 49.8 to 33.6 for *E. coli* under uncertainty sampling rather than random selection, from 19.4 to 5.2 for *K. pneumoniae*, and from 18.8 to 9.2 for *S. aureus*. These are reductions of 33, 73 and 51 percent. Major errors do not rise to compensate.

The mechanism is visible in the acquisition log, which records one line per isolate per round. Contrary to what might be expected, uncertainty sampling buys fewer resistant isolates than random selection does. At a budget of 400 the resistant share of what it bought was 17.2, 10.2 and 12.8 percent against arrival rates of 22.2, 14.2 and 18.8 percent. An isolate that the model confidently calls resistant lies far from the threshold, so the rule declines to pay for it. The isolates it does buy sit two to four times closer to the threshold than those it skips. It is therefore not a resistance detector. It finds the cases the model was about to get

wrong, which is a different and more useful thing, and it does so while spending less of the budget on cases the model already handles correctly.

E. Why the rules agree

TABLE III
Overlap of the twenty most used spectral bins between fitted models, as a Jaccard index. A value of 1.00 means identical lists and 0.00 means no bin in common.

Models compared	*E. coli*	*K. pneu.*	*S. aureus*
Random 400 vs uncertainty 400	0.60	0.38	0.54
Random 800 vs uncertainty 800	0.60	0.29	0.43
Random 800 vs retrain on all	0.67	0.38	0.60
Uncertainty 800 vs retrain on all	0.60	0.38	0.43
No retraining vs retrain on all	0.43	0.11	0.29
Mean over all fifteen model pairs	0.60	0.31	0.49

Table III compares the fitted models through their SHAP profiles. Models that differ only in how their labels were chosen agree with each other about as much as either agrees with the model retrained on everything. The comparison that stands out is the last one before the mean. The model that was never retrained is the outlier in every pair, and on *K. pneumoniae* it shares only two of twenty bins with the fully retrained model. Retraining changes what a model looks at; changing the acquisition rule at a fixed amount of retraining largely does not.

Counting how many of the six models place a given bin in their top twenty makes the structure clearer. For *E. coli*, 9 bins appear in all six lists out of 32 that appear anywhere, and the strongest lies at 2150 to 2156 daltons. For *S. aureus* the figure is 8 of 40, at 2405 to 2411 and 2852 to 2858 daltons. For *K. pneumoniae* it is only 2 of 57, which matches that pair having the fewest isolates, the lowest AUROC and the least stable explanations. A small shared core is found by every model, and the tail shifts with the training set. Because all six models are fitted to the same underlying relationship and the easily located part of it is located by all of them, similar accuracy is the expected outcome. We do not assign these masses to particular proteins. The low mass region is crowded, several known markers fall within a few daltons of one another, and a 3 dalton bin cannot separate them.

VI. DISCUSSION

The natural reading of a negative result is that the method was implemented badly. We think the better reading here is that label efficiency is the wrong objective for this problem. AL promises to reach a given accuracy with fewer labels. That promise is only valuable when the accuracy at stake is large. On this modality the entire gap between never retraining and retraining on everything is around five points of AUROC, and the training years already contain thousands of examples of the few spectral peaks that carry the signal. One more year of isolates adds mostly repetition. Choosing which repetitions to buy is a small lever on a problem whose bottleneck is elsewhere. The SHAP analysis supports this reading, since the models converge on the same bins regardless of how their labels were selected.

That does not make the acquisition rule irrelevant, but it moves the reason for choosing one. The rules differ in which patients are exposed to an error while the labels are being gathered, and that difference is large where the accuracy difference is not. A laboratory choosing a policy should therefore decide whether it is optimising next year's model or this year's reports. Our results suggest that a budget of a few hundred tests, selected by uncertainty, is a reasonable default: it costs the least, it is no worse than random by AUROC in all but one configuration, and it substantially reduces the resistant isolates that would otherwise be reported susceptible.

Several limitations qualify these conclusions. The released spectra carry no acquisition dates, so the arrival batches within 2017 are a deterministic pseudo ordering rather than true calendar quarters; the year boundaries themselves are real. The three pairs studied are those with enough labelled isolates for a temporal analysis, and the conclusions may not transfer to rarer organisms or to antibiotics with very low resistance prevalence. *S. aureus* with oxacillin has almost no resistant isolates at two of the four hospitals, which limits the power of the cross-site comparisons for that pair. The explanations are computed from one seed per configuration, so some of the disagreement in the tail of the ranking is sampling noise. Finally, all of this is a retrospective simulation on stored spectra. A prospective study would have to contend with laboratory turnaround, with clinicians who may override the model, and with instrument changes that our protocol holds fixed.

VII. CONCLUSION

We evaluated label acquisition strategies for the maintenance of MALDI-TOF resistance prediction models under a temporal protocol on DRIAMS. Across 165 paired comparisons spanning three species and antibiotic pairs, four hospitals and five budgets, no uncertainty, diversity or hybrid rule outperformed random selection, and two performed worse. The explanation is that retraining makes only 0.044 to 0.051 of AUROC available in total, and that models trained under different rules rely on the same small set of spectral bins. The quantity worth reporting is cost rather than label efficiency: roughly a third of a year of susceptibility tests recovers between half and three quarters of the benefit of testing everything. Uncertainty sampling remains worth using, not because it produces a better model, but because it reduces by a third to three quarters the resistant isolates that are reported susceptible while the labels are being collected. Future work should test whether these conclusions hold prospectively, and whether acquisition rules that target the clinical error directly, rather than model uncertainty, do better on the operational view.

REFERENCES

1. Antimicrobial Resistance Collaborators, “Global burden of bacterial antimicrobial resistance in 2019: a systematic analysis,” *The Lancet*, vol. 399, no. 10325, pp. 629-655, 2022.
2. C. Weis *et al.*, “Direct antimicrobial resistance prediction from clinical MALDI-TOF mass spectra using machine learning,” *Nature Medicine*, vol. 28, pp. 164-174, 2022.
3. B. Settles, “Active learning literature survey,” Computer Sciences Technical Report 1648, University of Wisconsin-Madison, 2009.
4. L. Schmidt-Santiago, A. Guerrero-Lopez, C. Sevilla-Salcedo, D. Rodriguez-Temporal, B. Rodriguez-Sanchez, and V. Gomez-Verdejo, “A systematic review of machine learning on clinical MALDI-TOF MS,” *Briefings in Bioinformatics*, vol. 27, 2026, doi:10.1093/bib/bbag208.

5. C. Weis, M. Horn, B. Rieck, A. Cuenod, A. Egli, and K. Borgwardt, “Topological and kernel-based microbial phenotype prediction from MALDI-TOF mass spectra,” *Bioinformatics*, vol. 36, no. S1, pp. i30-i38, 2020.
6. D. Duroux, P. P. Meyer, G. Visona, and N. Beerenwinkel, “Generalizable machine learning models for rapid antimicrobial resistance prediction in unseen health care settings,” *GigaScience*, 2026, doi:10.1093/gigascience/giaf156.
7. N. Corvelo Benz, L. Miranda, D. Chen, J. Sattler, and K. Borgwardt, “Detecting antimicrobial resistance through MALDI-TOF mass spectrometry with statistical guarantees using conformal prediction,” in *Research in Computational Molecular Biology (RECOMB)*, Springer, 2025, doi:10.1007/978-3-031-90252-9_5.
8. J. Chen, J. Yoon, S. Ebrahimi, S. O. Arik, S. Jha, and T. Pfister, “ASPEST: bridging the gap between active learning and selective prediction,” *Transactions on Machine Learning Research*, 2024.
9. Y. Chen, Z. Ding, and D. Wagner, “Continuous learning for Android malware detection,” in *USENIX Security Symposium*, 2023.
10. D. D. Lewis and W. A. Gale, “A sequential algorithm for training text classifiers,” in *ACM SIGIR*, 1994, pp. 3-12.
11. O. Sener and S. Savarese, “Active learning for convolutional neural networks: a core set approach,” in *International Conference on Learning Representations (ICLR)*, 2018.
12. A. Beygelzimer, S. Dasgupta, and J. Langford, “Importance weighted active learning,” in *International Conference on Machine Learning (ICML)*, 2009, pp. 49-56.
13. A. Krishnamurthy, A. Agarwal, T.-K. Huang, H. Daumé III, and J. Langford, “Active learning for cost-sensitive classification,” *Journal of Machine Learning Research*, vol. 20, no. 65, pp. 1-50, 2019.
14. W.-N. Hsu and H.-T. Lin, “Active learning by learning,” in *AAAI Conference on Artificial Intelligence*, 2015, pp. 2659-2665.
15. P.-Y. Lu, C.-L. Li, and H.-T. Lin, “A more robust baseline for active learning by injecting randomness to uncertainty sampling,” in *AI and HCI Workshop at ICML*, 2023.
16. F. Barbero, F. Pendlebury, F. Pierazzi, and L. Cavallaro, “Transcending Transcend: revisiting malware classification in the presence of concept drift,” in *IEEE Symposium on Security and Privacy*, 2022, pp. 805-823.
17. International Organization for Standardization, *ISO 20776-2:2021, Susceptibility testing of infectious agents and evaluation of performance of antimicrobial susceptibility test devices*, 2021.
18. S. M. Lundberg *et al.*, “From local explanations to global understanding with explainable AI for trees,” *Nature Machine Intelligence*, vol. 2, pp. 56-67, 2020.
19. T. Chen and C. Guestrin, “XGBoost: a scalable tree boosting system,” in *ACM SIGKDD*, 2016, pp. 785-794.
20. S. Holm, “A simple sequentially rejective multiple test procedure,” *Scandinavian Journal of Statistics*, vol. 6, pp. 65-70, 1979.