

■ Executive Regulatory Readiness Report: MDR UE 2017/745

Software as a Medical Device (SaMD) Qualification & Classification

Software Name: Tuscany Triage System (CDSS)
Version: v0.6.2
Applicable Framework: Regulation (EU) 2017/745 on Medical Devices (MDR)
Classification Rule: Annex VIII, Rule 11 (Software as a Medical Device - SaMD)
Target Device Class: Class IIa

1. Regulatory Qualification

Under Article 2(1) of Regulation (EU) 2017/745, software intended to provide information used to take decisions for diagnostic or therapeutic purposes qualifies as Medical Device Software (MDSW) / Software as a Medical Device (SaMD). Because the system analyzes physiological parameters to determine urgency codes, it qualifies as a Medical Device.

2. Classification Analysis (Rule 11)

Base Rule: Software providing diagnostic/therapeutic decision support is classified as Class IIa.
Final Qualification: Class IIa Medical Device Software (Human-in-the-Loop prevents escalation to Class III).

3. Standards Alignment Roadmap

- **EN ISO 13485:** Quality management systems for medical devices.
- **IEC 62304:** Medical device software lifecycle processes (Class B safety).
- **EN ISO 14971:** Risk management application to medical devices.