



Anton Paar DMA 5000 M - Measurement Results:

Software version: 2.95.15060.147

DMA Density serial number: 82885250

Alcolyzer serial number: 82792480

Xsample 520 serial number: 82886370

Sample Information:

- ▶ Unique Sample Id: 22965
- ▶ Date: 8/10/2020
- ▶ Time: 2:11:40 PM
- ▶ Sample Name: 3_2043
- ▶ Method: Alcolyzer Wine-WINE EXT

Measurement Result:

- ▶ Alcohol (% v/v): 11.34 %v/v
- ▶ Alcohol Condition: valid
- ▶ Density: 0.99348 g/cm³
- ▶ Specific Gravity: 0.995268
- ▶ Ethanol OIML-ITS-90 (% v/v): 3.232 % v/v