



## Anton Paar DMA 5000 M - Measurement Results:

Software version: 2.95.15060.147  
DMA Density serial number: 82885250  
Alcolyzer serial number: 82792480  
Xsample 520 serial number: 82886370

### Sample Information:

- ▶ Unique Sample Id: 22956
- ▶ Date: 8/10/2020
- ▶ Time: 12:09:40 PM
- ▶ Sample Name: 12\_2044
- ▶ Method: Alcolyzer Wine-WINE EXT

### Measurement Result:

- ▶ Alcohol (% v/v): 13.94 %v/v
- ▶ Alcohol Condition: valid
- ▶ Density: 0.98762 g/cm<sup>3</sup>
- ▶ Specific Gravity: 0.989399
- ▶ Ethanol OIML-ITS-90 (% v/v): 7.638 % v/v