



## Anton Paar DMA 5000 M - Measurement Results:

Software version: 2.95.15060.147  
DMA Density serial number: 82885250  
Alcolyzer serial number: 82792480  
Xsample 520 serial number: 82886370

### Sample Information:

- ▶ Unique Sample Id: 22947
- ▶ Date: 8/10/2020
- ▶ Time: 11:47:19 AM
- ▶ Sample Name: 3\_1061
- ▶ Method: Alcohol Only

### Measurement Result:

- ▶ Alcohol (% v/v): 13.37 %v/v
- ▶ Spec. Gravity (not visc.-corr.): ---
- ▶ Density: Out of range g/cm<sup>3</sup>
- ▶ Alcohol Condition: valid