



## Anton Paar DMA 5000 M - Measurement Results:

Software version: 2.95.15060.147  
DMA Density serial number: 82885250  
Alcolyzer serial number: 82792480  
Xsample 520 serial number: 82886370

### Sample Information:

- ▶ Unique Sample Id: 22918
- ▶ Date: 8/8/2020
- ▶ Time: 5:20:21 PM
- ▶ Sample Name: 2\_2039t
- ▶ Method: Alcolyzer Wine-WINE EXT

### Measurement Result:

- ▶ Alcohol (% v/v): 13.79 %v/v
- ▶ Alcohol Condition: valid
- ▶ Density: 0.99860 g/cm<sup>3</sup>
- ▶ Specific Gravity: 1.000403
- ▶ Ethanol OIML-ITS-90 (% v/v): Out of range % v/v