



## Anton Paar DMA 5000 M - Measurement Results:

Software version: 2.95.15060.147

DMA Density serial number: 82885250

Alcolyzer serial number: 82792480

Xsample 520 serial number: 82886370

### Sample Information:

- ▶ Unique Sample Id: 22967
- ▶ Date: 8/10/2020
- ▶ Time: 2:22:00 PM
- ▶ Sample Name: 5\_9030
- ▶ Method: Alcolyzer Wine-WINE EXT

### Measurement Result:

- ▶ Alcohol (% v/v): Out of range %v/v
- ▶ Alcohol Condition: Error: High Value
- ▶ Density: 0.89638 g/cm<sup>3</sup>
- ▶ Specific Gravity: 0.897990
- ▶ Ethanol OIML-ITS-90 (% v/v): 65.539 % v/v