www.tntextbooks.in

Figure 12. 3 Glucometer

mg/dL. Most glucometers are based on the electrochemical technology or colour reflectance principle (Figure12.3).

Significance

1. Handy and portable
2. Immediate results, the results are displayed in approximately 40 seconds.
3. Requires no calculation
4. No training is required for operating the instrument

Normal Glucose Values in blood

Normal glucose value:	70 – 100 mg/dL
Random blood sugar:	80 – 120 mg/dL
Fasting blood sugar:	70 – 110 mg/dL
Post – prandial blood sugar:	70 – 140mg/dL

Autoanalyser

Autoanalyser is a computer controlled device. It is used for quick estimation of several bio-chemical parameters like glucose, urea, cholesterol, enzymes and other proteins present in body fluids. The analyser is fully integrated with all the

Figure 12. 4 Autoanalyser