Attach this Cover Page as the first page of your report PDF submission. Answer in your own words; keep each answer to the space provided.

A. Student and team particulars

Name / Student ID	LE YI / G2605168D
Undergrad University / Degree	Zhengzhou University — B.Eng. in Industrial Design
Partner's name and ID (if any)	None (individual project)
Tentative Plan after NTU Master	Work in medical-device / healthcare-AI product roles in Singapore or China, and keep developing the home-haemodialysis monitoring-and-control concept from my undergraduate thesis.

B. Links (leave blank if not applicable)

Your CV Online and Github page	https://github.com/Ean214
Online repository (e.g., Github code, data, tools)	https://github.com/Ean214/md6329-idh-uf-control
Demo or presentation video	
Overleaf / report link	
LecTrace script or trace link	https://github.com/Ean214/md6329-idh-uf-control/tree/main/lectrace

Most relevant software tool in project	PyTorch (from-scratch Transformer encoder); XGBoost for the tabular baseline; PyVene for the intervention probes
Your project extends Assignments 1 or 2?	Yes — Assignment 2. The Transformer encoder is adapted from my Assignment 2 implementation (sinusoidal positional encoding, key-padding masks, multi-head attention), and the interpretability probe follows its observation-versus-intervention lesson.

Describe the software bug in your deep learning program (e.g., Pytorch, Tensorflow or ConvNetJS Javascript etc) that cost you the most time. How did you find it, and how did you know it was fixed? Did AI coding agent help you to find bugs?
The most expensive bug was not a crash but a misleading measurement. I ran an ablation that zeroed a component of a clinical-sequence Transformer and found the output probability essentially unchanged (about 1e-6), which I first read as "this component does not matter". Most of a day went into that wrong conclusion. The metric was the problem: predictions had saturated near 0.9999, and there the sigmoid slope p(1-p) is about 5.7e-05, so any real change is compressed to nothing on the probability scale. I found it by re-running the same ablation on the logit scale and by manually bypassing the intervention library, to rule out a hook error. It was fixed once all ablation reporting moved to logit scale and I added a whole-test-set sanity check: zeroing the same component across the full test set moved accuracy from 1.0000 to 0.5000 and changed the mean logit by 9.65. An AI coding agent did help — it recognised the saturation pattern and suggested the logit-scale check — but it had initially agreed with my wrong reading, which is why the manual cross-check and the whole-set check mattered more than the explanation. In this project the same tool was useful again when a pandas group-by over 160,000 dialysis sessions repeatedly died; it proposed a vectorised rewrite that cut the runtime of four dataset builds to about nineteen minutes in total.

Which of the Lecture Videos (https://www.youtube.com/playlist?list=PLORG7zBQosxw) can be improved and why?
I would improve Lecture 6 (EHR Part 2: RNN/LSTM, around 02:14), where the vanishing-gradient problem and the gating mechanism are explained verbally but without a worked numerical example. For a student coming from a product and clinical background — my own case — the gate arithmetic only became concrete when I implemented one LSTM cell myself. A short animated walk-through of a single time step (candidate state, forget and input gates, cell-state update) at the point where it is introduced would remove the main stumbling block. A smaller secondary point: Lecture 7 covers attention thoroughly (around 00:54), but the distinction between attention weights as an explanation and as an intervening variable only appears much later; raising it once in Lecture 7 would prevent a very common misreading, and it is exactly the confusion I made myself.

C. LecTrace — execution-explanation presentation

Approximate time spent learning LecTrace	About 2 hours — installation, one walk-through of the demo lecture, then writing and iterating on my own script.

LecTrace script written by yourself or mostly by AI?	Written by me; AI assistance was used for boilerplate and formatting. The content decisions — which execution states to show and why — are mine. Disclosed in the report's Software 3.0 section.

Q. Which execution states did you choose to show, and why those rather than others?

I chose five states, and each one is a place where a silent mistake would invalidate a headline number.
1. Session reconstruction: one patient's raw readings being cut into sessions at a six-hour gap. The unit runs three shifts and the last one crosses midnight, so keying sessions on the calendar date splits them in two and quietly destroys the weight fields. Showing the cut makes the pitfall visible.
2. The leakage boundary at T = 60 minutes: the readings that go into the model, and the later readings that are used only to build the label. If this boundary slips, every score becomes meaningless.
3. The patient-level split: the same patient ID never appearing in two splits. Rows, unlike patients, are not independent.
4. One decision point from the control simulation: the candidate ultrafiltration values, the predicted risk for each, and the chosen candidate together with the extra fluid and time it costs. This is the paper's central trade-off in one screen.
5. The ablation that surprised me: zeroing attention versus zeroing the feed-forward block in the Transformer, and the resulting logit change (about 1e-08 versus 0.042). It is the reason I do not present attention maps as explanations.
I chose these rather than the loss curve because the loss curve is the least informative part of the work: it looks healthy even when the evaluation is wrong.

Q. What was confusing or missing in LecTrace, and what one feature would you improve or add?

My main friction was that stepping through a long computation is only useful if the interesting state is visible, and it took me a while to learn which variables to mark for inspection. The viewer shows what you ask for, so the pedagogical work happens before the trace, not inside it. A smaller issue is reproducibility: with a stochastic training loop, re-running a step yields different numbers unless the seed is fixed explicitly, so the trace is not replayable by default.
The one feature I would add is an assertion step: let a traced value be compared against an expected value and fail loudly when they differ. My five states exist precisely to catch silent errors — the label window, the patient split, the chosen candidate — so having the tool assert them would turn a presentation into a regression test. That would have caught my saturation bug in seconds rather than a day.

By default, you consent to your anonymised answers being used in research on teaching. Let me know if otherwise. This choice doesn't affect your grade.

Sensor-Light Closed-Loop Ultrafiltration Control in Haemodialysis: An In-Silico Feasibility Study on Routine Machine Data

LE YI
Matriculation number: G2605168D
MSc in Artificial Intelligence in Medicine
Lee Kong Chian School of Medicine, Nanyang Technological University
`MD6329 Deep Learning for Healthcare AI`

Abstract

Intradialytic hypotension (IDH) is among the most frequent complications of haemodialysis. Clinical closed-loop systems that adjust the ultrafiltration rate (UFR) already exist, but they rely on optical relative-blood-volume sensors that a wearable device cannot easily carry. This project asks whether routinely recorded dialysis-machine data alone can support a predict-then-adjust loop, and what such a loop would cost. Using HEMOBP (CC BY 4.0; 1,072 patients recorded, 919 remaining after cleaning, 160,069 analysable sessions), we predict IDH in the remainder of a session from information available at a decision point T, with patient-level splits and AUPRC as the primary metric. Five models are compared. The best 60-minute model reaches AUPRC 0.2015 against a 0.0586 prevalence baseline, but patient-level bootstrap intervals are wide (95% CI 0.13–0.28) and the gap to the compact 1D-CNN+MLP is not established (difference 0.018, 95% CI -0.001 to 0.033), so we do not rank architectures. Raw probabilities are unusable as risks: they over-predict by a factor of four to six and every Brier score is worse than the no-skill value, which Platt scaling on validation repairs (0.052 versus 0.055). Modality ablation shows that intra-session blood pressure is the only indispensable modality (removing it costs 44% of AUPRC), while machine settings add no independent signal. In an in-silico simulation whose candidates use only decision-time information, holding the rate observed at T implies a predicted risk of 7.3%, and adapting down to about $0.75\times$ lowers it to 5.8% at the cost of 243 mL more unremoved fluid. The dynamics model is a conditional-risk model, not an identified counterfactual one, and candidate plans deviate from observed practice by 0.17 L/h; the control gains are therefore extrapolations, and validating the policy requires prospective data.

1 Problem Formulation

Haemodialysis replaces kidney function intermittently. Between sessions, patients accumulate fluid; during a four-hour session a machine must remove that fluid at a rate the cardiovascular system can tolerate. When the ultrafiltration rate (UFR) exceeds the capillary refill rate, effective circulating volume falls and intradialytic hypotension (IDH) follows. In our cohort, 11.2% of sessions reach a nadir systolic blood pressure (SBP) below 90 mmHg.

Three generations of countermeasures exist. First, reactive control: a proportional–integral–derivative (PID) controller, or nursing judgement, reacts only after blood pressure has already fallen.

39th Conference on Neural Information Processing Systems (NeurIPS 2025).

Second, blood-volume biofeedback: commercial systems such as Hemocontrol (Baxter) and the BVM/BTM-based feedback of Fresenius adjust UFR and dialysate conductivity in closed loop, guided by an optical sensor that measures relative blood volume [4, 11, 12]. Third, machine-learning prediction, which forecasts IDH from clinical variables but has rarely been evaluated as a component of a control loop [1, 10].

The second generation works, but it pays for its sensor: optical relative-blood-volume hardware is bulky and expensive, and a wearable artificial kidney cannot easily host it. This motivates the question we study here: *with routinely recorded machine data only — pump settings, dialysate parameters, and the sparse blood-pressure readings nurses already take — can a predict-then-adjust loop be built, and what does it cost?*

We decompose this into three testable sub-questions, each mapped to a claim from an accompanying engineering design project: (i) how early and how accurately can IDH be predicted (claim C1/C2); (ii) is multi-modal sensing necessary, and which modality matters (C3); (iii) how much IDH reduction does a predictive policy buy, and at what price in fluid removal and session time (C4/C5) — with a fourth, honest question about whether observational data can support causal claims at all (C6).

The work is deliberately positioned against the course material: irregular EHR time series (Lecture 5, ~00:38), recurrent models (Lecture 6, ~02:14), the Transformer (Lecture 7, ~00:27 and ~27:16), multi-modal fusion (Lecture 8, ~00:11), the reinforcement-learning view of sequential decisions (Lecture 10, ~00:27), interpretability (Lecture 11, ~05:11) and privacy in federated deployments (Lecture 12, ~00:39) [16]. Software 3.0 (generative AI) was used early here, to survey the closed-loop ultrafiltration literature and to test whether the design project's "prediction beats PID" framing was defensible. It was not (Section 2), and correcting that framing changed the research question rather than its wording.

2 Related Work

Closed-loop ultrafiltration. Santoro et al. introduced a multivariable closed-loop controller that adjusts UFR and dialysate conductivity to follow a target blood-volume trajectory, cutting severe hypotensive episodes from 26 to 3 in a crossover study [12]. Ronco et al. found fewer hypotensive episodes and better equilibrated Kt/V under biofeedback [11]. A recent narrative review located 28 clinical studies, of which 23 reported effectiveness and none reported adverse effects, while concluding that evidence for clinical adoption still requires randomised trials [4]. Our work does not compete with these systems; it asks what can be done when their sensor is unavailable.

IDH definitions and epidemiology. Prevalence depends heavily on definition [7], and different definitions carry different mortality associations [5]. Guidelines recommend keeping the ultrafiltration rate below roughly 13 mL/kg/h, alongside ultrafiltration profiling and cooler dialysate [3, 9]. We therefore report three definitions and justify the one we adopt.

Machine learning for IDH. Tree ensembles and recurrent networks dominate the recent literature; a bidirectional LSTM reached 78.5% accuracy and AUC 0.822 on 73,323 sessions [10], and a multi-scale temporal feature model reported 98.1% AUC with XGBoost on a public cohort [1]. A Temporal Fusion Transformer reached AUC 0.953 but at high compute cost. Two gaps motivate our design: prediction is rarely coupled to a control evaluation, and the causal status of the ultrafiltration signal is rarely examined. We address both, and we quantify the fluid-removal price of acting on a prediction. Our modelling choices also follow the course treatment of irregular clinical time series, recurrent gating, and attention over short sequences [16].

3 Data

Cohort. HEMOBP [8] contains dialysis-machine recordings linked to hospital demographic data, released under CC BY 4.0: 1,072 patients, 4,366,298 intra-session readings. Sessions are not a calendar day — the unit runs three shifts and the last crosses midnight — so we reconstruct sessions as contiguous runs of readings separated by more than six hours, and assign each to a dialysis day beginning at 08:00. This yields 197,269 sessions (82.7% match session-level weight records). Applying the cleaning rules below leaves **919 patients and 160,069 sessions**: 197,269 $\rightarrow$ 196,457




Figure 1: Cohort descriptives. Left: session duration. Right: mean UFR per session against the 13 mL/kg/h guideline threshold.

Table 1: IDH incidence under three definitions (session level, n=197,269).

Definition	Incidence	Use
nadir SBP $<$ 90 mmHg	11.24%	primary
nadir SBP $<$ 100 mmHg	20.73%	sensitivity
max $-$ nadir $\geq$ 20 mmHg	84.42%	rejected (non-specific)

(duration) $\rightarrow$ 162,563 (start weight) $\rightarrow$ 160,673 (interdialytic gain) $\rightarrow$ 160,069 (UFR). The weight filter is the largest single loss and removes 153 patients, so every modelling result describes 919 patients, not the 1,072 recorded.

Descriptive statistics. Mean session duration is 236 $\pm$ 24.5 minutes; mean UFR is 8.26 $\pm$ 3.38 mL/kg/h (median 7.8), i.e. mostly below the guideline threshold of 13; median interdialytic weight gain is 2.32 kg. Crucially, blood pressure is sparse: the median within-session interval between distinct readings is 50.7 minutes, and only 32% of intervals are at or below 20 minutes (Figure 1).

Outcome definition. Table 1 reports three candidate definitions. A drop of $\geq$20 mmHg applies to 84.4% of sessions and is therefore unusable as a target; we adopt nadir SBP $<$90 mmHg (11.24%) as primary and report nadir $<$100 mmHg (20.73%) for sensitivity.

Sampling design. For each session we form one sample per decision point $T \in \{30, 60, 90, 120\}$ minutes. Inputs use only readings with time $\leq T$; the label is the nadir SBP of readings after T. Sessions already below 90 mmHg before T are excluded, since the prediction would be moot; sessions with fewer than three pre-T readings are excluded. This yields 143,575 (T=30), 147,044 (T=60), 150,761 (T=90) and 151,182 (T=120) samples, with remaining-window IDH rates of 8.0%, 7.7%, 7.1% and 6.5% respectively. As a qualitative example, take one test session at T=60 min: 1.88 L of interdialytic gain, 0.46 L already removed, 1.42 L still to remove, 2.42 h remaining, and a clinician-chosen 0.48 L/h for the remainder. That plan removes 1.16 L and therefore leaves 266 mL unremoved, while SBP drifts from 148 to a nadir of 110 mmHg and no IDH event occurs. Sessions like this are where the label becomes noisy at the margin, and they are also where the fluid trade-off in E5 bites. Cleaning follows pre-registered bounds (duration 60–480 min, weight 20–200 kg, interdialytic gain -2 to 10 kg, UFR 0–20 mL/kg/h): 197,269 sessions reduce to 160,069 eligible.

Limitations of the data. HEMOBP is *in-centre* haemodialysis: four hours, supine, with sparse cuff measurements. It contains no photoplethysmography, electrocardiography, inertial or relative-blood-volume signals, and no symptom or nursing-intervention records. It therefore cannot validate a five-to-fifteen-minute warning window, and it cannot represent the ambulatory wearable scenario described in the accompanying design project. We return to this mismatch in the discussion.

4 Approach

Framing. The primary task is classification: given the state at T, will this session reach a nadir SBP below 90 mmHg afterwards? Trajectory prediction appears only *inside* the control simulation (Section 4), where a dynamics model must answer "what if the remaining ultrafiltration profile were different?" Reporting classification as the main result keeps the headline numbers interpretable and avoids over-claiming.

Features and modalities. Fifteen features fall into three groups, which we later ablate: (a) *demographic/static* — age, sex, diabetes, start weight, dry weight, interdialytic weight gain; (b) *pre-T physiology* — first SBP, nadir so far, SBP slope, count of distinct readings; (c) *machine settings* — mean UFR, blood flow, dialysate conductivity and temperature up to T. Sequence models additionally consume up to 96 readings of seven channels (SBP, DBP, UFR, blood flow, conductivity, dialysate temperature, normalised elapsed time) with an explicit padding mask.

Models. Five models are compared. (1) XGBoost [2] on tabular features is the clinical baseline. (2) A multilayer perceptron consumes static features only. (3) A 1D-CNN+MLP fusion mirrors the architecture of the accompanying design project: a convolutional branch over the reading sequence, an MLP branch over static features, concatenated before a shared head. (4) An LSTM [6] reads the masked sequence and takes the last valid hidden state. (5) A Transformer encoder built from scratch reuses the encoder written for the course assignment on clinical sequences [15]: sinusoidal positional encoding, multi-head scaled dot-product attention with key-padding masks, GELU feed-forward blocks and mean pooling over valid positions [13].

Evaluation protocol. Splits are by patient (70/15/15) with a fixed seed, because a patient contributes dozens of sessions and row-wise splitting would leak. AUPRC is the primary metric given class imbalance; AUROC, Brier score and sensitivity at 80% specificity are secondary. Because AUPRC is prevalence-sensitive, we report test prevalence alongside every value.

Training procedure. All four neural models share one recipe, so that E1 isolates architecture rather than tuning. Inputs are standardised with training-split statistics only: each sequence channel is centred and scaled over valid (non-pad) training entries and re-zeroed at padded positions, and the fifteen static features are standardised with training-split mean and standard deviation. Training minimises binary cross-entropy with `pos_weight` set to the training-split negative-to-positive ratio (17.1:1 at T=60); optimisation is Adam with learning rate 10^{-3} and batch size 256, for at most ten epochs, with early stopping of patience two epochs on validation AUPRC and restoration of the best-validation checkpoint before test evaluation. Dropout is 0.2 in the fusion and head layers and 0.1 inside the Transformer encoder, which uses d=64, four heads and two layers. Python, NumPy and PyTorch seeds are all fixed to 42. XGBoost uses the histogram tree method, 600 trees of depth 6, learning rate 0.05, 0.8 row and column subsampling, and `scale_pos_weight` equal to the same ratio. These defaults are revisited in E7.

Control policies (in silico). We train a *dynamics model* that adds two features to the tabular set: the mean UFR actually applied after T, and the remaining minutes. It is trained on clinicians' observed choices — which is precisely why it is a simulator, not an estimator of causal effect. On the test set we then compare three policies at the decision point: *constant* (keep the observed profile), *reactive* (a rule-based proxy for biofeedback: cut UFR by 30% if SBP has dropped $\geq$15 mmHg or nadir is below 100), and *predictive* (choose from a candidate grid of 0.6–$1.0\times$ the observed UFR, minimising predicted risk subject to a residual-fluid tolerance). For each policy we report mean predicted risk, mean unremoved fluid, and the extra minutes needed to still reach the prescribed fluid target.

Causal diagnosis. Two probes test whether the ultrafiltration signal is causally usable. First, a perturbation sweep: we vary the planned remaining UFR multiplier from 0.3 to 1.3 and record mean predicted risk, plus the crude observed association between actual UFR and IDH. Second, an intervention-style ablation on the Transformer: zeroing the attention output or the feed-forward output of each encoder block (the intervention tooling follows 14) and measuring the change on the logit scale (not the probability scale, which saturates). Figure 6 shows the perturbation and association results.

Table 2: E1: five models at T=60 min (test n=22,407, prevalence 5.86%). Intervals are patient-level bootstrap (B=1,000). Brier scores are raw and all exceed the no-skill 0.0552 (E8).

Model	AUPRC	95% CI	AUROC	Sens@80%spec	Brier	Params
XGBoost	0.1675	0.118–0.221	0.733	0.548	0.116	—
MLP	0.1721	0.107–0.257	0.758	0.562	0.150	10,369
1D-CNN+MLP	0.1837	0.117–0.268	0.768	0.573	0.136	24,961
LSTM	0.1804	0.119–0.247	0.761	0.569	0.175	32,129
Transformer	0.2015	0.131–0.282	0.774	0.587	0.164	107,265




Figure 2: E9: test AUPRC with patient-level 95% bootstrap intervals. The intervals overlap across all five models, so the architecture ordering in Table 2 is not established.

Software 3.0. Generative AI assistance was used throughout, under the course honour code, and is disclosed here. It was most valuable in three places: refactoring the session builder after pandas-groupby iteration failed on 160k sessions (a vectorised rewrite cut runtime from >20 min to 19 min for four horizons); diagnosing a misleading ablation result in an earlier assignment that turned out to be sigmoid saturation rather than a modelling failure; and localising a device-placement bug in a diagnostic model. It was least reliable on domain judgement — choosing the IDH definition, deciding what the control comparison may legitimately claim, and recognising that the biofeedback literature makes "predictive versus PID" the wrong framing. All numerical results in this report were produced by scripts in the accompanying repository and can be regenerated.

5 Experiments

E1: five-model comparison. Table 2 reports the 60-minute horizon on 22,407 test samples (prevalence 5.86%). The from-scratch Transformer has the highest point estimate, but patient-level bootstrap intervals are wide and overlap: it exceeds the 1D-CNN+MLP by 0.018 with a 95% interval of -0.001 to 0.033 (Figure 2). We therefore *do not* claim an architecture ranking; what the comparison supports is that a compact fused model is competitive at a quarter of the parameters. Two cautions follow. Validation AUPRC (0.30–0.35) exceeds test AUPRC because the best-validation checkpoint is chosen on validation, so that estimate is optimistically biased — a winner's-curse effect rather than distribution shift. And every Brier score exceeds the no-skill value $p(1-p) = 0.0552$, so they cannot be read as calibration quality; E8 diagnoses why and repairs it.

E2: compute budget. Table 3 shows why the fusion model is the reasonable deployment choice: 0.271 ms per sample on CPU and 3.31 ms for a batch of 256, against 1.487 ms and 147.2 ms for the Transformer — 44× faster at batch scale for 91% of the ranking quality. Latency was measured on a laptop (RTX 3070 Ti, CPU fallback) and is indicative, not an embedded-MCU benchmark. LSTM FLOPs are under-counted by the profiler because recurrent terms are not attributed; we do not compare on that column.

E3: prediction horizon. Table 4 shows a non-monotone pattern: quality peaks at T=90 min for the fusion model (AUPRC 0.2395, AUROC 0.807). Two effects trade off — later decision points have more observed history, but their remaining windows contain fewer events, so prevalence




Figure 3: E1: AUPRC by model against the no-skill prevalence line.

Table 3: E2: accuracy versus cost (T=60). Latency on CPU; FLOPs per sample.

Model	Params	FLOPs	Lat. bs1	Lat. bs256
XGBoost	600 trees	—	0.003 ms	—
1D-CNN+MLP	24,961	1.22 M	0.271 ms	3.31 ms
LSTM	32,129	0.06 M*	0.544 ms	9.85 ms
Transformer	107,265	23.69 M	1.487 ms	147.2 ms

falls from 6.21% to 4.75%. Cross-T comparisons therefore mix window and prevalence effects and must not be read as "later is better". The practically important point is the floor: with blood pressure sampled every 50.7 minutes at the median, prediction horizons below roughly one sampling interval cannot be validated on this data. A five-to-fifteen-minute warning window, as specified in the accompanying design project, requires continuous sensing — which is a direct argument for its photoplethysmography specification rather than a refutation of it.

E4: modality ablation. Table 5 answers the sensor question. Pre-T blood-pressure history alone nearly matches the full feature set (0.1592 versus 0.1675), and removing it costs 44% of AUPRC — the largest single loss. Machine settings alone are weak (0.0751), and removing them slightly *improves* the model (0.1740), indicating little independent signal at T=60; this is consistent with clinicians already reacting to blood pressure, so the UF setting partly encodes that reaction. Static demographics contribute a small but non-zero amount. Within this data, then, physiological sensing carries the signal and additional machine-side channels do not. We stop short of the stronger claim that adding a high-frequency sensor such as photoplethysmography would improve prediction: the current data contain no such signal, so that remains a hypothesis for the next study rather than a result here. Caveat: this ablation uses XGBoost; sequence models might extract more from the dense machine stream, which we leave as follow-up.

E5: the price of prediction. An earlier version of this experiment anchored its candidate actions on the mean UFR *actually delivered after* T — a quantity that does not exist at decision time — and its numbers are withdrawn. The revised design generates candidates only from what a controller knows at T: the mean rate delivered so far, anchor, with candidates $\{1.0, 0.9, 0.8, 0.7, 0.6\} \times \mathrm{anchor}$. Three arms share that anchor: *hold* (keep it), *reactive* (cut 30% if SBP has fallen $\geq$15 mmHg or nadir is below 100), and *predictive* (minimise predicted risk subject to a residual-fluid tolerance). The plan clinicians actually followed is a descriptive reference, not an arm, because it is not a choice available at T. Risks are Platt-recalibrated before any level is quoted.

At T=60 min and 0.5 L tolerance (Table 6), holding the observed rate leaves the least fluid (330 mL) but implies the highest predicted risk (7.3%). The reactive rule reduces that to 7.1% and leaves 428 mL. The predictive arm reaches 5.8% with 573 mL residual — numerically the same risk as the clinicians' own plan, which instead left 699 mL. We do not read that as the model beating clinical practice: the observed plan is in-distribution while the candidate arms are partly extrapolated.

Table 4: E3: horizon study. 1D-CNN+MLP AUPRC/AUROC by decision point.

T (min)	30	60	90	120
Test prevalence	6.21%	5.86%	5.20%	4.75%
XGBoost AUPRC	—	0.1675	0.1965	—
1D-CNN+MLP AUPRC	0.1963	0.1837	**0.2395**	0.2266
1D-CNN+MLP AUROC	0.757	0.768	0.807	**0.824**




Figure 4: Left: E3 horizon curve against the prevalence (no-skill) line. Right: E4 modality ablation.

Loosening the tolerance moves the predictive arm along the curve (5.9% at 0.25 L, 5.5% at 1.0 L), a monotone trade-off and the only feature of this table we would defend.

Two limits bound the interpretation. First, the dynamics model is a conditional-risk model $P(\text{IDH} \mid \text{state}, \text{plan})$ trained on clinicians' observed plans; nothing here identifies a counterfactual, so the risk differences are model extrapolations rather than causal effects. Second, the candidates sit 0.17 L/h from the mean observed plan — a 43% deviation — and although 89% of candidate values fall inside the observed support, that support is wide, so support alone does not establish credibility.

E6: is the ultrafiltration signal causally usable? Figure 6 reports both probes. The perturbation sweep is monotone and in the expected direction: raising the planned remaining UFR multiplier from 0.3 to 1.3 raises mean predicted risk from 0.183 to 0.265 (gradient $+0.089$). The crude association agrees: binned by actual post-decision UFR, observed IDH rises from 3.1% to 11.8%. This is necessary for the control simulation to make sense — but it is not sufficient to establish causality, because confounding by indication predicts the same pattern: patients who are more fluid-overloaded, and therefore more fragile, receive higher ultrafiltration and are also more likely to become hypotensive. The gradient's existence licenses the simulation; its magnitude cannot be decomposed into causal and confounded parts from this data.

A second probe is more surprising. Zeroing the attention output of each Transformer encoder block changes the logit by about 10^{-8} — effectively nothing — whereas zeroing the feed-forward output changes it by 0.042 and 0.067 at layers 0 and 1. The predictive signal therefore flows through the feed-forward path, not through attention, over a sequence of roughly 22 readings with sparse blood pressure. Two consequences follow: attention-heavy architectures are hard to justify here (reinforcing E1/E2), and attention maps must not be presented as clinical explanations — the same observation-versus-intervention lesson from our earlier coursework.

E7: hyperparameter selection. The brief asks for training and hyperparameter selection, so we swept the two choices with the largest effect. For XGBoost, tree depth dominates: AUPRC peaks at depth 3 (0.1904 with 600 trees) and falls monotonically to 0.1675 at depth 6 and 0.1520 at depth 9, while adding trees beyond 600 does not help (Figure 7). With roughly 100k training rows and a signal concentrated in a few physiological features, deeper trees overfit. For the 1D-CNN+MLP, the learning rate matters more than the class-weight correction: AUPRC rises from 0.1817 at 10^{-3} to 0.1926 at 3×10^{-4} and falls to 0.1696 at 3×10^{-3}, while disabling the positive-class weight costs about five relative percent. Retuning therefore changes the E1 ordering at the margin: XGBoost moves from 0.1675 to 0.1904 and the fusion model from 0.1837 to 0.1926, leaving the from-scratch

Table 5: E4: modality ablation (XGBoost AUPRC, T=60). "Only" uses one group; "w/o" drops one group.

Feature set	AUPRC
All 15 features	0.1675
Only pre-T physiology	0.1592
Only demographics/static	0.0821
Only machine settings	0.0751
Without pre-T physiology	0.0935
Without demographics/static	0.1513
Without machine settings	0.1740

Table 6: E5 (revised): decision-time arms at T=60 min, tolerance 0.5 L, with the observed plan as a descriptive reference. Risks are Platt-recalibrated, in silico, and conditional — not causal.

Arm	Predicted IDH risk	Residual fluid	Extra time
Observed plan (reference, not an arm)	0.058	699 mL	120 min
Hold the rate observed at T	0.073	330 mL	31 min
Reactive rule (biofeedback proxy)	0.071	428 mL	74 min
Predictive (model-chosen UFR)	0.058	573 mL	129 min

Transformer (0.2015) ahead by a much smaller margin than E1 suggests. We keep E1 on identical default settings because that comparison holds the pipeline constant, and report the tuned numbers as evidence that architecture ranking at this signal strength is sensitive to tuning — a caution for anyone deploying these models.

E8: are the probabilities usable as risks? No, not before recalibration. Table 2 already shows that every Brier score exceeds the no-skill value $p(1-p) = 0.0552$, and Figure 8 shows why: mean predicted probabilities are 0.24–0.35 against an observed 0.0586, an over-prediction of four to six times, with calibration slopes of 0.56–0.82 rather than 1. The cause is the training recipe, not the architecture: `scale_pos_weight` (and `pos_weight`) reweights the loss by 17:1, which shifts the effective prior, and no step maps the resulting scores back onto the observed event rate. Fitting a one-parameter Platt map on the validation split and applying it to test repairs the probabilities without touching the ranking: Brier falls to 0.0515–0.0525 for all five models, below the no-skill 0.0552, while AUPRC is unchanged by construction. Two consequences follow. First, "best Brier among five" in E1 does not mean well calibrated — it means least bad before repair. Second, every risk level quoted in E5 is recalibrated; raw scores would have overstated risk by roughly a factor of four, and a control policy optimising raw scores would be optimising a distorted objective.

E9: is the model ordering real? No. Bootstrap resampling by patient (B=1,000) gives 95% intervals of 0.118–0.221 (XGBoost), 0.117–0.268 (1D-CNN+MLP) and 0.131–0.282 (Transformer); the paired difference between the Transformer and the fusion model is 0.018 with an interval of -0.001 to 0.033, positive in 96.6% of resamples. A 96.6% one-sided figure is suggestive but the interval crosses zero, and after the hyperparameter sweep in E7 the gap narrows to 0.009. We therefore report the comparison as "competitive at a quarter of the parameters" rather than "better". Any product decision that hinges on the ranking would need either far more patients or an external cohort.

Software 3.0 in the experimental workflow. Two of the seven experiments exist because of an AI-assisted check rather than a planned one: the probability-versus-logit diagnosis that produced the whole-test-set sanity check, and the suggestion to test attention against feed-forward ablation once the attention heatmaps looked uniform. In the other direction, the first session builder proposed for this project died on 160k sessions; the vectorised rewrite was suggested, but confirming that no patient crossed a split still had to be done by hand. The recurring pattern was that assistance shortened the distance from a symptom to a candidate cause, and contributed nothing to deciding which claim the evidence licenses.




Figure 5: E5: the IDH-versus-fluid trade-off (left) and its sensitivity to the residual-fluid tolerance (right).




Figure 6: E6: left, predicted risk under UFR perturbation; right, crude observed UFR–IDH association. Both are consistent with a causal effect and with confounding.

6 Discussion and Conclusion

What we can now say. Three questions were asked at the outset. First, how early and how accurately: with routine machine data and sparse cuff pressures, IDH is predictable at a 60-minute horizon with AUPRC 0.2015 against 0.0586 prevalence (AUROC 0.774) — but with an interval of 0.13–0.28, and no established advantage over a compact fused model at a quarter of the parameters. Second, whether multi-modal sensing is necessary: only intra-session blood pressure is indispensable; machine settings contribute almost no independent signal at T=60. Third, what prediction buys and costs: within the simulation, moving from holding the observed rate to a model-chosen reduction lowers predicted risk from 7.3% to 5.8% while leaving 243 mL more fluid behind — a trade-off whose direction is robust but whose magnitude is not identified, because the model is conditional rather than counterfactual.

Boundaries we will not overstep. Every control number here is in silico under a model trained on clinicians' own decisions; it is not clinical validation. The dose–response gradient is equally compatible with confounding by indication, so the E5 gains are upper bounds. The attention-ablation result means explanation should not be read off attention weights. And the cohort is in-centre haemodialysis with a cuff measurement every 50.7 minutes: it cannot speak to the ambulatory wearable scenario, which lacks exactly the high-frequency modalities (photoplethysmography, inertial, pressure waveforms) whose value this dataset cannot measure — the ablation says blood pressure history matters, not that a new sensor would help. Scaling from 4-hour supine sessions to 6–8-hour ambulatory therapy, and from sparse to continuous sensing, remains an extrapolation that this dataset cannot adjudicate.




Figure 7: E7: hyperparameter selection. Left: XGBoost AUPRC over tree depth and ensemble size. Right: 1D-CNN+MLP learning-rate sweep with and without positive-class weighting.




Figure 8: E8: reliability diagrams before (left) and after (right) Platt recalibration. Raw scores sit far above the diagonal; the one-parameter map restores them without changing the ranking.

Implications for the design project. The results strengthen three claims and weaken one. Strengthened: continuous physiological sensing deserves the sensor budget (E4); a compact convolutional-plus-MLP model is the right deployment shape (E1/E2); and a clinician-set safety envelope with a hard physical override is not bureaucratic decoration but a necessary condition, precisely because the causal strength of ultrafiltration cannot be established retrospectively (E6). Weakened: the design project's framing of its contribution as "prediction beats PID" is historically inaccurate — 28 clinical studies of closed-loop ultrafiltration biofeedback predate it [4]. The defensible framing is narrower and stronger: *replacing the optical relative-blood-volume sensor with routine machine data and inference*. The five-to-fifteen-minute warning specification also needs restating, since it is not verifiable on routine data and depends on the continuous sensing the device is meant to provide.

What the project taught me. The value of a healthcare-AI contribution sits less in the highest AUPRC than in knowing which claim the data can carry. The most useful result here is a negative one — that the control policy cannot be validated retrospectively — and it took a controlled model comparison, a modality ablation and an intervention probe to establish it rather than assert it.

Future work. Four directions follow directly. External validation on a second centre, since all results here come from one cohort. A properly validated calibration step deployed with the model: Platt scaling works but was fitted on a single split, so it needs external confirmation before any threshold is used clinically. A targeted-causal analysis — within-patient comparisons or instrumental variables — to bound the causal share of the ultrafiltration gradient. And a high-frequency simulation arm, calibrated to the distributions reported here, to test whether the ambulatory scenario preserves or reverses the conclusions.

Conclusion. Sensor-light closed-loop ultrafiltration control is buildable as software and evaluable offline, and its central trade-off — hypotension risk against fluid removal and session time — can be quantified. What it cannot be, on retrospective routine data, is causally validated. Naming that boundary is not a weakness of the study; it is the finding that should shape the next experiment.

References

[1] Amjad Al-Mousa et al. Intradialytic hypotension detection using explainable machine learning techniques aided with multi-scale temporal features. *Journal of Reliable Intelligent Environments (Springer)*, 2025. doi: 10.1007/s13721-025-00689-w.

[2] Tianqi Chen and Carlos Guestrin. XGBoost: A scalable tree boosting system. *Proceedings of the 22nd ACM SIGKDD International Conference on Knowledge Discovery and Data Mining*, pages 785–794, 2016.

[3] Chinese Society of Nephrology Expert Group. Expert consensus on the prevention and treatment of intradialytic hypotension. Chinese clinical guideline (published in Chinese), 2022.

[4] Zijun Dong, Lemuel Rivera Fuentes, Sharon Rao, and Peter Kotanko. Closed loop ultrafiltration feedback control in hemodialysis: A narrative review. *Toxins*, 16(8):351, 2024. doi: 10.3390/toxins16080351.

[5] Jennifer E. Flythe, Hui Xue, Katherine E. Lynch, Gary C. Curhan, and Steven M. Brunelli. Association of mortality risk with various definitions of intradialytic hypotension. *Journal of the American Society of Nephrology*, 26(3):724–734, 2015. doi: 10.1681/ASN.2014020222.

[6] Sepp Hochreiter and Jürgen Schmidhuber. Long short-term memory. *Neural Computation*, 9 (8):1735–1780, 1997.

[7] Johanna Kuipers et al. The prevalence of intradialytic hypotension in patients on conventional hemodialysis: a systematic review with meta-analysis. *American Journal of Nephrology*, 49 (6):497–506, 2019. doi: 10.1159/000500877.

[8] Yi-Chun Lin et al. Dataset supporting blood pressure prediction for the management of chronic hemodialysis. *Scientific Data*, 6:313, 2019. doi: 10.1038/s41597-019-0319-8.

[9] National Kidney Foundation. K/doqi clinical practice guidelines for cardiovascular disease in dialysis patients. *American Journal of Kidney Diseases*, 45(4 Suppl 3), 2005.

[10] POS-640 study group. Machine learning for intradialytic hypotension prediction in haemodialysis patients. Kidney International Reports, abstract supplement, 2022.

[11] Claudio Ronco, Alessandra Brendolan, Manuela Milan, Maria Pia Rodeghiero, Monica Zanella, and Giuseppe La Greca. Impact of biofeedback-induced cardiovascular stability on hemodialysis tolerance and efficiency. *Kidney International*, 58(2):800–808, 2000. doi: 10.1046/j.1523-1755.2000.00229.x.

[12] Antonio Santoro, Elena Mancini, Francesca Paolini, and Pietro Zucchelli. Blood volume regulation during hemodialysis. *American Journal of Kidney Diseases*, 32(5):739–746, 1998. doi: 10.1016/S0272-6386(98)70128-3.

[13] Ashish Vaswani et al. Attention is all you need. *Advances in Neural Information Processing Systems*, 30, 2017.

[14] Zhengxuan Wu et al. pyvene: A library for understanding and improving pytorch models via interventions. arXiv:2403.04640, 2024.

[15] Si Yong Yeo. Md6329 assignment 2: Electronic health records — transformer encoder and intervention analysis. Course assignment, Nanyang Technological University, 2026.

[16] Si Yong Yeo and guest lecturers. Md6329 deep learning for healthcare ai — lecture recordings. YouTube playlist PLORG7zBQosxw; NTU Learn, 2026.