

# Daily Log

---

Tuesday, May 04, 2021 10:42 AM

10:42 starting cooling process

12:46 turn on the dummy heaters

12:58 turn on the reservoir heater(high)

13:25 turn on the reservoir heater(low)

14:31 superheat problem(after 90 second, stop)

14:40 set chiller temperature to -65

15:31 set chillier temperature to -70

15:52 turn off dummy heater and reservoir heater/chiller