Instructions Test Strip:

1. Using the forceps, insert one testing strip into testing tube #1 for 2 seconds and remove quickly.
2. Place the strip onto the sheet of paper and let sit for 30 seconds and compare results using the chart provided.
3. Read the results within another 30 seconds and record the number onto the corresponding box.
4. Repeat steps 1-3 for the other 3 samples of water.

Instructions Turbidity:

1. Place sample vial #1 and use a KimTech wipe to clean the outside. Place the vial in the turbidimeter.
2. Press the “read” button and record the number onto the corresponding “turbidity” box.
3. Repeat steps 1-2 for the other 3 samples of water.

