



Faculty of Engineering
Department of Biomedical Engineering

Ethics Declaration for The
Non-Contact Clinical Thermometer

Part Number: 01-MIY-007

Rev-1

Date: 18/12/2025



CHANGE RECORDS

FIRST RELEASE

CREATED BY

Name	Title	Date	Sign.
Gülşen Demir	System Engineer	18.12.2025	
Augusta Nwosu	Electrical Engineer	18.12.2025	

CHECKED BY

Name	Title	Date	Sign.
Fadime Taş	Quality Engineer	18.12.2025	

APPROVED BY

Name	Title	Date	Sign.
Efe Ateş	Program Manager	18.12.2025	

REVISIONS

Rev. No	Change Description	Date	QA Approval
---------	--------------------	------	-------------

Doc. No: Part Number: 01-MIY-007	Ethics Declaration for Non-Contact Type Clinical Thermometer	Rev:01	Page 2/4
ALL RIGHTS RESERVED. REPRODUCTION OR ISSUE TO THIRD PARTIES IN ANY FORM WHATEVER IS NOT PERMITTED WITHOUT WRITTEN AUTHORITY FROM THE PROPRIETORS.			



TABLE OF CONTENTS

CHANGE RECORDS.....	2
FIRST RELEASE.....	2
REVISIONS.....	2
1. PURPOSE.....	4
2. SCOPE.....	4
3. ETHICAL DECLARATIONS.....	4
4. REFERENCES.....	4

Doc. No: Part Number: 01-MIY-007	Ethics Declaration for Non-Contact Type Clinical Thermometer	Rev:01	Page 3/4
ALL RIGHTS RESERVED. REPRODUCTION OR ISSUE TO THIRD PARTIES IN ANY FORM WHATEVER IS NOT PERMITTED WITHOUT WRITTEN AUTHORITY FROM THE PROPRIETORS.			



1. PURPOSE

The purpose of this ethics declaration is to evaluate and address the ethical considerations associated with the design and development of a non-contact body temperature measurement device intended for users. The report aims to ensure that ethical principles such as user safety, privacy, accuracy, and respect for human dignity are upheld throughout the design process and potential use of the device.

2. SCOPE

The scope of this ethics declaration covers ethical aspects associated with the design, development, and intended clinical use of the non-contact thermometer. The device is designed to measure body temperature without physical contact, supporting hygienic practices in clinical environments. This report applies to the academic design phase of the project and does not include human testing or regulatory approval processes.

3. ETHICAL DECLARATIONS

The device adheres to the following medical standards. ISO 80601-2-56:2017: Medical electrical equipment – Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement. It defines the accuracy requirements $\pm 0.5^{\circ}\text{C}$ in the 34°C - 42°C range and testing methods for infrared skin thermometers. IEC 60601-1 (General Safety): Medical electrical equipment – Part 1: General requirements for basic safety and essential performance. Ensuring the use of low-voltage (3.3V) DC power inherently complies with the safety limits for "Applied Parts." ISO 15223-1:2016, Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied — Part 1: General requirements

4. REFERENCES

[1] "BS EN ISO 80601-2-56:2017+A1:2020 | 31 May 2020 | BSI Knowledge," *Bsigroup.com*, May 31, 2020.

<https://knowledge.bsigroup.com/products/medical-electrical-equipment-particular-requirements-for-basic-safety-and-essential-performance-of-clinical-thermometers-for-body-temperature-measurement> (accessed Dec. 22, 2025).

Doc. No: Part Number: 01-MIY-007	Ethics Declaration for Non-Contact Type Clinical Thermometer	Rev:01	Page 4/4
ALL RIGHTS RESERVED. REPRODUCTION OR ISSUE TO THIRD PARTIES IN ANY FORM WHATEVER IS NOT PERMITTED WITHOUT WRITTEN AUTHORITY FROM THE PROPRIETORS.			