

INSTRUCTIONS FOR USE

Transform S H Files

Stainless steel Hedstrom hand files

FOR PROFESSIONAL DENTAL USE

01 Intended use and product range

For manual root canal filing and dentine removal, by dental professionals trained in endodontics. Identify the file by its article number, tip size, taper and length.

Tip sizes	6, 8, 10, 15, 20, 25, 30, 35, 40, 45, 50, 55, 60, 70, 80
Taper	.02
Lengths	21 and 25 mm

Not all size–length combinations are available. Range and final device specifications require approval.

02 Composition and presentation

Stainless steel working part with a manual handle and depth stop. This draft covers the planned sterile, single-use presentation. Final component materials and coatings require confirmation.

03 Contraindications and safety

Do not use in patients allergic to a constituent material. Check the alloy declaration for patients with metal allergies. Do not force a file through a blocked or unidentifiable path.

Use rubber dam isolation, eye protection and standard infection-control precautions. Sharp instruments can cause injury. Misuse or damage may cause fracture, canal transportation, perforation or cross-contamination.

Do not use expired products, open or damaged sterile packs, or files with corrosion, cracks, unwound flutes or other damage. Do not reuse or resterilise single-use files.

Hedstrom files cut on withdrawal. Do not rotate an engaged H file or connect it to a motor.

04 Preparation and operating mode

Manual use only. Do not connect to a motor.

Before issue: confirm manual technique and movement limits.

Assess clinical findings, radiographs and canal anatomy. Prepare access and determine working length. Open the pack aseptically before use; check file identity and stop position. Maintain a controlled manual grip.

05 Clinical sequence

1. Establish an accessible canal and working length using suitable exploration instruments before introducing a Hedstrom file.
2. Select a file that enters passively. Set the stop to the intended depth without wedging the cutting flutes.
3. Apply controlled withdrawal filing strokes. Do not rotate or use a reaming motion with the file engaged in the canal.
4. Withdraw, clean and inspect frequently, and irrigate between insertions. Increase file size only where the canal permits passive insertion.
5. If the file binds, stop and release it carefully without twisting. Discard any damaged instrument and reassess access and size selection.

06 Irrigation and instrument care

Maintain irrigation and remove debris during preparation. Clear flutes without damaging edges. Do not immerse the handle or entire instrument in irrigant. Before issue, confirm compatible solutions, concentrations and contact times for this device.

07 Storage disposal and reporting

Protect unopened packs from moisture, sunlight and damage. Before issue, confirm temperature, humidity and shelf-life limits. Follow the final pack expiry date and batch information. Discard used files in an approved sharps container under the practice's clinical-waste procedure.

Record article and batch details where practicable. Report defects or adverse incidents to the NZ distributor, retaining packaging and relevant device details.

08 New Zealand distributor

ENDOTECHNZ LIMITED

3c/14 Beaumont St, Auckland 1010, New Zealand
Steve Shepherd · 021 543 414