INSTRUCTIONS FOR USE

Transform S ET

Nickel-titanium root canal shaping files with .04 and .06 tapers

FOR PROFESSIONAL DENTAL USE

01 Intended use and product range

For mechanical preparation and shaping of root canals, by dental professionals trained in endodontics. Identify the file by its article number, tip size, taper and length.

Shaping files	17–45 in .04 and .06; additional 15/.04
Shaping lengths	21, 25 and 29 mm
Orifice opener	17/.08, 17 mm

Not all size–length combinations are available. Range and final device specifications require approval.

02 Composition and presentation

Nickel-titanium working part with a handpiece shank and depth stop. This draft covers the planned sterile, single-use presentation. Final component materials and coatings require confirmation.

03 Contraindications and safety

Do not use in patients allergic to a constituent material. Nickel-titanium contains nickel. Avoid powered instrumentation in severe, abrupt apical curves where safe progression cannot be achieved.

Use rubber dam isolation, eye protection and standard infection-control precautions. Sharp instruments can cause injury. Misuse or damage may cause fracture, canal transportation, perforation or cross-contamination.

Do not use expired products, open or damaged sterile packs, or files with corrosion, cracks, unwound flutes or other damage. Do not reuse or resterilise single-use files.

Use minimal pressure. Stop when a file binds or ceases to advance. Inspect throughout treatment; fatigue may occur without visible damage.

04 Preparation and operating mode

Continuous clockwise rotation with a suitable torque-controlled endodontic motor.

Before issue: confirm speed, torque and the program for each file group. Do not use the draft protocol clinically.

Assess clinical findings, radiographs and canal anatomy. Prepare access and determine working length. Open the pack aseptically before use; check file identity and stop position. Confirm secure handpiece retention and the approved program.

05 Clinical sequence

1. Explore the canal with small hand files, confirm working length and establish a smooth, reproducible glide path.
2. If required, use the 17/.08 opener only for coronal access. Protect thin dentine and the furcation.
3. Choose .04 or .06 taper according to anatomy. Introduce a shaping file that follows the prepared path without binding, and enlarge progressively using the approved ET sequence.
4. When advancement becomes difficult, withdraw the file and inspect it. Irrigate, recapitulate with a small hand file and reassess the path before selecting a smaller instrument. Never force progression.
5. Recheck working length and gauge apical size before choosing any further enlargement. The 15/.04 is an additional small instrument; its place in the sequence must follow the final approved ET protocol.

06 Irrigation and instrument care

Maintain irrigation and remove debris during preparation. Clear flutes without damaging edges. Do not immerse the handle or entire instrument in irrigant. Before issue, confirm compatible solutions, concentrations and contact times for this device.

07 Storage disposal and reporting

Protect unopened packs from moisture, sunlight and damage. Before issue, confirm temperature, humidity and shelf-life limits. Follow the final pack expiry date and batch information. Discard used files in an approved sharps container under the practice's clinical-waste procedure.

Record article and batch details where practicable. Report defects or adverse incidents to the NZ distributor, retaining packaging and relevant device details.

08 New Zealand distributor

ENDOTECHNZ LIMITED

3c/14 Beaumont St, Auckland 1010, New Zealand
Steve Shepherd · 021 543 414