

INSTRUCTIONS FOR USE

Transform S CT

Constant taper nickel-titanium root canal shaping files

FOR PROFESSIONAL DENTAL USE

01 Intended use and product range

For mechanical preparation and shaping of root canals, by dental professionals trained in endodontics. Identify the file by its article number, tip size, taper and length.

Shaping files	.04: sizes 15–80; .06: sizes 15–50
Shaping lengths	21, 25 and 31 mm; availability varies by size
Expeditor	27/.04, 21 mm
Orifice opener	20/.08, 16 mm

Not all size–length combinations are available. Range and final device specifications require approval.

02 Composition and presentation

Nickel-titanium working part with a handpiece shank and depth stop. This draft covers the planned sterile, single-use presentation. Final component materials and coatings require confirmation.

03 Contraindications and safety

Do not use in patients allergic to a constituent material. Nickel-titanium contains nickel. Avoid powered instrumentation in severe, abrupt apical curves where safe progression cannot be achieved.

Use rubber dam isolation, eye protection and standard infection-control precautions. Sharp instruments can cause injury. Misuse or damage may cause fracture, canal transportation, perforation or cross-contamination.

Do not use expired products, open or damaged sterile packs, or files with corrosion, cracks, unwound flutes or other damage. Do not reuse or resterilise single-use files.

Use minimal pressure. Stop when a file binds or ceases to advance. Inspect throughout treatment; fatigue may occur without visible damage.

04 Preparation and operating mode

Continuous clockwise rotation with a suitable torque-controlled endodontic motor.

Speed 500–600 rpm · Initial torque 1.8–2.3 N·cm

Maximum torque 2.9 N·cm

05 Clinical sequence

1. Confirm coronal patency with a size 10 hand file. Never advance a rotary file beyond the path already explored manually.
2. Introduce the expeditor until first wall engagement, then withdraw. Engagement before half its cutting length indicates a small canal; halfway or deeper indicates medium; loose entry to full length indicates large.
3. Choose the crown-down group: small 30 → 25 → 20 → 15; medium 40 → 35 → 30 → 25; large 50 → 45 → 40 → 35. The extra-large .04 group is 80 → 70 → 60 → 55.
4. After the second rotary instrument, determine working length with a size 10 hand file and apex locator. Continue crown-down to engagement. Finish at the first file that engages at length, or one size larger. If size 15 reaches length, return with size 20; do not finish below size 20.
5. For extra-small canals, use .04 taper: 25 → 30; determine working length; return to 25, then 20 and, if needed, 15. If 15 is needed, finish with 20. A size 15 hand-file glide path can assist narrow-canal preparation.
6. Use three gentle engagements per series, withdrawing completely after the third; each series takes 2–3 seconds. Use two series before changing files. Wipe flutes after each series; maintain lubrication, irrigation, movement and a firm finger rest. Never force advancement.
7. Obturate with a matching cone using the selected obturation technique.

06 Irrigation and instrument care

Maintain irrigation and remove debris during preparation. Clear flutes without damaging edges. Do not immerse the handle or entire instrument in irrigant. Before issue, confirm compatible solutions, concentrations and contact times for this device.

07 Storage disposal and reporting

Protect unopened packs from moisture, sunlight and damage. Before issue, confirm temperature, humidity and shelf-life limits. Follow the final pack expiry date and batch information. Discard used files in an approved sharps container under the practice's clinical-waste procedure.

Record article and batch details where practicable. Report defects or adverse incidents to the NZ distributor, retaining packaging and relevant device details.

08 New Zealand distributor

ENDOTECHNZ LIMITED

3c/14 Beaumont St, Auckland 1010, New Zealand
Steve Shepherd · 021 543 414