Catheterization Kit

Category: **Medical, Healthcare & First Aid** Type: **Full SOP** Priority: **High** 5 labs

Covers these lab items: Advanced Male and Female Catheterization Kit; Folley catheter; Red rubber catheter for Basti Administration; Rubber Catheter

What it is used for

A catheterization kit contains sterile instruments and supplies used by trained healthcare professionals for urinary catheterization or other specified clinical procedures. Correct aseptic technique, patient identification, and appropriate catheter selection help reduce infection and injury.

Main parts

- Catheter drains or delivers fluid through the intended anatomical passage.
- Sterile gloves help maintain aseptic technique during the procedure.
- Sterile drape provides a clean working field.
- Antiseptic or cleansing supplies prepare the specified area.
- Lubricating gel reduces friction during catheter insertion when indicated.
- Drainage bag and tubing collect urine in urinary catheter kits.
- Syringe, sterile water, or other accessories are included only when specified for the kit and procedure.

Before you start

1. Only trained and authorised healthcare personnel should perform catheterization.
2. Confirm the patient's identity, procedure, indication, consent, and relevant allergies according to clinical policy.
3. Explain the procedure and provide privacy and appropriate positioning.
4. Check the package seal, expiry date, sterility indicators where provided, and kit contents.
5. Gather the correct catheter type and size and all required sterile supplies.
6. Perform hand hygiene and prepare a clean area using the facility's infection-control procedure.

How to use

1. Open the kit using aseptic technique and avoid touching sterile items with unclean hands.
2. Put on the required PPE and sterile gloves according to clinical protocol.
3. Prepare the patient and clean the specified anatomical area using the approved procedure.
4. Arrange the sterile drape and supplies without contaminating the sterile field.
5. Apply the approved lubricant or antiseptic only as directed for the procedure.
6. Insert the catheter only using the approved clinical technique and without forcing it.
7. For urinary catheterization, follow the device instructions for confirming placement before any specified balloon inflation.
8. Connect the appropriate drainage or delivery system using aseptic technique when required.

9. Secure the catheter, maintain appropriate drainage, and monitor the patient for discomfort or complications.
10. Dispose of single-use items safely, document the procedure, and clean reusable equipment according to policy.

⚠ Safety

- Never perform catheterization without appropriate clinical training, authorisation, and a valid indication.
- Do not use a package that is open, damaged, expired, or no longer sterile.
- Never force the catheter if resistance, severe pain, or unexpected bleeding occurs; stop and seek clinical assistance.
- Use aseptic technique to reduce the risk of introducing infection.
- Inflate a catheter balloon only when the specific device instructions and confirmed placement permit it.
- Follow clinical policy for sharps, contaminated materials, and medical waste disposal.

Care & storage

- Store unopened kits in a clean, dry area within the specified storage conditions.
- Check package integrity and expiry dates before each use.
- Dispose of single-use catheters and supplies after use; do not reuse them.
- Report missing, damaged, contaminated, or expired kit components and replace the kit as required.

Common mistakes to avoid

- Using a non-sterile or expired catheterization kit.
- Touching sterile components with unclean hands or contaminating the sterile field.
- Forcing insertion despite resistance, pain, or bleeding.
- Inflating a urinary catheter balloon before confirming correct placement according to device instructions.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.