Basti Yantra



Category: **Beauty, Wellness & Fitness** Type: **Full SOP** Priority: **High** 2 labs

Covers these lab items: Basti; Basti yantra

What it is used for

A Basti Yantra is a device used in traditional Ayurvedic practice to administer a prescribed enema preparation through the rectal route under appropriate clinical supervision. Correct patient assessment, hygienic preparation, suitable equipment, and careful administration are essential to reduce injury, infection, and other complications.

Main parts

- Reservoir or container holds the prescribed liquid preparation.
- Connecting tube carries the preparation from the reservoir.
- Nozzle or rectal tip is designed for the intended route of administration.
- Clamp or flow control regulates liquid delivery where fitted.
- Connectors join the reservoir, tube, and nozzle.
- Support or stand holds the reservoir where included.
- Disposable or reusable components depend on the device design and instructions.

Before you start

1. Only qualified Ayurvedic or healthcare practitioners trained in the procedure should administer Basti.
2. Confirm the patient's identity, indication, prescribed preparation, volume, temperature, and relevant contraindications.
3. Explain the procedure and obtain informed consent according to clinical policy.
4. Inspect the reservoir, tube, nozzle, connectors, and flow control for damage, blockage, or leakage.
5. Confirm that reusable components have been properly cleaned and processed, or that disposable items are sterile and within expiry.
6. Prepare the prescribed solution using approved ingredients and the correct clinical method; do not use unprescribed substances.
7. Perform hand hygiene, wear appropriate gloves and PPE, and prepare a private, clean treatment area.

How to use

1. Position the patient as directed by the qualified practitioner and maintain privacy and comfort.
2. Prepare and verify the prescribed liquid, ensuring its volume and temperature meet the clinical instructions.

3. Assemble the device according to the manufacturer's instructions and check that connections are secure.
4. Ensure the nozzle or rectal tip is clean, appropriate for the patient, and lubricated with an approved lubricant when indicated.
5. Prime the tubing as required by the device instructions while preventing contamination and excessive air entry.
6. Administer the preparation gently using the prescribed technique and controlled flow under clinical supervision.
7. Monitor the patient for pain, abdominal discomfort, dizziness, bleeding, nausea, or other unexpected reactions.
8. Stop administration immediately if significant pain, bleeding, resistance, or other concerning symptoms occur and seek clinical assessment.
9. Complete the procedure according to the prescribed protocol and assist the patient with safe recovery and hygiene.
10. Dispose of single-use items and clinical waste according to facility policy; clean reusable components using validated procedures.
11. Document the preparation, administered amount, patient response, and any adverse event as required.

⚠ Safety

- Do not perform the procedure without appropriate clinical training, a valid indication, and informed consent.
- Do not administer an unprescribed preparation or exceed the prescribed volume, temperature, or rate.
- Do not force the nozzle or fluid if resistance or significant pain occurs.
- Seek clinical assessment before use when contraindications or warning signs such as rectal bleeding, severe abdominal pain, suspected obstruction, or suspected perforation are present.
- Use appropriate infection-control precautions and never reuse items designated for single use.
- Follow clinical protocols for patients who are medically vulnerable or require additional monitoring.

Care & storage

- Clean and process reusable components immediately after use according to validated instructions.
- Inspect the tube, nozzle, reservoir, and connectors for cracks, residue, blockage, and leakage.
- Dry components completely where required and store them in a clean, protected area.
- Check expiry and packaging integrity of sterile disposable components before use.
- Remove damaged equipment from service and arrange authorised replacement or repair.

Common mistakes to avoid

- Using the device without confirming the patient's suitability and prescribed preparation.

- Using inadequately cleaned or contaminated components.
- Applying excessive force or delivering the preparation too quickly.
- Ignoring pain, bleeding, resistance, or other signs of complications.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.