Infrared Lamp



Category: **Medical, Healthcare & First Aid** Type: **Full SOP** Priority: **High** 2 labs

Covers these lab items: Infrared Lamp

What it is used for

An infrared lamp emits infrared radiation to provide controlled surface warming for approved therapeutic or wellness applications. Use only for the intended purpose and follow the device instructions, prescribed exposure time, and required safety precautions.

Main parts

- Infrared bulb or heating element
- Reflector and protective lamp housing
- Adjustable stand or positioning arm
- Power switch and intensity control, if fitted
- Timer, if fitted
- Power cord, plug, and stable base

Before you start

1. Confirm that the device is approved and suitable for the intended application.
2. Inspect the bulb, housing, stand, cord, plug, and controls for damage.
3. Position the lamp on a stable surface with adequate clearance.
4. Check the required distance, exposure time, and settings in the device instructions or care plan.
5. Assess the user's skin sensitivity and ability to sense heat, following clinical guidance.
6. Remove flammable materials and ensure the treatment area is dry.

How to use

1. Switch off the lamp before adjusting its position or replacing accessories.
2. Position the lamp at the recommended distance from the treatment area.
3. Secure the stand or positioning arm to prevent accidental movement.
4. Connect the device to a suitable electrical outlet.
5. Switch on the lamp and select the approved setting.
6. Set the exposure time according to the device instructions or authorised care plan.
7. Keep the treated area within the recommended exposure zone.
8. Monitor the user's comfort and skin response throughout the session.
9. Stop immediately if excessive heat, pain, burning, or unusual redness occurs.
10. Switch off the lamp at the end of the session.
11. Allow the bulb and housing to cool before moving or cleaning the device.
12. Disconnect the power when required and store the lamp safely.
13. Record the treatment and report any equipment faults or adverse response.

⚠ Safety

- Use the lamp only for its approved purpose and under appropriate supervision.
- Do not look directly at the infrared source or exceed the recommended exposure.
- Keep the lamp away from water, flammable materials, and combustible surfaces.
- Avoid use on areas with reduced heat sensation unless specifically authorised by a qualified clinician.
- Do not leave the lamp operating unattended during treatment.
- Do not touch the bulb or housing while hot; disconnect power before maintenance.

Care & storage

- Clean the housing only after the lamp has cooled and been disconnected.
- Inspect the bulb, reflector, stand, and electrical cable regularly.
- Replace bulbs only with the specified compatible type.
- Keep vents and reflector surfaces free from dust.
- Arrange authorised servicing if the lamp flickers, overheats, or becomes unstable.

Common mistakes to avoid

- Placing the lamp too close to the skin.
- Exceeding the authorised exposure time or intensity.
- Leaving the user unattended during heating.
- Moving or cleaning the lamp before it has cooled.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.

Prepared by - Subodh Kumar