High Frequency Machine



Category: **Beauty, Wellness & Fitness** Type: **Full SOP** Priority: **High** 1 labs

Covers these lab items: High frequency

What it is used for

To use a high-frequency beauty device for approved skin-care treatments according to the device manual and professional training. To apply the correct electrode and settings while protecting the client from burns, irritation, and electrical hazards.

Main parts

- Main unit and control dial
- High-frequency handpiece
- Glass electrodes
- Electrode holder or connector
- Power switch and cord
- Timer or intensity control, if fitted

Before you start

1. Confirm the device is intended for the planned treatment and is in good working condition.
2. Inspect the power cord, handpiece, connectors, and glass electrodes for cracks or damage.
3. Check the client's treatment history, skin condition, and contraindications according to professional protocol.
4. Remove metal jewellery from the treatment area and ensure the skin is clean and dry.
5. Explain the procedure and obtain informed consent according to workplace policy.
6. Ensure the intensity control is at minimum and the device is switched off before fitting an electrode.

How to use

1. Wash hands and prepare the treatment area using the approved hygiene procedure.
2. Select the correct intact glass electrode for the intended treatment.
3. Fit the electrode securely into the handpiece while the device is switched off.
4. Position the client comfortably and provide suitable eye protection if required by the device instructions.
5. Connect the device to the approved power source and switch it on at the lowest setting.
6. Increase intensity gradually only within the device instructions and client comfort limits.
7. Keep the electrode in controlled contact or use the specified technique for the selected treatment.
8. Avoid prolonged treatment on one spot and monitor the skin continuously for discomfort or excessive redness.

9. Reduce intensity to zero and switch off before removing or changing the electrode.
10. Disconnect the handpiece and carefully remove the electrode after the device has stopped.
11. Clean reusable components only with manufacturer-approved products and disinfect them as directed.
12. Record the treatment and report any unexpected skin reaction.

⚠ Safety

- Use only by trained beauty or skin-care professionals within their scope of practice.
- Do not use on clients with contraindications listed in the manual, such as certain implanted electronic devices or other specified medical conditions.
- Do not use near flammable products, alcohol vapours, oxygen-enriched environments, or other fire hazards.
- Never fit or remove a glass electrode while the device is energized.
- Do not use cracked electrodes, wet hands, damaged cables, or a device with faulty insulation.
- Stop immediately if the client experiences pain, burning, unusual tingling, or unexpected skin changes.

Care & storage

- Clean and disinfect handpieces and reusable electrodes according to the manual.
- Inspect glass electrodes for cracks before and after every treatment.
- Store electrodes in protective holders to prevent breakage.
- Keep the unit dry and avoid pulling the power cord.
- Arrange inspection and servicing if the unit overheats, sparks abnormally, or malfunctions.

Common mistakes to avoid

- Starting at excessive intensity instead of increasing gradually.
- Using a cracked electrode or fitting it while the device is on.
- Applying the treatment over wet skin or near flammable products.
- Ignoring contraindications or client discomfort.
- Using unapproved cleaning chemicals that can damage electrodes or the handpiece.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.