Nadi Swedana Yantra



Category: **Beauty, Wellness & Fitness** Type: **Full SOP** Priority: **High** 1 labs

Covers these lab items: Nadee Swedan Yantra

What it is used for

A Nadi Swedana Yantra is a traditional Ayurvedic steam apparatus used to direct steam toward a selected body area under trained supervision. It requires careful control of steam flow, temperature and exposure to reduce burn risk.

Main parts

- Identify the steam-generating vessel or boiler.
- Locate the water-filling opening and specified water-level indicator, if fitted.
- Identify the heating element or fuel-fired heating arrangement.
- Locate the steam outlet pipe and directing nozzle.
- Identify the steam-flow control valve, where provided.
- Locate the temperature and pressure indicators, if fitted.
- Identify the support frame, insulated handles and power or fuel controls.

Before you start

1. Use the apparatus only under a trained Ayurvedic or healthcare practitioner's supervision.
2. Read the operating instructions and confirm the equipment is designed for this use.
3. Inspect the vessel, pipe, nozzle, seals and controls for damage or leaks.
4. Fill with the specified water quantity before switching on the heater.
5. Place the apparatus on a stable, heat-resistant surface in a ventilated area.
6. Keep clean towels, suitable protective equipment and first-aid supplies accessible.
7. Explain the procedure and check for heat sensitivity or other relevant precautions.

How to use

1. Position the apparatus securely away from the user and other people.
2. Fill the vessel with the specified amount of clean water.
3. Close the filling opening and confirm all required connections are secure.
4. Switch on the heater or operate the approved heat source as instructed.
5. Allow steam to develop while monitoring the specified operating indicators.
6. Adjust the steam flow using the designated control, if provided.
7. Test steam delivery safely away from the body before beginning exposure.
8. Direct steam toward the intended area from the practitioner-recommended distance.
9. Monitor skin response and comfort continuously throughout the procedure.
10. Stop immediately if excessive heat, discomfort or skin redness develops.
11. Switch off the heat source and allow the apparatus to cool before draining.

⚠ Safety

- Never direct steam toward the face, eyes, wounds or sensitive areas without specific clinical guidance.
- Keep hands and skin away from the nozzle and escaping steam.
- Never open a hot pressurised vessel or block the steam outlet.
- Do not leave the heating apparatus unattended during operation.
- Use only the specified water quantity and never operate the heater dry.
- Stop exposure immediately if burning, dizziness or unusual discomfort occurs.

Care & storage

- Switch off and cool the apparatus completely before cleaning.
- Drain remaining water safely and clean the vessel and nozzle as instructed.
- Inspect pipes, seals, valves and electrical components for damage after use.
- Store the dry apparatus in a clean, ventilated and secure area.

Common mistakes to avoid

- Starting the heater without checking the required water level.
- Directing hot steam too close to skin or for too long.
- Opening the vessel before it has cooled and pressure has safely released.
- Using damaged pipes or blocked nozzles that cause unsafe steam release.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.