Phosphine Gas Monitor



Category: **Safety & PPE** Type: **Short SOP** Priority: **High** 6 labs

Covers these lab items: Phosphine Alert personal monitor; Phosphine gas Monitor

What it is used for

A phosphine gas monitor detects or warns of phosphine gas in the surrounding atmosphere. Use a suitable, functioning monitor to assess potential exposure and follow the site's emergency procedure.

How to use

1. Inspect the monitor, sensor openings, battery, display, and alarm indicators.
2. Confirm that the monitor is suitable for phosphine and within its calibration period.
3. Perform the prescribed bump test or function check before use.
4. Switch on the monitor in a safe area and allow its startup checks to finish.
5. Wear or position the monitor as specified by the manufacturer and site procedure.
6. Observe readings and alarms while working in the monitored area.
7. If an alarm activates, leave the area immediately and follow the site emergency plan.
8. Report alarms and arrange testing, calibration, or servicing before returning the monitor to service.

⚠ Safety

- Never enter a suspected contaminated area to test the gas without authorised training and protection.
- Do not ignore alarms or assume an area is safe from a single reading.
- Use the site's emergency response and evacuation procedures when an alarm occurs.
- Only trained personnel should perform calibration and sensor maintenance.

Care & storage

- Keep the monitor clean, dry, and within the specified operating conditions.
- Charge or replace batteries and complete scheduled bump tests and calibration.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.