AED Trainer

Category: **Medical, Healthcare & First Aid** Type: **Full SOP** Priority: **Medium** 1 labs

Covers these lab items: AED Trainer with Adult Pad

What it is used for

An AED trainer simulates the operation of an automated external defibrillator for first-aid and CPR training. It teaches learners to follow voice prompts, place training pads correctly, and maintain scene safety without delivering a therapeutic shock.

Main parts

- AED trainer unit with control buttons
- Adult training electrode pads
- Training pad cable and connector
- Speaker and voice-prompt system
- Training remote control if supplied
- Carrying case and batteries or power adapter

Before you start

1. Confirm that the unit is an AED trainer and not a clinical defibrillator.
2. Inspect the trainer, pad cables, connectors, controls, and casing for damage.
3. Check the battery level or approved power supply before starting the session.
4. Select the appropriate training scenario and language for the learners.
5. Explain scene safety, emergency response, CPR, and the purpose of AED training.
6. Check that adult training pads are clearly marked and that no therapeutic shock can be delivered.

How to use

1. Switch on the AED trainer and select the required training scenario.
2. Explain that learners must first check scene safety and responsiveness in a real emergency.
3. Demonstrate calling emergency services and asking someone to obtain an actual AED.
4. Position the training mannequin or practice partner as directed by the scenario.
5. Expose the chest area and demonstrate correct placement of the adult training pads.
6. Connect the training pad cable to the trainer if it is not already connected.
7. Follow the voice prompts and demonstrate checking that nobody touches the mannequin during simulated analysis.
8. When prompted, demonstrate the simulated shock sequence or the no-shock response without using a real defibrillator.
9. Resume CPR when the scenario instructs and follow the trainer's compression and ventilation prompts where available.
10. Reset the scenario, remove the training pads carefully, and switch off the trainer after practice.

⚠ Safety

- Use only designated training pads and accessories with the AED trainer.
- Never connect training pads to a person for therapeutic use or to a clinical defibrillator unless explicitly approved for that purpose.
- Make clear that the trainer does not provide a real shock and is not emergency treatment equipment.
- Keep cables arranged to prevent tripping and protect connectors from liquids.
- Use an actual AED and contact emergency services in a real cardiac emergency.
- Clean reusable accessories according to the manufacturer's instructions between learners.

Care & storage

- Wipe the trainer and reusable accessories with approved cleaning materials after each session.
- Inspect pad cables, connectors, buttons, and casing for wear or damage.
- Replace batteries and worn training pads according to the manufacturer's instructions.
- Store the trainer and pads in a clean, dry case and check completeness after use.

Common mistakes to avoid

- Using a clinical defibrillator instead of a trainer can create serious danger during practice.
- Placing pads incorrectly can teach unsafe or ineffective technique.
- Skipping the clear-area instruction can lead learners to overlook shock safety in real emergencies.
- Using damaged cables or depleted batteries can interrupt the training scenario.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.