DO Meter



Category: **Laboratory Instruments & Glassware** Type: **Short SOP** Priority: **Medium**

1 labs

Covers these lab items: Dissolved Oxygen (DO) meter

What it is used for

A dissolved oxygen (DO) meter measures the oxygen dissolved in a water sample. Calibrate the meter as instructed and handle the probe carefully for reliable readings.

How to use

1. Inspect the meter, probe, cable, and sensor membrane or cap for damage.
2. Switch on the meter and check battery status and measurement units.
3. Calibrate the meter according to the manufacturer's procedure and required conditions.
4. Rinse the probe with clean water and gently remove excess droplets as instructed.
5. Place the probe in the sample at the required depth without trapping air bubbles.
6. Wait for the reading to stabilize and record DO, temperature, units, and sample details.
7. Rinse the probe after measurement and store it according to its probe type.

⚠ Safety

- Follow laboratory precautions for unknown or contaminated water samples.
- Keep electrical connectors dry and do not immerse parts beyond their rated limit.
- Do not touch or damage the sensor membrane or sensitive probe tip.

Care & storage

- Clean the probe and maintain its storage cap or solution according to the manufacturer's instructions.

Disclaimer: This SOP is prepared from general information available on the internet. For detailed instructions, refer to the manufacturer's original manual / file.