Northfield Metrology Ltd. · 1200 Industrial Parkway, Suite 4 · Example City, ON

Certificate of Calibration

Certificate No. NM-2026-04417 · Issued 14 October 2026

Instrument

Description	Digital outside micrometer, 0–25 mm
Manufacturer / model	Precisa MX-25
Serial number	PX25-88310
Customer	Lakeshore Components Inc.

Results

Nominal (mm)	Measured (mm)	Error (µm)	Tolerance (µm)	Result
2.500	2.501	+1	±2	Pass
10.300	10.299	−1	±2	Pass
17.600	17.602	+2	±2	Pass
25.000	25.001	+1	±2	Pass

Measurements were made at 20.0 °C ± 0.5 °C using gauge blocks traceable to national standards. The expanded uncertainty is 0.8 µm (k = 2, approximately 95% confidence).

Approval

Technician name: _____

Date of review: _____

☐ Instrument is within tolerance and fit for use

Comments: _____

This certificate may not be reproduced except in full without written approval of the laboratory. This is a fictional sample document for trying out Pagewright.

Notes and Conditions

Method

The micrometer was inspected for damage, cleaned, and checked for flatness and parallelism of the measuring faces before calibration. Readings were taken at four points across the range using grade 0 gauge blocks.

Traceability

Reference standards are calibrated annually by an accredited laboratory. Reference block set GB-112 was last calibrated on 3 March 2026 (report RL-55102).

Validity

Calibration results relate only to the item listed on page 1 at the time of test. The recommended recalibration interval is 12 months, subject to frequency of use and the customer’s quality procedures.

Revision history

Revision	Date	Description
A	14 Oct 2026	Initial issue