

Point-by-point response to the reviewers

The comments from the reviewers are quoted in *italic*, whereas the revised article parts are reported in upright font.

Reviewer #1

I could not find a significant advance in this work. X-ray Talbot-Lau interferometers with high-energy x-rays have been already reported (Ref. 18-20, 23, D. Stutman et al., Appl. Opt. 49 (2010) 4677 etc.). The authors mention that the edge-on-illuminated grating interferometry breaks the current limitations of x-ray phase-contrast imaging, but I don't think that there is clear evidence showing that the authors' method is much more promising than the methods reported in the previous papers. If the authors want to show that the edge-on-illumination method is by far the most promising for high-energy and large-field-of-view x-ray phase imaging, I think that the authors have to quantitatively compare the method with those previously reported in terms of the total exposure time, spatial resolution, signal-to-noise ratio, and so on, and show that the authors' method is much better than the others.

We think that the experiments presented in the current manuscript show indeed significant progress with respect to the cited references. The method presented by Stutman et al. has a fundamental limitation that prevents it from achieving curved structures and large aspect ratios. Moreover, those experiments were performed on a source with 60 kVp, which is significantly lower than the 160 kVp of our source and the 100 keV used as the design energy for our interferometer.

The other works also present results in a much lower energy range or very different conditions, thus preventing a meaningful comparison. Reference 23 reported images taken at 30 and 60 keV, reference 20 also has a design energy of 47.9 keV. Finally, references 18 and 19 present Talbot interferometry at 82 and 123 keV respectively, but only the first result is published on a peer reviewed journal, and both experiments were performed at the ESRF synchrotron, with a monochromatic and coherent beam that is incomparable to a conventional tube.

In order to address the concerns of reviewer #1, the short review of the previous results in the introductory paragraph has been updated:

The vast majority of phase-sensitive techniques, including crystal analyzer based [5, 6] or interferometric [7, 8] methods rely on X-ray beams of high spatial and temporal coherence, which is available only at synchrotron sources. Inline phase contrast [9–11] and Talbot interferometry [12–14] need high spatial coherence but are available on polychromatic microfocus sources. Phase-contrast imaging using X-ray beams of low temporal and spatial coherence such as conventional low-brilliance X-ray tubes have been demonstrated with coded apertures [15] and Talbot-Lau interferometry [16]. Analyzer-based systems have been recently extended to tube sources [17, 18] but only at energies up to the tungsten K_α line at 60 keV. In addition to phase sensitivity, analyzer-based, Talbot interferometry and coded-aperture approaches also provide (with different retrieval mechanisms) information about the integrated local small angle scattering power from microscopic density fluctuations in a specimen [19]. This signal is known under the name of dark-field, scatter or visibility reduction contrast. High-energy Talbot interferometry has been reported so far using a synchrotron source at nominal energies of 82 keV [20] and 123 keV [21]. Using a low-brilliance X-ray tube, 2 Talbot-Lau interferometry was applied so far at 60 keV mean energy [22]. Medical imaging applications may benefit from phase contrast at higher energies: chest or abdominal radiography or CT require energies between 100 and 150 kVp. Other potential applications are homeland security or chip failure analysis, which require high energies for the visualization of materials of high density and atomic number.

Furthermore, the authors mention that the results in Fig. 3 show the benefit of the phase nature of the image, but I think that the authors should provide a more detailed explanation about the contrasts in the images because the contrast of soldering points underneath the resistors should not be reduced in the differential attenuation image for monochromatic x-rays. The authors should explain quantitatively why the contrast is reduced. Otherwise, the authors' assertion is suspicious.

A paragraph has been added to discuss the different beam hardening behaviour of the absorption and differential phase image:

resistors, while in the phase image, they can clearly be identified. The reduced contrast of the soldering points in the absorption image is due to beam hardening. The spectrum impinging on these soldering point is hardened by the resistors in the upper layer, resulting in lower absorption contrast. Due to the weaker energy dependency of phase shifts ($1/E$ compared to $1/E^3$), phase-contrast images are less sensitive to beam hardening [31], which explains the lower contrast reduction of the soldering points underneath the resistors in the phase image of the chip. This result shows the benefit of the phase nature in high-energy X-ray imaging, which may be useful to identify flaws in multilayered structures such as electronic chips.

Reviewer #2

However, this solution comes with significant drawbacks in terms of very limited fields of view 3 cm horizontally, a single detector pixel in the vertical direction), which at the moment seem hard to overcome.

There is no fundamental limitation on the horizontal field of view in our approach. On the contrary, our solution actually solves precisely this issue, by easily matching the spherical wave front with the curved structures. Any limitation to horizontal the field of view given by large glancing angles is therefore removed. The gratings could already be fabricated on larger wafers with the existing technology.

The limitations in the vertical field of view do not come at the expense of an increased dose with our proposed scanning solution. Larger areas could in principle be scanned with short exposure times by stacking an array of edge-illuminated gratings. This scanning technique is already offered in commercial X-ray diagnostic devices, such as the Philips MicroDose mammography system, to reduce scattering. References are now provided in the following paragraph:

Radiographic 2D imaging can be obtained by scanning the sample or a thin fan beam. The scanning technique has been demonstrated to deliver less dose than the conventional approach based on the illumination of a large area. In digital mammography, for instance, where dose is a critical issue, Philips' MicroDose system combines a scanning approach with an highly collimated fan beam [27]. Thanks to the high collimation, the dose deposited on patients has been reported to be significantly lower than with other instruments based on the illumination of a large area detector [28].

Moreover, in the methods section the authors talk of a 5 % visibility, which is even lower than that of the quoted papers by Pfeiffer's group (refs 18, 19), which, at least nominally, reach even higher energies (123 keV). Hence, the reduction in field-of-view does not seem to be counterbalanced by significant improvements in visibility. This is also supported by the relatively poor quality of the differential phase image compared to the derivative of the absorption image, where the only advantage seems to be the visibility of the soldering points beneath the resistors (primarily due to the fact that these do not produce area contrast in the DPC image), while everywhere else the contrast of the differentiated absorption image seems to be higher.

The visibility in our experiment is still low, and this obviously affects the quality of the images, but we also mentioned that this is the first attempt in building gratings with our new design, while the fabrication of traditional structures has been developed for over a decade now. The remarks above about references 18 and 19 (synchrotron source, unpublished results) should also be taken into account here.

Finally, the novelty of the result is diminished by previous demonstrations that high energies in phase-contrast x-ray imaging could be reached by other

methods (e.g. Ignatyev et al JAP 110 (2011) 014906), albeit in that case image quality seems to be lower.

In the paper from Ignatyev et al., an admittedly lower image quality was achieved again with softer X-rays, from 60 to 100 kVp. Thus, we claim that our results are a significantly more promising than the technique described therein.

Reviewer #3

An extended object is imaged by axial scanning, which results in very long exposure times due to phase stepping. The legend of Fig. 3 indicates that one image line of 0.1 mm width is acquired in 6 min. The authors admit this situation in the last sentences of the manuscripts without giving a convincing solution to the problem.

This is indeed an issue, as the exposure time is mostly affected by the low visibility. The manuscript was amended to include a more detailed discussion of our plans to reduce the exposure time. We mention now that raising the visibility to 15 %, a value already reached in lower energy experiments and in some sections of our gratings, would immediately cut the exposure by 9 times:

visibility of the gratings (5 %). The exposure time was chosen in order to get a low noise in the differential phase image. The signal-to-noise ratio (SNR) is proportional to the visibility and the square root of the exposure time [33]. This implies that the exposure times can easily drop by an order of magnitude as these gratings become comparable in quality to those developed in the last ten years. Smaller regions of these gratings actually exhibit a visibility up to 14 % already, indicating that this goal is reachable as the fabrication becomes more reliable and uniform. The detector is also relatively inefficient (30 %) leaving an additional margin for improvement.

The authors suggest that the only imaging methods that can be used with conventional x-ray sources are the in-line phase contrast technique and Talbot interferometry. They fail to mention that analyzer-based imaging systems with x-ray tube sources have been built and successfully used for imaging rather large biological objects. The current exposure times are shorter or comparable to those of the present work, but there are clear indications how the exposure times in x-ray tube based ABI can be reduced by an order of magnitude or more. On the other hand, the upper limit of photon energy is that of the tungsten Kalpha line, 60 keV. For references, see Nesch et al.(2009), Rev. Sci. Instrum. 80, 093702, and Parham et al.(2009), Acad.Radiol. 16, 911-917.

We now included references to these papers, but we point out that our experiments do not have any intrinsic limit to the photon energy.