

Section 14 – Sterilization and shelf life

The DripAssist is not provided sterile and it is not intended to be sterilized by the user. The device does not require sterilization to be used since it only contacts outer surfaces of standard IV administration sets.

[REDACTED]

The device is powered by a single AA battery supplied by the user. A battery indicator light is lit when the battery needs replacement. The battery compartment is easily accessible for the user to replace the battery.