

Section 15 – Biocompatibility

The DripAssist does not include any direct or indirect patient contact components. The device is in contact with outside surfaces of IV administration set components and no patient contact occurs. Therefore, biocompatibility testing was not conducted.

Materials used for the DripAssist housing panels and other external parts consist of [REDACTED] [REDACTED] for the exterior housing and of [REDACTED] for the membrane switch and front decals. All these materials are RoHs compliant.

These materials have a long history of safe use in both medical and consumer products.