MICRO PRECISION

MICRO PRECISION CALIBRATION, INC
2165 N. Glassell St.,
Orange, CA 92865
714-901-5659

ilac-MRA

ANAB
ANSI National Accreditation Board
ACCREDITED
ISO/IEC 17025
CALIBRATION AND DIMENSIONAL MEASUREMENT
AC-1969

Certificate of Calibration

Date: Aug 2, 2024 | **Cert No. 5523631031124783**

Customer:

SILQ TECHNOLOGIES CORPORATION

607 CHARLES E YOUNG DR E
LOS ANGELES CA 90095

		Work Order #:	LA-90065228
		Purchase Order #:	0000131
MPC Control #:	LAF5769	Serial Number:	1125311515
Asset ID:	LAF5769	Department:	N/A
Gage Type:	ANALYTICAL BALANCE	Performed By:	JUAN ACEVES
Manufacturer:	METTLER TOLEDO	Received Condition:	IN TOLERANCE
Model Number:	AB54-S	Returned Condition:	IN TOLERANCE
Size:	51 g	Cal. Date:	August 02, 2024
Temp/RH:	22.0°C / 48.0%	Cal. Interval:	12 MONTHS
Location:	Calibration performed at MPC facility	Cal. Due Date:	August 02, 2025

Calibration Notes:

See attached data sheet (2 pages).

Standards Used to Calibrate Equipment

I.D.	Description.	Model	Serial	Manufacturer	Cal. Due Date	Traceability #
DS3215	WEIGHT SET 29 PCS	E1 E2	589/J232/J234	FUYUE WEIGHT	Sep 30, 2024	2964214A

Procedures Used in this Event

Procedure Name	Description
MPC-WEI-001 Rev. 08	Weighing Instruments, General, Rev.08, Jul-15-2024

Calibrating Technician: JUAN ACEVES

QC Approval: ILYA VAKS

STATEMENTS OF PASS OR FAIL CONFORMANCE: The uncertainty of measurement has been taken into account when determining compliance with specification. All measurements and test results guard banded to ensure the probability of false-accept does not exceed 2% in compliance with ANSI/NCSL Z540.3-2006.

THE CALIBRATION REPORT STATUS:
PASS- Term used when compliance statement is given, and the measurement result is PASS.
PASS^{z}- Term used when compliance statement is given, and the measurement result is conditional passed or PASS^{z}.
FAIL- Term used when compliance statement is given, and the measurement result is FAIL.
FAIL^{z}- Term used when compliance statement is given, and the measurement result is conditional failed or FAIL^{z}.
REPORT OF VALUE - Term used when reported measurement is not requiring compliance statement in report.
ADJUSTED- When adjustments are made to an instrument which changes the value of measurement from what was measured as found to new value as left.
LIMITED - When an instrument fails calibration but is still functional in a limited manner.

The expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%, unless otherwise stated.
This calibration report complies with ISO/IEC 17025:2017, ANSI/NCSL Z540.3-2006 and ANSI/NCSL Z540.1-1994. Calibration cycles and resulting due dates were submitted/approved by the customer. Any number of factors may cause an instrument to drift out of tolerance before the next scheduled calibration. Recalibration cycles should be based on frequency of use, environmental conditions and customer's established systematic accuracy. All standards are traceable to SI through the National Institute of Standards and Technology (NIST) and/or recognized national or international standards laboratories. Services rendered include proper manufacturer's service instruction and are warranted for no less than thirty (30) days. The information on this report pertains only to the instrument identified, this may not be reproduced in part or in a whole without the prior written approval of the issuing MP Calibration Laboratory.